

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN122418\
 Data File : VN053186.D
 Acq On : 24 Dec 2018 15:40
 Operator : MD\SY
 Sample : J6188-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 BU-03-122118

Quant Time: Dec 25 00:22:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 18 13:03:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1253983	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1943847	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1787395	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	721854	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	624027	49.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.08%	
35) Dibromofluoromethane	7.59	113	545169	61.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.80%	
50) Toluene-d8	10.09	98	2406834	50.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.98%	
62) 4-Bromofluorobenzene	12.40	95	820339	49.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.04%	
Target Compounds						
16) Acetone	3.82	43	23759	9.522	ug/l	91
18) Methyl Acetate	4.33	43	5953	1.507	ug/l #	87
20) Methylene Chloride	4.55	84	78462	5.765	ug/l	98
43) Isopropyl Acetate	8.17	43	45499	1.239	ug/l #	87
95) Naphthalene	15.13	128	26970	2.075	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN122418\
Data File : VN053186.D
Acq On : 24 Dec 2018 15:40
Operator : MD\SY
Sample : J6188-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
BU-03-122118

Quant Time: Dec 25 00:22:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 18 13:03:37 2018
Response via : Initial Calibration

