

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN122418\
 Data File : VN053188.D
 Acq On : 24 Dec 2018 16:33
 Operator : MD\SY
 Sample : J6511-13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-58(0-5)-GRAB

Quant Time: Dec 25 00:28:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 18 13:03:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1313154	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2057034	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1886594	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	736664	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	663933	49.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.64%	
35) Dibromofluoromethane	7.59	113	559883	59.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.16%	
50) Toluene-d8	10.09	98	2545133	50.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.92%	
62) 4-Bromofluorobenzene	12.40	95	842672	47.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.18%	
Target Compounds						
16) Acetone	3.82	43	18057	6.911	ug/l	93
18) Methyl Acetate	4.32	43	4284	1.346	ug/l	99
20) Methylene Chloride	4.56	84	38067	2.671	ug/l	96
43) Isopropyl Acetate	8.17	43	53552	1.497	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN122418\
Data File : VN053188.D
Acq On : 24 Dec 2018 16:33
Operator : MD\SY
Sample : J6511-13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
SB-58(0-5)-GRAB

Quant Time: Dec 25 00:28:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 18 13:03:37 2018
Response via : Initial Calibration

