

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN122418\
 Data File : VN053190.D
 Acq On : 24 Dec 2018 17:26
 Operator : MD\SY
 Sample : J6511-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-60(0-5)-GRAB

Quant Time: Dec 25 00:34:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 18 13:03:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1335337	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2062730	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1799484	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	652275	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	665872	49.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.28%	
35) Dibromofluoromethane	7.59	113	560990	59.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.08%	
50) Toluene-d8	10.09	98	2515462	49.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.46%	
62) 4-Bromofluorobenzene	12.40	95	779787	43.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.82%	
Target Compounds						
16) Acetone	3.82	43	17470	6.575	ug/l	92
18) Methyl Acetate	4.34	43	5844	1.467	ug/l	92
20) Methylene Chloride	4.56	84	34288	2.366	ug/l	95
43) Isopropyl Acetate	8.17	43	52510	1.441	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN122418\
Data File : VN053190.D
Acq On : 24 Dec 2018 17:26
Operator : MD\SY
Sample : J6511-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
SB-60(0-5)-GRAB

Quant Time: Dec 25 00:34:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 18 13:03:37 2018
Response via : Initial Calibration

