

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN122418\
 Data File : VN053191.D
 Acq On : 24 Dec 2018 17:53
 Operator : MD\SY
 Sample : J6511-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-61(0-5)-GRAB

Quant Time: Dec 25 00:36:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 18 13:03:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1305079	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2026367	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1759697	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	627221	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	652930	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
35) Dibromofluoromethane	7.59	113	544975	58.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.76%	
50) Toluene-d8	10.09	98	2469122	49.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.38%	
62) 4-Bromofluorobenzene	12.40	95	756872	43.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.78%	
Target Compounds						
16) Acetone	3.82	43	17575	6.768	ug/l	93
18) Methyl Acetate	4.33	43	5806	1.475	ug/l #	65
20) Methylene Chloride	4.55	84	33866	2.391	ug/l	99
43) Isopropyl Acetate	8.17	43	55325	1.622	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN122418\
Data File : VN053191.D
Acq On : 24 Dec 2018 17:53
Operator : MD\SY
Sample : J6511-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
SB-61(0-5)-GRAB

Quant Time: Dec 25 00:36:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 18 13:03:37 2018
Response via : Initial Calibration

