

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN122418\
 Data File : VN053192.D
 Acq On : 24 Dec 2018 18:20
 Operator : MD\SY
 Sample : J6511-17
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-62(0-5)-GRAB

Quant Time: Dec 25 00:38:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 18 13:03:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1277067	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1968805	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1760092	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	632236	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	632457	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
35) Dibromofluoromethane	7.59	113	530993	59.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.08%	
50) Toluene-d8	10.09	98	2404769	49.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.62%	
62) 4-Bromofluorobenzene	12.40	95	764243	45.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.18%	
Target Compounds						
16) Acetone	3.82	43	16136	6.350	ug/l	96
18) Methyl Acetate	4.33	43	6103	1.511	ug/l #	83
20) Methylene Chloride	4.55	84	32151	2.320	ug/l	98
43) Isopropyl Acetate	8.17	43	52998	1.585	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN122418\
Data File : VN053192.D
Acq On : 24 Dec 2018 18:20
Operator : MD\SY
Sample : J6511-17
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
SB-62(0-5)-GRAB

Quant Time: Dec 25 00:38:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 18 13:03:37 2018
Response via : Initial Calibration

