

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN122418\
 Data File : VN053195.D
 Acq On : 24 Dec 2018 19:40
 Operator : MD\SY
 Sample : J6511-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-65(0-5)-GRAB

Quant Time: Dec 25 00:47:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 18 13:03:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1264320	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1957733	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1817321	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	713747	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	641697	50.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.02%	
35) Dibromofluoromethane	7.59	113	531015	59.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.76%	
50) Toluene-d8	10.09	98	2433598	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
62) 4-Bromofluorobenzene	12.40	95	818957	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
Target Compounds						
16) Acetone	3.82	43	17050	6.778	ug/l	99
18) Methyl Acetate	4.33	43	5584	1.471	ug/l	93
20) Methylene Chloride	4.56	84	33942	2.473	ug/l	96
43) Isopropyl Acetate	8.17	43	51491	1.524	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN122418\
Data File : VN053195.D
Acq On : 24 Dec 2018 19:40
Operator : MD\SY
Sample : J6511-20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
SB-65(0-5)-GRAB

Quant Time: Dec 25 00:47:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 18 13:03:37 2018
Response via : Initial Calibration

