

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN122418\
 Data File : VN053194.D
 Acq On : 24 Dec 2018 19:13
 Operator : MD\SY
 Sample : J6511-19
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-64(0-5)-GRAB

Quant Time: Dec 25 00:43:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 18 13:03:37 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1264565	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1977764	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1732469	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	647195	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	641144	49.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.92%	
35) Dibromofluoromethane	7.59	113	548072	60.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.32%	
50) Toluene-d8	10.09	98	2404830	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
62) 4-Bromofluorobenzene	12.40	95	754112	44.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.58%	
Target Compounds						
16) Acetone	3.82	43	18083	7.187	ug/l	91
18) Methyl Acetate	4.33	43	5461	1.461	ug/l #	78
20) Methylene Chloride	4.55	84	34747	2.532	ug/l	93
43) Isopropyl Acetate	8.17	43	53737	1.609	ug/l #	85


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN122418\
Data File : VN053194.D
Acq On : 24 Dec 2018 19:13
Operator : MD\SY
Sample : J6511-19
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
SB-64(0-5)-GRAB

Quant Time: Dec 25 00:43:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 18 13:03:37 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. -0.00 min

Lab File: VN053194.D

Acq: 24 Dec 2018 19:13

Instrument :

MSVOA_N

ClientSampleId :


SB-64(0-5)-GRAB

Tot Ion:168 Resp: 1264565

Ion Ratio Lower Upper

168 100

99 49.7 40.2 60.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN

7.67

600000

500000

400000



300000

200000

100000

0

Time-->



#16

Acetone

Concen: 7.187 ug/l

RT: 3.82 min Scan# 638

Delta R.T. 0.01 min

Lab File: VN053194.D


Acq: 24 Dec 2018 19:13

Tgt Ion: 43 Resp: 18083

Ion Ratio Lower Upper

43 100

58 36.9 25.4 38.0



Abundance Ion 43.00 (42.70 to 43.70): VN

Ion 58.00 (57.70 to 58.70): VN

3.82


6000


4000

2000

0

Time-->









#18
Methyl Acetate
Concen: 1.461 ug/l
RT: 4.33 min Scan# 796
Delta R.T. 0.01 min
Lab File: VN053194.D
Acq: 24 Dec 2018 19:13




Instrument :
MSVOA_N
ClientSampleId :
SB-64(0-5)-GRAB


Tot Ion: 43 Resp: 5461
Ion Ratio Lower Upper
43 100
74 13.0 19.2 28.8#



#20
Methylene Chloride
Concen: 2.532 ug/l
RT: 4.55 min Scan# 865
Delta R.T. -0.00 min
Lab File: VN053194.D
Acq: 24 Dec 2018 19:13

Tgt Ion: 84 Resp: 34747
Ion Ratio Lower Upper
84 100
49 111.2 96.8 145.2
51 33.4 29.9 44.9
86 62.6 51.8 77.6









#33
 1,2-Dichloroethane-d4
 Concen: Below ug/l
 RT: 8.03 min Scan# 1945
 Delta R.T. -0.00 min
 Lab File: VN053194.D
 Acq: 24 Dec 2018 19:13




Instrument :
 MSVOA_N
 ClientSampleId :
 SB-64(0-5)-GRAB

Tot Ion: 65 Resp: 641144
 Ion Ratio Lower Upper
 65 100
 67 55.4 0.0 110.4



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN053194.D
 Acq: 24 Dec 2018 19:13

Tgt Ion: 114 Resp: 1977764
 Ion Ratio Lower Upper
 114 100
 63 19.7 0.0 39.8
 88 15.1 0.0 31.0









#35
 Dibromofluoromethane
 Concen: 50.000 ug/l
 RT: 7.59 min Scan# 1810
 Delta R.T. 0.00 min
 Lab File: VN053194.D
 Acq: 24 Dec 2018 19:13




Instrument :
 MSVOA_N
 ClientSampleId :
 SB-64(0-5)-GRAB


Tot Ion	Ratio	Lower	Upper
113	100		
111	102.1	83.0	124.4
192	22.6	17.5	26.3



#43
 Isopropyl Acetate
 Concen: 1.609 ug/l
 RT: 8.17 min Scan# 1989
 Delta R.T. 0.01 min
 Lab File: VN053194.D
 Acq: 24 Dec 2018 19:13

Tgt Ion	Ratio	Lower	Upper
43	100		
61	17.8	22.9	34.3#
87	11.9	10.6	15.8





#50

Toluene-d8

Concen: N.D. ug/l

RT: 10.09 min Scan# 2587

Delta R.T. -0.00 min

Lab File: VN053194.D

Acq: 24 Dec 2018 19:13

Instrument :

MSVOA_N

ClientSampleId :





SB-64(0-5)-GRAB

Tot Ion: 98 Resp: 2404830

Ion Ratio Lower Upper

98 100

100 64.0 51.6 77.4



#62

4-Bromofluorobenzene

Concen: N.D. ug/l

RT: 12.40 min Scan# 3305

Delta R.T. -0.00 min

Lab File: VN053194.D

Acq: 24 Dec 2018 19:13




Tgt Ion: 95 Resp: 754112


Ion Ratio Lower Upper

95 100

174 89.1 0.0 178.8

176 86.0 0.0 173.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. -0.00 min
Lab File: VN053194.D
Acq: 24 Dec 2018 19:13

Instrument :
MSVOA_N
ClientSampled :
SB-64(0-5)-GRAB

Tot Ion:117 Resp: 1732469
Ion Ratio Lower Upper
117 100
82 52.7 42.8 64.2
119 32.0 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.35 min Scan# 3599
Delta R.T. 0.00 min
Lab File: VN053194.D
Acq: 24 Dec 2018 19:13

Tgt Ion:152 Resp: 647195
Ion Ratio Lower Upper
152 100
115 56.7 28.3 85.0
150 158.4 0.0 347.8

