

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122420\
 Data File : VN065270.D
 Acq On : 24 Dec 2020 18:47
 Operator : JC/MD
 Sample : L5159-10DL 250X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Dec 25 02:41:24 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	252858	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	414783	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	435513	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	192374	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	145751	51.06	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 102.12%			
35) Dibromofluoromethane	7.553	113	130117	52.00	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 104.00%			
50) Toluene-d8	10.061	98	509651	54.34	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 108.68%			
62) 4-Bromofluorobenzene	12.376	95	207138	55.25	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 110.50%			
Target Compounds						Qvalue
3) Chloromethane	2.145	50	572	0.21	ug/l #	41
11) Tert butyl alcohol	4.782	59	8910	19.08	ug/l #	91
13) Acrolein	3.586	56	186	2.99	ug/l #	14
16) Acetone	3.788	43	20629	6.75	ug/l	93
18) Methyl Acetate	4.290	43	5876	2.29	ug/l #	80
19) Methyl tert-butyl Ether	5.000	73	4724	0.60	ug/l	91
20) Methylene Chloride	4.505	84	1009	0.36	ug/l #	79
22) Diisopropyl ether	5.910	45	24254	2.98	ug/l #	1
25) 2-Butanone	6.814	43	7832	4.89	ug/l	100
31) Cyclohexane	7.624	56	5935	1.41	ug/l #	26
36) 1,1-Dichloropropene	7.624	75	16282	4.32	ug/l #	50
37) Ethyl Acetate	6.814	43	7832	2.10	ug/l #	66
38) Carbon Tetrachloride	7.624	117	21893	5.53	ug/l #	16
39) Methylcyclohexane	9.045	83	1171	0.26	ug/l #	84
40) Benzene	8.003	78	647610	57.72	ug/l	98
41) Methacrylonitrile	7.180	41	632	0.32	ug/l #	15
42) 1,2-Dichloroethane	8.003	62	5903	1.42	ug/l #	73
43) Isopropyl Acetate	8.235	43	10747	1.67	ug/l #	59
49) 1,4-Dioxane	9.174	88	161	3.18	ug/l #	26
51) 4-Methyl-2-Pentanone	9.955	43	1893	0.52	ug/l #	82
52) Toluene	10.126	92	229712	32.48	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.778	63	599	0.32	ug/l #	46
59) 2-Hexanone	10.733	43	1474	0.55	ug/l	63
67) Ethyl Benzene	11.486	91	11252	0.70	ug/l	99
68) m/p-Xylenes	11.592	106	25083	4.17	ug/l	98
69) o-Xylene	11.923	106	9402	1.61	ug/l	99
71) Bromoform	12.373	173	1076	0.36	ug/l #	1
74) N-amyl acetate	11.971	43	3668	0.63	ug/l #	1
76) 1,2,3-Trichloropropane	12.376	75	104251	26.56	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.707	105	6008	0.48	ug/l	94
81) trans-1,4-Dichloro-2-b...	12.376	75	104251	66.98	ug/l #	15
84) 1,2,4-Trimethylbenzene	13.016	105	10375	0.85	ug/l	99
85) sec-Butylbenzene	13.016	105	10375	0.73	ug/l #	56
95) Naphthalene	15.103	128	567	3.43	ug/l #	69

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122420\
Data File : VN065270.D
Acq On : 24 Dec 2020 18:47
Operator : JC/MD
Sample : L5159-10DL 250X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Dec 25 02:41:24 2020
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#)=qualifier out of range (m)=manual integration (+)=signals summed						

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122420\
Data File : VN065270.D
Acq On : 24 Dec 2020 18:47
Operator : JC/MD
Sample : L5159-10DL 250X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Dec 25 02:41:24 2020
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.627 min Scan# 1831
Delta R.T. 0.000 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:168 Resp: 252858
Ion Ratio Lower Upper
168 100
99 49.4 43.0 64.4



#3
Chloromethane
Concen: 0.21 ug/l
RT: 2.145 min Scan# 126
Delta R.T. 0.099 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47


Tgt Ion: 50 Resp: 572
Ion Ratio Lower Upper
50 100
52 0.0 26.6 39.8#








#11
Tert butyl alcohol
Concen: 19.08 ug/l
RT: 4.782 min Scan# 946
Delta R.T. 0.010 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47

Instrument :
MSVOA_N
ClientSampled :







Tgt Ion: 59 Resp: 8910
Ion Ratio Lower Upper
59 100
57 13.1 7.9 11.9#



#13
Acrolein
Concen: 2.99 ug/l
RT: 3.586 min Scan# 574
Delta R.T. 0.013 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47

Tgt Ion: 56 Resp: 186
Ion Ratio Lower Upper
56 100
55 0.0 56.5 84.7#









#16
Acetone
Concen: 6.75 ug/l
RT: 3.788 min Scan# 637
Delta R.T. 0.003 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 20629
Ion Ratio Lower Upper
43 100
58 36.0 25.7 38.5



#18
Methyl Acetate
Concen: 2.29 ug/l
RT: 4.290 min Scan# 793
Delta R.T. 0.003 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47

Tgt Ion: 43 Resp: 5876
Ion Ratio Lower Upper
43 100
74 14.9 20.2 30.2#









#19
Methyl tert-butyl Ether
Concen: 0.60 ug/l
RT: 5.000 min Scan# 1014
Delta R.T. -0.004 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 73 Resp: 4724
Ion Ratio Lower Upper
73 100
57 25.8 17.4 26.0



#20
Methylene Chloride
Concen: 0.36 ug/l
RT: 4.505 min Scan# 860
Delta R.T. 0.003 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47


Tgt Ion: 84 Resp: 1009
Ion Ratio Lower Upper
84 100
49 81.5 93.4 140.0#
51 40.4 28.2 42.4
86 69.8 51.7 77.5








#22
Diisopropyl ether
Concen: 2.98 ug/l
RT: 5.910 min Scan# 1297
Delta R.T. -0.004 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47

Instrument :
MSVOA_N
ClientSampled :







Tgt Ion	Ratio	Lower	Upper
45	100		
43	7.6	46.2	69.2#
87	0.0	24.1	36.1#
59	202.8	9.0	13.4#



#25
2-Butanone
Concen: 4.89 ug/l
RT: 6.814 min Scan# 1578
Delta R.T. 0.016 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47


Tgt Ion	Ratio	Lower	Upper
43	100		
72	26.1	21.0	31.6








#31
Cyclohexane
Concen: 1.41 ug/l
RT: 7.624 min Scan# 1830
Delta R.T. 0.013 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 56 Resp: 5935
Ion Ratio Lower Upper
56 100
69 140.0 27.5 41.3#
84 123.9 73.9 110.9#



#33
1,2-Dichloroethane-d4
Concen: 51.06 ug/l
RT: 7.991 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47

Tgt Ion: 65 Resp: 145751
Ion Ratio Lower Upper
65 100
67 52.1 0.0 108.2









#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.553 min Scan# 2119
Delta R.T. 0.003 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:114	Resp: 414783
Ion Ratio	Lower Upper
114 100	
63 19.5	0.0 38.6
88 15.2	0.0 32.6



#35
Dibromofluoromethane
Concen: 52.00 ug/l
RT: 7.553 min Scan# 1808
Delta R.T. 0.003 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47

Tgt Ion:113	Resp: 130117
Ion Ratio	Lower Upper
113 100	
111 101.8	81.6 122.4
192 23.2	16.2 24.2









#36
1,1-Dichloropropene
Concen: 4.32 ug/l
RT: 7.624 min Scan# 1830
Delta R.T. -0.129 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 75 Resp: 16282
Ion Ratio Lower Upper
75 100
110 9.4 17.8 53.3#
77 0.0 25.0 37.4#



#37
Ethyl Acetate
Concen: 2.10 ug/l
RT: 6.814 min Scan# 1578
Delta R.T. -0.077 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47


Tgt Ion: 43 Resp: 7832
Ion Ratio Lower Upper
43 100
61 0.0 12.8 19.2#
70 0.0 8.8 13.2#








#38
Carbon Tetrachloride
Concen: 5.53 ug/l
RT: 7.624 min Scan# 1830
Delta R.T. -0.109 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 117 Resp: 21893
Ion Ratio Lower Upper
117 100
119 4.9 76.7 115.1#
121 0.0 25.4 38.2#



#39
Methylcyclohexane
Concen: 0.26 ug/l
RT: 9.045 min Scan# 2272
Delta R.T. 0.000 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47

Tgt Ion: 83 Resp: 1171
Ion Ratio Lower Upper
83 100
55 61.5 60.2 90.2
98 57.8 37.4 56.2#









#40
Benzene
Concen: 57.72 ug/l
RT: 8.003 min Scan# 1948
Delta R.T. -0.001 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 78 Resp: 647610
Ion Ratio Lower Upper
78 100
77 23.6 17.9 26.9



#41
Methacrylonitrile
Concen: 0.32 ug/l
RT: 7.180 min Scan# 1692
Delta R.T. 0.051 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47

Tgt Ion: 41 Resp: 632
Ion Ratio Lower Upper
41 100
39 4.7 51.0 76.6#
67 0.0 79.8 119.8#
52 0.0 31.5 47.3#





#42

1,2-Dichloroethane

Concen: 1.42 ug/l

RT: 8.003 min Scan# 1948

Delta R.T. -0.084 min

Lab File: VN065270.D





Acq: 24 Dec 2020 18:47

Tgt Ion: 62 Resp: 5903

Ion Ratio Lower Upper

62 100

98 0.0 0.0 19.6



#43

Isopropyl Acetate

Concen: 1.67 ug/l

RT: 8.235 min Scan# 2020

Delta R.T. 0.103 min

Lab File: VN065270.D

Acq: 24 Dec 2020 18:47




Tgt Ion: 43 Resp: 10747


Ion Ratio Lower Upper

43 100

61 0.0 17.2 25.8#

87 0.0 11.4 17.0#









#49
1,4-Dioxane
Concen: 3.18 ug/l
RT: 9.174 min Scan# 2312
Delta R.T. 0.003 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 88 Resp: 161
Ion Ratio Lower Upper
88 100
43 0.0 24.2 36.4#
58 0.0 52.5 78.7#



#50
Toluene-d8
Concen: 54.34 ug/l
RT: 10.061 min Scan# 2588
Delta R.T. 0.003 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47

Tgt Ion: 98 Resp: 509651
Ion Ratio Lower Upper
98 100
100 64.3 51.7 77.5









#51
4-Methyl-2-Pentanone
Concen: 0.52 ug/l
RT: 9.955 min Scan# 2555
Delta R.T. 0.000 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 1893
Ion Ratio Lower Upper
43 100
58 29.5 32.4 48.6#



#52
Toluene
Concen: 32.48 ug/l
RT: 10.126 min Scan# 2608
Delta R.T. -0.000 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47

Tgt Ion: 92 Resp: 229712
Ion Ratio Lower Upper
92 100
91 174.4 139.1 208.7









#58
2-Chloroethyl Vinyl ether
Concen: 0.32 ug/l
RT: 9.778 min Scan# 2500
Delta R.T. 0.115 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 63 Resp: 599
Ion Ratio Lower Upper
63 100
106 0.0 22.7 34.1#



#59
2-Hexanone
Concen: 0.55 ug/l
RT: 10.733 min Scan# 2797
Delta R.T. 0.009 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47

Tgt Ion: 43 Resp: 1474
Ion Ratio Lower Upper
43 100
58 29.4 28.3 84.9









#62
4-Bromofluorobenzene
Concen: 55.25 ug/l
RT: 12.376 min Scan# 3308
Delta R.T. 0.000 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 95 Resp: 207138
Ion Ratio Lower Upper
95 100
174 91.9 0.0 160.4
176 90.2 0.0 158.4



#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.380 min Scan# 2998
Delta R.T. -0.000 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47

Tgt Ion: 117 Resp: 435513
Ion Ratio Lower Upper
117 100
82 53.5 44.2 66.2
119 32.3 25.8 38.6









#67
Ethyl Benzene
Concen: 0.70 ug/l
RT: 11.486 min Scan# 3031
Delta R.T. 0.004 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 91 Resp: 11252
Ion Ratio Lower Upper
91 100
106 30.9 25.2 37.8



#68
m/p-Xylenes
Concen: 4.17 ug/l
RT: 11.592 min Scan# 3064
Delta R.T. -0.000 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47

Tgt Ion:106 Resp: 25083
Ion Ratio Lower Upper
106 100
91 204.7 161.0 241.4









#69
o-Xylene
Concen: 1.61 ug/l
RT: 11.923 min Scan# 3167
Delta R.T. 0.003 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:106 Resp: 9402
Ion Ratio Lower Upper
106 100
91 212.6 105.5 316.5



#71
Bromoform
Concen: 0.36 ug/l
RT: 12.373 min Scan# 3307
Delta R.T. 0.273 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47

Tgt Ion:173 Resp: 1076
Ion Ratio Lower Upper
173 100
175 1028.1 24.3 72.8#
254 0.0 0.0 0.0









#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.318 min Scan# 3601
Delta R.T. 0.000 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:152 Resp: 192374
Ion Ratio Lower Upper
152 100
115 55.4 29.7 89.1
150 158.4 0.0 350.6



#74
N-ethyl acetate
Concen: 0.63 ug/l
RT: 11.971 min Scan# 3182
Delta R.T. -0.071 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47

Tgt Ion: 43 Resp: 3668
Ion Ratio Lower Upper
43 100
70 0.0 34.5 51.7#
55 205.3 20.4 30.6#
61 0.0 20.0 30.0#









#76
1,2,3-Trichloropropane
Concen: 26.56 ug/l
RT: 12.376 min Scan# 3308
Delta R.T. -0.151 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 75 Resp: 104251
Ion Ratio Lower Upper
75 100
77 1.1 0.0 0.0#



#80
1,3,5-Trimethylbenzene
Concen: 0.48 ug/l
RT: 12.707 min Scan# 3411
Delta R.T. -0.001 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47


Tgt Ion: 105 Resp: 6008
Ion Ratio Lower Upper
105 100
120 53.3 24.6 73.6








#81
trans-1,4-Dichloro-2-butene
Concen: 66.98 ug/l
RT: 12.376 min Scan# 3308
Delta R.T. 0.103 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47


Instrument :
MSVOA_N
ClientSampled :





Tgt Ion: 75 Resp: 104251
Ion Ratio Lower Upper
75 100
53 0.4 67.0 100.4#
89 0.0 39.6 59.4#




#84
1,2,4-Trimethylbenzene
Concen: 0.85 ug/l
RT: 13.016 min Scan# 3507
Delta R.T. 0.003 min
Lab File: VN065270.D
Acq: 24 Dec 2020 18:47



Tgt Ion: 105 Resp: 10375
Ion Ratio Lower Upper
105 100
120 45.3 23.1 69.2









#85
 sec-Butylbenzene
 Concen: 0.73 ug/l
 RT: 13.016 min Scan# 3507
 Delta R.T. -0.129 min
 Lab File: VN065270.D
 Acq: 24 Dec 2020 18:47

Instrument :
 MSVOA_N
 ClientSampleId :

Tgt Ion:105 Resp: 10375
 Ion Ratio Lower Upper
 105 100
 134 0.0 10.1 30.1#



#95
 Naphthalene
 Concen: 3.43 ug/l
 RT: 15.103 min Scan# 4156
 Delta R.T. 0.007 min
 Lab File: VN065270.D
 Acq: 24 Dec 2020 18:47

Tgt Ion:128 Resp: 567
 Ion Ratio Lower Upper
 128 100
 127 0.0 10.3 15.5#
 129 0.0 8.7 13.1#

