

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122420\
 Data File : VN065250.D
 Acq On : 24 Dec 2020 9:50
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 25 02:33:26 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
 Quant Title : Sw846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	50.000	49.950	0.1	96	0.00
3 P	Chloromethane	50.000	52.436	-4.9	103	0.00
4 C	Vinyl Chloride	50.000	50.893	-1.8#	98	0.00
5 T	Bromomethane	50.000	50.585	-1.2	101	0.00
6 T	Chloroethane	50.000	49.304	1.4	96	0.00
7 T	Trichlorofluoromethane	50.000	47.291	5.4	93	0.00
8 T	Diethyl Ether	50.000	49.538	0.9	101	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.554	2.9	98	0.00
10 T	Methyl Iodide	50.000	49.080	1.8	94	0.00
11 T	Tert butyl alcohol	250.000	240.500	3.8	95	0.00
12 CM	1,1-Dichloroethene	50.000	46.059	7.9#	95	0.00
13 T	Acrolein	250.000	208.804	16.5	83	0.00
14 T	Allyl chloride	50.000	49.490	1.0	98	0.00
15 T	Acrylonitrile	250.000	261.288	-4.5	102	0.00
16 T	Acetone	250.000	262.847	-5.1	104	0.00
17 T	Carbon Disulfide	50.000	50.536	-1.1	98	0.00
18 T	Methyl Acetate	50.000	51.297	-2.6	101	0.00
19 T	Methyl tert-butyl Ether	50.000	50.705	-1.4	102	0.00
20 T	Methylene Chloride	50.000	47.481	5.0	103	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.789	-1.6	101	0.00
22 T	Diisopropyl ether	50.000	50.774	-1.5	103	0.00
23 T	Vinyl Acetate	250.000	258.419	-3.4	101	0.00
24 P	1,1-Dichloroethane	50.000	51.144	-2.3	101	0.00
25 T	2-Butanone	250.000	258.750	-3.5	101	0.00
26 T	2,2-Dichloropropane	50.000	48.076	3.8	98	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.465	-2.9	102	0.00
28 T	Bromochloromethane	50.000	47.857	4.3	104	0.00
29 T	Tetrahydrofuran	250.000	245.788	1.7	99	0.00
30 C	Chloroform	50.000	51.119	-2.2#	102	0.00
31 T	Cyclohexane	50.000	46.575	6.8	97	0.00
32 T	1,1,1-Trichloroethane	50.000	49.189	1.6	98	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.855	2.3	98	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	99	0.00
35 S	Dibromofluoromethane	50.000	51.163	-2.3	100	0.00
36 T	1,1-Dichloropropene	50.000	51.103	-2.2	100	0.00
37 T	Ethyl Acetate	50.000	51.833	-3.7	100	0.00
38 T	Carbon Tetrachloride	50.000	49.357	1.3	96	0.00
39 T	Methylcyclohexane	50.000	51.768	-3.5	99	0.00
40 TM	Benzene	50.000	51.880	-3.8	102	0.00
41 T	Methacrylonitrile	50.000	47.858	4.3	98	0.00
42 TM	1,2-Dichloroethane	50.000	51.509	-3.0	102	0.00
43 T	Isopropyl Acetate	50.000	51.345	-2.7	98	0.00
44 TM	Trichloroethene	50.000	51.911	-3.8	101	0.00
45 C	1,2-Dichloropropane	50.000	53.091	-6.2#	103	0.00
46 T	Dibromomethane	50.000	53.007	-6.0	104	0.00
47 T	Bromodichloromethane	50.000	52.512	-5.0	102	0.00
48 T	Methyl methacrylate	50.000	52.448	-4.9	102	0.00
49 T	1,4-Dioxane	1000.000	1043.682	-4.4	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122420\
 Data File : VN065250.D
 Acq On : 24 Dec 2020 9:50
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 25 02:33:26 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
 Quant Title : Sw846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

50 S	Toluene-d8	50.000	50.728	-1.5	99	0.00
51 T	4-Methyl-2-Pentanone	250.000	257.090	-2.8	98	0.00
52 CM	Toluene	50.000	52.102	-4.2#	101	0.00
53 T	t-1,3-Dichloropropene	50.000	51.962	-3.9	101	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.816	-3.6	102	0.00
55 T	1,1,2-Trichloroethane	50.000	52.895	-5.8	102	0.00
56 T	Ethyl methacrylate	50.000	51.990	-4.0	101	0.00
57 T	1,3-Dichloropropane	50.000	53.530	-7.1	103	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	253.460	-1.4	100	0.00
59 T	2-Hexanone	250.000	259.425	-3.8	98	0.00
60 T	Dibromochloromethane	50.000	52.573	-5.1	102	0.00
61 T	1,2-Dibromoethane	50.000	52.730	-5.5	102	0.00
62 S	4-Bromofluorobenzene	50.000	47.618	4.8	99	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	100	0.00
64 T	Tetrachloroethene	50.000	49.510	1.0	101	0.00
65 PM	Chlorobenzene	50.000	51.838	-3.7	102	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.499	-3.0	101	0.00
67 C	Ethyl Benzene	50.000	51.166	-2.3#	101	0.00
68 T	m/p-Xylenes	100.000	103.831	-3.8	102	0.00
69 T	o-Xylene	50.000	51.521	-3.0	101	0.00
70 T	Styrene	50.000	52.480	-5.0	102	0.00
71 P	Bromoform	50.000	51.944	-3.9	100	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	101	0.00
73 T	Isopropylbenzene	50.000	48.162	3.7	100	0.00
74 T	N-aryl acetate	50.000	49.675	0.7	98	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.042	-0.1	103	0.00
76 T	1,2,3-Trichloropropane	50.000	49.247	1.5	106	0.00
77 T	Bromobenzene	50.000	50.015	-0.0	103	0.00
78 T	n-propylbenzene	50.000	48.603	2.8	100	0.00
79 T	2-Chlorotoluene	50.000	48.608	2.8	99	0.00
80 T	1,3,5-Trimethylbenzene	50.000	48.515	3.0	101	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.673	0.7	100	0.00
82 T	4-Chlorotoluene	50.000	49.771	0.5	102	0.00
83 T	tert-Butylbenzene	50.000	48.703	2.6	101	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.054	-0.1	103	0.00
85 T	sec-Butylbenzene	50.000	48.619	2.8	101	0.00
86 T	p-Isopropyltoluene	50.000	49.586	0.8	102	0.00
87 T	1,3-Dichlorobenzene	50.000	51.473	-2.9	104	0.00
88 T	1,4-Dichlorobenzene	50.000	50.840	-1.7	103	0.00
89 T	n-Butylbenzene	50.000	50.535	-1.1	102	0.00
90 T	Hexachloroethane	50.000	50.533	-1.1	100	0.00
91 T	1,2-Dichlorobenzene	50.000	51.770	-3.5	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.746	-1.5	99	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.369	-0.7	107	0.00
94 T	Hexachlorobutadiene	50.000	48.440	3.1	104	0.00
95 T	Naphthalene	50.000	48.561	2.9	105	0.00
96 T	1,2,3-Trichlorobenzene	50.000	48.746	2.5	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122420\
Data File : VN065250.D
Acq On : 24 Dec 2020 9:50
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Dec 25 02:33:26 2020
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
Quant Title : Sw846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6