

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122420\
 Data File : VN065272.D
 Acq On : 24 Dec 2020 19:36
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 25 02:42:12 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
 Quant Title : Sw846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	50.000	49.802	0.4	95	0.00
3 P	Chloromethane	50.000	49.085	1.8	96	0.00
4 C	Vinyl Chloride	50.000	49.270	1.5#	95	0.00
5 T	Bromomethane	50.000	48.906	2.2	97	0.00
6 T	Chloroethane	50.000	49.592	0.8	96	0.00
7 T	Trichlorofluoromethane	50.000	49.202	1.6	97	0.00
8 T	Diethyl Ether	50.000	48.794	2.4	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.532	2.9	98	0.00
10 T	Methyl Iodide	50.000	52.686	-5.4	102	0.00
11 T	Tert butyl alcohol	250.000	255.419	-2.2	101	0.00
12 CM	1,1-Dichloroethene	50.000	47.467	5.1#	98	0.00
13 T	Acrolein	250.000	217.773	12.9	87	0.00
14 T	Allyl chloride	50.000	47.815	4.4	95	0.00
15 T	Acrylonitrile	250.000	261.990	-4.8	102	0.00
16 T	Acetone	250.000	216.391	13.4	86	0.00
17 T	Carbon Disulfide	50.000	49.300	1.4	96	0.00
18 T	Methyl Acetate	50.000	52.676	-5.4	104	0.00
19 T	Methyl tert-butyl Ether	50.000	49.797	0.4	100	0.00
20 T	Methylene Chloride	50.000	45.903	8.2	99	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.810	0.4	99	0.00
22 T	Diisopropyl ether	50.000	48.593	2.8	99	0.00
23 T	Vinyl Acetate	250.000	251.851	-0.7	98	0.00
24 P	1,1-Dichloroethane	50.000	49.475	1.0	98	0.00
25 T	2-Butanone	250.000	250.483	-0.2	98	0.00
26 T	2,2-Dichloropropane	50.000	41.439	17.1	84	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.328	-0.7	100	0.00
28 T	Bromochloromethane	50.000	45.788	8.4	99	0.00
29 T	Tetrahydrofuran	250.000	253.552	-1.4	102	0.00
30 C	Chloroform	50.000	50.317	-0.6#	100	0.00
31 T	Cyclohexane	50.000	45.153	9.7	94	0.00
32 T	1,1,1-Trichloroethane	50.000	49.904	0.2	99	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.672	0.7	100	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	102	0.00
35 S	Dibromofluoromethane	50.000	50.860	-1.7	102	0.00
36 T	1,1-Dichloropropene	50.000	48.323	3.4	97	0.00
37 T	Ethyl Acetate	50.000	50.822	-1.6	101	0.00
38 T	Carbon Tetrachloride	50.000	49.513	1.0	100	0.00
39 T	Methylcyclohexane	50.000	47.083	5.8	93	0.00
40 TM	Benzene	50.000	49.337	1.3	100	0.00
41 T	Methacrylonitrile	50.000	45.375	9.3	96	0.00
42 TM	1,2-Dichloroethane	50.000	49.218	1.6	101	0.00
43 T	Isopropyl Acetate	50.000	50.062	-0.1	99	0.00
44 TM	Trichloroethene	50.000	49.386	1.2	99	0.00
45 C	1,2-Dichloropropane	50.000	49.372	1.3#	99	0.00
46 T	Dibromomethane	50.000	50.113	-0.2	101	0.00
47 T	Bromodichloromethane	50.000	49.872	0.3	100	0.00
48 T	Methyl methacrylate	50.000	50.187	-0.4	101	0.00
49 T	1,4-Dioxane	1000.000	1068.527	-6.9	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122420\
 Data File : VN065272.D
 Acq On : 24 Dec 2020 19:36
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 25 02:42:12 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
 Quant Title : Sw846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

50 S	Toluene-d8	50.000	49.579	0.8	100	0.00
51 T	4-Methyl-2-Pentanone	250.000	257.311	-2.9	102	0.00
52 CM	Toluene	50.000	50.009	-0.0#	100	0.00
53 T	t-1,3-Dichloropropene	50.000	47.682	4.6	95	0.00
54 T	cis-1,3-Dichloropropene	50.000	47.680	4.6	97	0.00
55 T	1,1,2-Trichloroethane	50.000	50.570	-1.1	101	0.00
56 T	Ethyl methacrylate	50.000	50.334	-0.7	101	0.00
57 T	1,3-Dichloropropane	50.000	50.247	-0.5	100	0.00
58 T	2-Chloroethyl vinyl ether	250.000	241.548	3.4	98	0.00
59 T	2-Hexanone	250.000	258.284	-3.3	101	0.00
60 T	Dibromochloromethane	50.000	50.846	-1.7	102	0.00
61 T	1,2-Dibromoethane	50.000	50.938	-1.9	102	0.00
62 S	4-Bromofluorobenzene	50.000	47.046	5.9	101	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	103	0.00
64 T	Tetrachloroethene	50.000	48.034	3.9	101	0.00
65 PM	Chlorobenzene	50.000	49.481	1.0	100	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.610	-1.2	102	0.00
67 C	Ethyl Benzene	50.000	49.192	1.6#	100	0.00
68 T	m/p-Xylenes	100.000	98.943	1.1	100	0.00
69 T	o-Xylene	50.000	49.598	0.8	100	0.00
70 T	Styrene	50.000	50.413	-0.8	100	0.00
71 P	Bromoform	50.000	51.686	-3.4	102	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	104	0.00
73 T	Isopropylbenzene	50.000	46.443	7.1	99	0.00
74 T	N-aryl acetate	50.000	49.041	1.9	100	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.412	1.2	105	0.00
76 T	1,2,3-Trichloropropane	50.000	49.404	1.2	109	0.00
77 T	Bromobenzene	50.000	48.868	2.3	103	0.00
78 T	n-propylbenzene	50.000	45.821	8.4	97	0.00
79 T	2-Chlorotoluene	50.000	46.557	6.9	98	0.00
80 T	1,3,5-Trimethylbenzene	50.000	46.346	7.3	99	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	45.196	9.6	93	0.00
82 T	4-Chlorotoluene	50.000	46.935	6.1	99	0.00
83 T	tert-Butylbenzene	50.000	46.548	6.9	99	0.00
84 T	1,2,4-Trimethylbenzene	50.000	47.124	5.8	99	0.00
85 T	sec-Butylbenzene	50.000	45.774	8.5	97	0.00
86 T	p-Isopropyltoluene	50.000	45.815	8.4	97	0.00
87 T	1,3-Dichlorobenzene	50.000	49.203	1.6	102	0.00
88 T	1,4-Dichlorobenzene	50.000	48.327	3.3	100	0.00
89 T	n-Butylbenzene	50.000	44.603	10.8	92	0.00
90 T	Hexachloroethane	50.000	48.274	3.5	98	0.00
91 T	1,2-Dichlorobenzene	50.000	50.086	-0.2	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	53.285	-6.6	106	0.00
93 T	1,2,4-Trichlorobenzene	50.000	47.394	5.2	103	0.00
94 T	Hexachlorobutadiene	50.000	42.509	15.0	93	0.00
95 T	Naphthalene	50.000	48.603	2.8	108	0.00
96 T	1,2,3-Trichlorobenzene	50.000	47.657	4.7	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122420\
Data File : VN065272.D
Acq On : 24 Dec 2020 19:36
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Dec 25 02:42:12 2020
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
Quant Title : Sw846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6