

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122624\
 Data File : VN085311.D
 Acq On : 26 Dec 2024 11:13
 Operator : JC\MD
 Sample : VN1226WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1226WBL01

Quant Time: Dec 27 00:42:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121824W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 19 06:31:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	159396	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	294503	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.859	117	251735	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	102404	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	123859	44.458	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	88.920%	
35) Dibromofluoromethane	8.165	113	95259	46.008	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	92.020%	
50) Toluene-d8	10.559	98	357973	47.303	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	94.600%	
62) 4-Bromofluorobenzene	12.847	95	118525	43.850	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	87.700%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122624\
Data File : VN085311.D
Acq On : 26 Dec 2024 11:13
Operator : JC\MD
Sample : VN1226WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1226WBL01

Quant Time: Dec 27 00:42:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121824W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 19 06:31:00 2024
Response via : Initial Calibration

