

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122624\
 Data File : VN085315.D
 Acq On : 26 Dec 2024 14:12
 Operator : JC\MD
 Sample : P5371-06RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPLP-C11-3-MRE

Quant Time: Dec 27 00:45:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121824W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 19 06:31:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	152126	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	275786	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	242276	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	109298	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	114507	43.065	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	86.140%		
35) Dibromofluoromethane	8.165	113	63325	32.660	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	65.320%#		
50) Toluene-d8	10.565	98	339611	47.922	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.840%		
62) 4-Bromofluorobenzene	12.847	95	124635	49.240	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	98.480%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	21792	23.744	ug/l	94
27) cis-1,2-Dichloroethene	7.483	96	6743	3.025	ug/l	79
30) Chloroform	7.965	83	11768	2.965	ug/l	99
40) Benzene	8.606	78	40115	4.776	ug/l	96
52) Toluene	10.630	92	35830	7.095	ug/l	99
67) Ethyl Benzene	11.959	91	396222	42.447	ug/l	100
68) m/p-Xylenes	12.065	106	39601	11.343	ug/l	99
69) o-Xylene	12.394	106	37171	11.327	ug/l	99
73) Isopropylbenzene	12.694	105	31660	4.015	ug/l	99
78) n-propylbenzene	13.035	91	15233	1.575	ug/l	96
80) 1,3,5-Trimethylbenzene	13.171	105	24697	3.755	ug/l	94
84) 1,2,4-Trimethylbenzene	13.476	105	98233	14.613	ug/l	99
86) p-Isopropyltoluene	13.723	119	5582	2.660	ug/l #	72
95) Naphthalene	15.635	128	4350456	723.774	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122624\
Data File : VN085315.D
Acq On : 26 Dec 2024 14:12
Operator : JC\MD
Sample : P5371-06RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPLP-C11-3-MRE

Quant Time: Dec 27 00:45:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121824W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 19 06:31:00 2024
Response via : Initial Calibration

