

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN122718\
 Data File : VN053204.D
 Acq On : 27 Dec 2018 11:30
 Operator : MD\SY
 Sample : J6571-03 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 30061

Quant Time: Dec 28 01:15:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 18 13:03:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1193292	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1832570	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1614934	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	615519	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	587147	48.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.98%	
35) Dibromofluoromethane	7.59	113	517172	61.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.56%	
50) Toluene-d8	10.09	98	2237397	49.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.58%	
62) 4-Bromofluorobenzene	12.40	95	722541	45.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.60%	
Target Compounds						
16) Acetone	3.82	43	58823	24.775	ug/l #	72
20) Methylene Chloride	4.55	84	6894	0.532	ug/l	93
25) 2-Butanone	6.85	43	4198	1.054	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN122718\
Data File : VN053204.D
Acq On : 27 Dec 2018 11:30
Operator : MD\SY
Sample : J6571-03 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
30061

Quant Time: Dec 28 01:15:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 18 13:03:37 2018
Response via : Initial Calibration

