

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122722\
 Data File : VN075978.D
 Acq On : 27 Dec 2022 10:59
 Operator : JC\MD
 Sample : VN1227WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1227WBL01

Quant Time: Dec 28 04:43:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	343642	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	537903	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	463695	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	193976	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	174018	46.547	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.100%
35) Dibromofluoromethane	8.171	113	161712	48.545	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.100%
50) Toluene-d8	10.571	98	612354	49.166	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.340%
62) 4-Bromofluorobenzene	12.853	95	192511	46.667	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122722\
Data File : VN075978.D
Acq On : 27 Dec 2022 10:59
Operator : JC\MD
Sample : VN1227WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1227WBL01

Quant Time: Dec 28 04:43:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

