

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122722\
 Data File : VN075983.D
 Acq On : 27 Dec 2022 13:11
 Operator : JC\MD
 Sample : N6210-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Dec 28 04:44:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	335869	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	532518	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	467371	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	198849	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	174470	47.748	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.500%
35) Dibromofluoromethane	8.171	113	158942	48.196	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.400%
50) Toluene-d8	10.571	98	607899	49.302	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.600%
62) 4-Bromofluorobenzene	12.853	95	194063	47.519	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122722\
Data File : VN075983.D
Acq On : 27 Dec 2022 13:11
Operator : JC\MD
Sample : N6210-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Dec 28 04:44:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

