

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122724\
 Data File : VN085327.D
 Acq On : 27 Dec 2024 13:12
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Dec 28 01:40:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 28 01:36:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	120723	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	221370	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	188737	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	72957	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	96912	56.284	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	112.560%	
35) Dibromofluoromethane	8.159	113	73263	53.050	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	106.100%	
50) Toluene-d8	10.565	98	270098	53.297	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	106.600%	
62) 4-Bromofluorobenzene	12.847	95	79232	47.076	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	94.160%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122724\
Data File : VN085327.D
Acq On : 27 Dec 2024 13:12
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Dec 28 01:40:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
Quant Title : SW846 8260
QLast Update : Sat Dec 28 01:36:16 2024
Response via : Initial Calibration

