

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122820\
 Data File : VN065274.D
 Acq On : 28 Dec 2020 9:55
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Quant Time: Dec 29 02:17:53 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	297121	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	443761	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	404273	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	199404	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	159144	47.45	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.90%		
35) Dibromofluoromethane	7.550	113	133287	49.79	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	99.58%		
50) Toluene-d8	10.058	98	501451	49.97	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	99.94%		
62) 4-Bromofluorobenzene	12.373	95	187259	47.48	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	94.96%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.840	85	138584	50.41	ug/l	98
3) Chloromethane	2.045	50	163566	52.28	ug/l	100
4) Vinyl Chloride	2.174	62	168966	51.91	ug/l	99
5) Bromomethane	2.524	94	115478	49.91	ug/l	100
6) Chloroethane	2.676	64	105888	50.33	ug/l	99
7) Trichlorofluoromethane	2.991	101	233178	47.88	ug/l	99
8) Diethyl Ether	3.380	74	91654	47.80	ug/l	98
9) 1,1,2-Trichlorotrifluo...	3.717	101	132961	48.68	ug/l	98
10) Methyl Iodide	3.907	142	208200	50.45	ug/l	95
11) Tert butyl alcohol	4.769	59	125339	228.37	ug/l	99
12) 1,1-Dichloroethene	3.695	96	136459	46.76	ug/l	97
13) Acrolein	3.573	56	69097	186.77	ug/l	100
14) Allyl chloride	4.274	41	237545	49.38	ug/l	97
15) Acrylonitrile	4.936	53	327446	246.89	ug/l	99
16) Acetone	3.782	43	355211	252.46	ug/l	98
17) Carbon Disulfide	4.004	76	375444	50.27	ug/l	99
18) Methyl Acetate	4.286	43	149119	49.51	ug/l	97
19) Methyl tert-butyl Ether	5.000	73	460762	49.71	ug/l	100
20) Methylene Chloride	4.502	84	154599	47.24	ug/l	97
21) trans-1,2-Dichloroethene	4.984	96	146067	50.98	ug/l	99
22) Diisopropyl ether	5.910	45	476712	49.78	ug/l	98
23) Vinyl Acetate	5.849	43	1994187	247.09	ug/l	98
24) 1,1-Dichloroethane	5.798	63	267050	50.84	ug/l	99
25) 2-Butanone	6.794	43	463420	246.10	ug/l	100
26) 2,2-Dichloropropane	6.775	77	243971	48.62	ug/l	100
27) cis-1,2-Dichloroethene	6.778	96	171946	51.57	ug/l	100
28) Bromochloromethane	7.151	49	112101	46.96	ug/l	98
29) Tetrahydrofuran	7.174	42	282960	232.37	ug/l	98
30) Chloroform	7.328	83	275518	50.95	ug/l	99
31) Cyclohexane	7.611	56	233598	47.16	ug/l	96
32) 1,1,1-Trichloroethane	7.527	97	244744	49.96	ug/l	99
36) 1,1-Dichloropropene	7.749	75	208244	51.61	ug/l	99
37) Ethyl Acetate	6.888	43	189332	47.34	ug/l	97
38) Carbon Tetrachloride	7.733	117	213062	50.28	ug/l	100
39) Methylcyclohexane	9.045	83	245101	50.56	ug/l	99
40) Benzene	8.000	78	614801	51.22	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122820\
 Data File : VN065274.D
 Acq On : 28 Dec 2020 9:55
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Quant Time: Dec 29 02:17:53 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.126	41	90692	43.58	ug/l	90
42) 1,2-Dichloroethane	8.084	62	217017	48.92	ug/l	98
43) Isopropyl Acetate	8.129	43	333765	48.42	ug/l	98
44) Trichloroethene	8.801	130	181131	52.28	ug/l	97
45) 1,2-Dichloropropane	9.084	63	162672	52.13	ug/l	97
46) Dibromomethane	9.174	93	107277	50.62	ug/l	92
47) Bromodichloromethane	9.370	83	227275	50.86	ug/l	100
48) Methyl methacrylate	9.167	41	165015	49.32	ug/l	96
49) 1,4-Dioxane	9.171	88	51054	941.43	ug/l	97
51) 4-Methyl-2-Pentanone	9.952	43	933184	240.79	ug/l	97
52) Toluene	10.122	92	393696	52.04	ug/l	100
53) t-1,3-Dichloropropene	10.351	75	255281	50.69	ug/l	99
54) cis-1,3-Dichloropropene	9.807	75	274681	51.33	ug/l	96
55) 1,1,2-Trichloroethane	10.531	97	153623	51.32	ug/l	96
56) Ethyl methacrylate	10.399	69	231802	49.99	ug/l	94
57) 1,3-Dichloropropane	10.675	76	256830	52.00	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.662	63	501787	248.82	ug/l	99
59) 2-Hexanone	10.724	43	686223	240.86	ug/l	96
60) Dibromochloromethane	10.871	129	183958	51.56	ug/l	99
61) 1,2-Dibromoethane	10.978	107	163150	50.77	ug/l	100
64) Tetrachloroethene	10.601	164	206323	52.55	ug/l	95
65) Chlorobenzene	11.402	112	430367	52.30	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.479	131	166002	51.84	ug/l	99
67) Ethyl Benzene	11.482	91	766377	51.68	ug/l	99
68) m/p-Xylenes	11.592	106	584008	104.55	ug/l	99
69) o-Xylene	11.920	106	282815	52.15	ug/l	99
70) Styrene	11.936	104	482615	53.34	ug/l	99
71) Bromoform	12.100	173	143518	52.06	ug/l #	100
73) Isopropylbenzene	12.222	105	755355	49.74	ug/l	100
74) N-amyl acetate	12.042	43	293702	48.57	ug/l	96
75) 1,1,2,2-Tetrachloroethane	12.476	83	195091	47.64	ug/l	98
76) 1,2,3-Trichloropropane	12.527	75	284748	70.00	ug/l #	100
77) Bromobenzene	12.498	156	201381	50.74	ug/l	91
78) n-propylbenzene	12.563	91	852323	49.71	ug/l	98
79) 2-Chlorotoluene	12.646	91	506994	50.22	ug/l	98
80) 1,3,5-Trimethylbenzene	12.704	105	645212	49.82	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.273	75	77790	48.22	ug/l	92
82) 4-Chlorotoluene	12.746	91	521201	50.13	ug/l	98
83) tert-Butylbenzene	12.968	119	548083	49.67	ug/l	100
84) 1,2,4-Trimethylbenzene	13.013	105	644804	51.14	ug/l	99
85) sec-Butylbenzene	13.145	105	722277	49.34	ug/l	99
86) p-Isopropyltoluene	13.260	119	669739	49.91	ug/l	99
87) 1,3-Dichlorobenzene	13.254	146	346172	52.15	ug/l	97
88) 1,4-Dichlorobenzene	13.334	146	343511	51.55	ug/l	98
89) n-Butylbenzene	13.585	91	575700	50.19	ug/l	99
90) Hexachloroethane	13.846	117	119195	51.66	ug/l	85
91) 1,2-Dichlorobenzene	13.627	146	334770	52.52	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.241	75	40022	47.44	ug/l	94
93) 1,2,4-Trichlorobenzene	14.881	180	188024	47.76	ug/l	99
94) Hexachlorobutadiene	14.977	225	118355	47.04	ug/l	100
95) Naphthalene	15.093	128	463439	45.20	ug/l	100
96) 1,2,3-Trichlorobenzene	15.264	180	169737	46.60	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122820\
Data File : VN065274.D
Acq On : 28 Dec 2020 9:55
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Quant Time: Dec 29 02:17:53 2020
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.624 min Scan# 1830
Delta R.T. -0.003 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:168 Resp: 297121
Ion Ratio Lower Upper
168 100
99 49.4 43.0 64.4



#2
Dichlorodifluoromethane
Concen: 50.41 ug/l
RT: 1.840 min Scan# 31
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 85 Resp: 138584
Ion Ratio Lower Upper
85 100
87 31.6 16.4 49.4









#3
Chloromethane
Concen: 52.28 ug/l
RT: 2.045 min Scan# 95
Delta R.T. -0.001 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 50 Resp: 163566
Ion Ratio Lower Upper
50 100
52 33.3 26.6 39.8



#4
Vinyl Chloride
Concen: 51.91 ug/l
RT: 2.174 min Scan# 135
Delta R.T. 0.003 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 62 Resp: 168966
Ion Ratio Lower Upper
62 100
64 32.5 26.5 39.7









#5
Bromomethane
Concen: 49.91 ug/l
RT: 2.524 min Scan# 244
Delta R.T. -0.004 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 94 Resp: 115478
Ion Ratio Lower Upper
94 100
96 94.0 75.4 113.0



#6
Chloroethane
Concen: 50.33 ug/l
RT: 2.676 min Scan# 291
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 64 Resp: 105888
Ion Ratio Lower Upper
64 100
66 32.3 25.6 38.4









#7
 Trichlorofluoromethane
 Concen: 47.88 ug/l
 RT: 2.991 min Scan# 389
 Delta R.T. -0.003 min
 Lab File: VN065274.D
 Acq: 28 Dec 2020 9:55




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion:101 Resp: 233178
 Ion Ratio Lower Upper
 101 100
 103 63.6 50.5 75.7



#8
 Diethyl Ether
 Concen: 47.80 ug/l
 RT: 3.380 min Scan# 510
 Delta R.T. -0.000 min
 Lab File: VN065274.D
 Acq: 28 Dec 2020 9:55

Tgt Ion: 74 Resp: 91654
 Ion Ratio Lower Upper
 74 100
 45 96.4 48.9 146.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 48.68 ug/l
RT: 3.717 min Scan# 615
Delta R.T. 0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:101 Resp: 132961
Ion Ratio Lower Upper
101 100
85 43.0 34.4 51.6
151 82.3 64.0 96.0



#10
Methyl Iodide
Concen: 50.45 ug/l
RT: 3.907 min Scan# 674
Delta R.T. 0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55


Tgt Ion:142 Resp: 208200
Ion Ratio Lower Upper
142 100
127 40.0 35.0 52.6
141 14.2 11.8 17.6








#11
Tert butyl alcohol
Concen: 228.37 ug/l
RT: 4.769 min Scan# 942
Delta R.T. -0.003 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050







Tgt Ion: 59 Resp: 125339
Ion Ratio Lower Upper
59 100
57 10.3 7.9 11.9



#12
1,1-Dichloroethene
Concen: 46.76 ug/l
RT: 3.695 min Scan# 608
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 96 Resp: 136459
Ion Ratio Lower Upper
96 100
61 159.6 123.4 185.2
98 62.6 49.8 74.8









#13
Acrolein
Concen: 186.77 ug/l
RT: 3.573 min Scan# 570
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 56 Resp: 69097
Ion Ratio Lower Upper
56 100
55 70.8 56.5 84.7



#14
Allyl chloride
Concen: 49.38 ug/l
RT: 4.274 min Scan# 788
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 41 Resp: 237545
Ion Ratio Lower Upper
41 100
39 65.9 50.9 76.3
76 33.9 29.0 43.4









#15
Acrylonitrile
Concen: 246.89 ug/l
RT: 4.936 min Scan# 994
Delta R.T. -0.007 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 53 Resp: 327446
Ion Ratio Lower Upper
53 100
52 82.7 65.5 98.3
51 36.2 28.6 43.0



#16
Acetone
Concen: 252.46 ug/l
RT: 3.782 min Scan# 635
Delta R.T. -0.003 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 43 Resp: 355211
Ion Ratio Lower Upper
43 100
58 32.9 25.7 38.5









#17
Carbon Disulfide
Concen: 50.27 ug/l
RT: 4.004 min Scan# 704
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 375444
Ion Ratio Lower Upper
76 100
78 8.9 7.4 11.0



#18
Methyl Acetate
Concen: 49.51 ug/l
RT: 4.286 min Scan# 792
Delta R.T. -0.001 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 43 Resp: 149119
Ion Ratio Lower Upper
43 100
74 23.5 20.2 30.2









#19
Methyl tert-butyl Ether
Concen: 49.71 ug/l
RT: 5.000 min Scan# 1014
Delta R.T. -0.004 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 73 Resp: 460762
Ion Ratio Lower Upper
73 100
57 21.7 17.4 26.0



#20
Methylene Chloride
Concen: 47.24 ug/l
RT: 4.502 min Scan# 859
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 84 Resp: 154599
Ion Ratio Lower Upper
84 100
49 121.6 93.4 140.0
51 37.4 28.2 42.4
86 64.4 51.7 77.5









#21
trans-1,2-Dichloroethene
Concen: 50.98 ug/l
RT: 4.984 min Scan# 1009
Delta R.T. 0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 96 Resp: 146067
Ion Ratio Lower Upper
96 100
61 137.8 108.4 162.6
98 64.6 51.3 76.9



#22
Diisopropyl ether
Concen: 49.78 ug/l
RT: 5.910 min Scan# 1297
Delta R.T. -0.004 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 45 Resp: 476712
Ion Ratio Lower Upper
45 100
43 59.0 46.2 69.2
87 28.5 24.1 36.1
59 11.3 9.0 13.4









#23
 Vinyl Acetate
 Concen: 247.09 ug/l
 RT: 5.849 min Scan# 1278
 Delta R.T. -0.003 min
 Lab File: VN065274.D
 Acq: 28 Dec 2020 9:55




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 43 Resp: 1994187
 Ion Ratio Lower Upper
 43 100
 86 10.4 8.9 13.3



#24
 1,1-Dichloroethane
 Concen: 50.84 ug/l
 RT: 5.798 min Scan# 1262
 Delta R.T. -0.000 min
 Lab File: VN065274.D
 Acq: 28 Dec 2020 9:55

Tgt Ion: 63 Resp: 267050
 Ion Ratio Lower Upper
 63 100
 98 6.9 3.5 10.5
 100 4.1 2.3 6.8








#25
2-Butanone
Concen: 246.10 ug/l
RT: 6.794 min Scan# 1572
Delta R.T. -0.004 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 463420
Ion Ratio Lower Upper
43 100
72 26.3 21.0 31.6



#26
2,2-Dichloropropane
Concen: 48.62 ug/l
RT: 6.775 min Scan# 1566
Delta R.T. -0.003 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 77 Resp: 243971
Ion Ratio Lower Upper
77 100
97 23.1 11.7 35.0









#27
cis-1,2-Dichloroethene
Concen: 51.57 ug/l
RT: 6.778 min Scan# 1567
Delta R.T. -0.004 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:	96	Resp:	171946
Ion	Ratio	Lower	Upper
96	100		
61	143.6	0.0	285.8
98	65.3	0.0	130.6



#28
Bromochloromethane
Concen: 46.96 ug/l
RT: 7.151 min Scan# 1683
Delta R.T. -0.004 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion:	49	Resp:	112101
Ion	Ratio	Lower	Upper
49	100		
129	2.0	0.0	3.8
130	84.3	66.2	99.2





#29

Tetrahydrofuran

Concen: 232.37 ug/l

RT: 7.174 min Scan# 1690

Delta R.T. -0.003 min

Lab File: VN065274.D

Acq: 28 Dec 2020 9:55





Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:	42	Resp:	282960
Ion	Ratio	Lower	Upper
42	100		
72	44.0	36.6	54.8
71	41.7	33.6	50.4



#30

Chloroform

Concen: 50.95 ug/l




RT: 7.328 min Scan# 1738


Delta R.T. -0.004 min

Lab File: VN065274.D

Acq: 28 Dec 2020 9:55


Tgt Ion:	83	Resp:	275518
Ion	Ratio	Lower	Upper
83	100		
85	64.5	52.6	78.8








#31
Cyclohexane
Concen: 47.16 ug/l
RT: 7.611 min Scan# 1826
Delta R.T. 0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050







Tgt Ion: 56 Resp: 233598
Ion Ratio Lower Upper
56 100
69 33.7 27.5 41.3
84 88.1 73.9 110.9



#32
1,1,1-Trichloroethane
Concen: 49.96 ug/l
RT: 7.527 min Scan# 1800
Delta R.T. -0.003 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 97 Resp: 244744
Ion Ratio Lower Upper
97 100
99 64.0 51.7 77.5
61 42.8 33.9 50.9









#33
1,2-Dichloroethane-d4
Concen: 47.45 ug/l
RT: 7.987 min Scan# 1943
Delta R.T. -0.004 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 65 Resp: 159144
Ion Ratio Lower Upper
65 100
67 52.9 0.0 108.2



#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.550 min Scan# 2118
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 114 Resp: 443761
Ion Ratio Lower Upper
114 100
63 19.5 0.0 38.6
88 15.6 0.0 32.6








#35
Dibromofluoromethane
Concen: 49.79 ug/l
RT: 7.550 min Scan# 1807
Delta R.T. 0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:113 Resp: 133287
Ion Ratio Lower Upper
113 100
111 101.1 81.6 122.4
192 23.0 16.2 24.2



#36
1,1-Dichloropropene
Concen: 51.61 ug/l
RT: 7.749 min Scan# 1869
Delta R.T. -0.004 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 75 Resp: 208244
Ion Ratio Lower Upper
75 100
110 36.2 17.8 53.3
77 30.6 25.0 37.4





#37

Ethyl Acetate

Concen: 47.34 ug/l

RT: 6.888 min Scan# 1601

Delta R.T. -0.003 min

Lab File: VN065274.D

Acq: 28 Dec 2020 9:55





Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:	43	Resp:	189332
Ion	Ratio	Lower	Upper
43	100		
61	14.2	12.8	19.2
70	10.8	8.8	13.2



#38

Carbon Tetrachloride

Concen: 50.28 ug/l




RT: 7.733 min Scan# 1864


Delta R.T. 0.000 min

Lab File: VN065274.D

Acq: 28 Dec 2020 9:55

Tgt Ion:	117	Resp:	213062
Ion	Ratio	Lower	Upper
117	100		
119	95.9	76.7	115.1
121	31.2	25.4	38.2









#39
Methylcyclohexane
Concen: 50.56 ug/l
RT: 9.045 min Scan# 2272
Delta R.T. 0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:	83	Resp:	245101
Ion	Ratio	Lower	Upper
83	100		
55	74.3	60.2	90.2
98	46.4	37.4	56.2



#40
Benzene
Concen: 51.22 ug/l
RT: 8.000 min Scan# 1947
Delta R.T. -0.004 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55


Tgt Ion:	78	Resp:	614801
Ion	Ratio	Lower	Upper
78	100		
77	23.4	17.9	26.9








#41
Methacrylonitrile
Concen: 43.58 ug/l
RT: 7.126 min Scan# 1675
Delta R.T. -0.003 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050







Tgt Ion:	41	Resp:	90692
Ion Ratio	Lower	Upper	
41	100		
39	62.6	51.0	76.6
67	83.2	79.8	119.8
52	34.6	31.5	47.3



#42
1,2-Dichloroethane
Concen: 48.92 ug/l
RT: 8.084 min Scan# 1973
Delta R.T. -0.003 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion:	62	Resp:	217017
Ion Ratio	Lower	Upper	
62	100		
98	9.1	0.0	19.6









#43
Isopropyl Acetate
Concen: 48.42 ug/l
RT: 8.129 min Scan# 1987
Delta R.T. -0.003 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 333765
Ion Ratio Lower Upper
43 100
61 20.5 17.2 25.8
87 13.4 11.4 17.0



#44
Trichloroethene
Concen: 52.28 ug/l
RT: 8.801 min Scan# 2196
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 130 Resp: 181131
Ion Ratio Lower Upper
130 100
95 96.3 0.0 197.8









#45
1,2-Dichloropropane
Concen: 52.13 ug/l
RT: 9.084 min Scan# 2284
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 63 Resp: 162672
Ion Ratio Lower Upper
63 100
65 29.8 25.0 37.4



#46
Dibromomethane
Concen: 50.62 ug/l
RT: 9.174 min Scan# 2312
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 93 Resp: 107277
Ion Ratio Lower Upper
93 100
95 83.3 68.8 103.2
174 116.0 83.1 124.7









#47
Bromodichloromethane
Concen: 50.86 ug/l
RT: 9.370 min Scan# 2373
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 227275
Ion Ratio Lower Upper
83 100
85 64.4 51.8 77.8
127 8.6 6.8 10.2



#48
Methyl methacrylate
Concen: 49.32 ug/l
RT: 9.167 min Scan# 2310
Delta R.T. 0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 41 Resp: 165015
Ion Ratio Lower Upper
41 100
69 84.5 71.8 107.8
39 56.1 44.1 66.1









#49
1,4-Dioxane
Concen: 941.43 ug/l
RT: 9.171 min Scan# 2311
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 88 Resp: 51054
Ion Ratio Lower Upper
88 100
43 31.7 24.2 36.4
58 67.9 52.5 78.7



#50
Toluene-d8
Concen: 49.97 ug/l
RT: 10.058 min Scan# 2587
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 98 Resp: 501451
Ion Ratio Lower Upper
98 100
100 64.5 51.7 77.5








#51
4-Methyl-2-Pentanone
Concen: 240.79 ug/l
RT: 9.952 min Scan# 2554
Delta R.T. -0.003 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 933184
Ion Ratio Lower Upper
43 100
58 38.9 32.4 48.6



#52
Toluene
Concen: 52.04 ug/l
RT: 10.122 min Scan# 2607
Delta R.T. -0.004 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 92 Resp: 393696
Ion Ratio Lower Upper
92 100
91 174.0 139.1 208.7









#53
t-1,3-Dichloropropene
Concen: 50.69 ug/l
RT: 10.351 min Scan# 2678
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 75 Resp: 255281
Ion Ratio Lower Upper
75 100
77 31.6 25.8 38.6



#54
cis-1,3-Dichloropropene
Concen: 51.33 ug/l
RT: 9.807 min Scan# 2509
Delta R.T. 0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 75 Resp: 274681
Ion Ratio Lower Upper
75 100
77 31.9 25.5 38.3
39 48.9 35.6 53.4









#55
1,1,2-Trichloroethane
Concen: 51.32 ug/l
RT: 10.531 min Scan# 2734
Delta R.T. -0.003 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 97	Resp: 153623
Ion Ratio	Lower Upper
97 100	
83 85.6	72.1 108.1
85 54.4	46.8 70.2
99 63.2	49.6 74.4



#56
Ethyl methacrylate
Concen: 49.99 ug/l
RT: 10.399 min Scan# 2693
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 69	Resp: 231802
Ion Ratio	Lower Upper
69 100	
41 65.7	49.4 74.0
39 36.6	25.9 38.9









#57
1,3-Dichloropropane
Concen: 52.00 ug/l
RT: 10.675 min Scan# 2779
Delta R.T. -0.004 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 256830
Ion Ratio Lower Upper
76 100
78 32.0 26.1 39.1



#58
2-Chloroethyl Vinyl ether
Concen: 248.82 ug/l
RT: 9.662 min Scan# 2464
Delta R.T. -0.001 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 63 Resp: 501787
Ion Ratio Lower Upper
63 100
106 28.7 22.7 34.1









#59
2-Hexanone
Concen: 240.86 ug/l
RT: 10.724 min Scan# 2794
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 686223
Ion Ratio Lower Upper
43 100
58 53.8 28.3 84.9



#60
Dibromochloromethane
Concen: 51.56 ug/l
RT: 10.871 min Scan# 2840
Delta R.T. -0.001 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 129 Resp: 183958
Ion Ratio Lower Upper
129 100
127 77.1 39.1 117.2









#61
1,2-Dibromoethane
Concen: 50.77 ug/l
RT: 10.978 min Scan# 2873
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:107 Resp: 163150
Ion Ratio Lower Upper
107 100
109 95.0 76.0 114.0



#62
4-Bromofluorobenzene
Concen: 47.48 ug/l
RT: 12.373 min Scan# 3307
Delta R.T. -0.003 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 95 Resp: 187259
Ion Ratio Lower Upper
95 100
174 90.7 0.0 160.4
176 88.2 0.0 158.4









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.379 min Scan# 2998
Delta R.T. -0.001 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:117 Resp: 404273
Ion Ratio Lower Upper
117 100
82 53.1 44.2 66.2
119 32.9 25.8 38.6



#64
Tetrachloroethene
Concen: 52.55 ug/l
RT: 10.601 min Scan# 2756
Delta R.T. 0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion:164 Resp: 206323
Ion Ratio Lower Upper
164 100
166 128.3 101.4 152.0
129 88.1 76.1 114.1
131 84.7 74.2 111.2









#65
Chlorobenzene
Concen: 52.30 ug/l
RT: 11.402 min Scan# 3005
Delta R.T. -0.003 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:112 Resp: 430367
Ion Ratio Lower Upper
112 100
114 31.4 25.1 37.7



#66
1,1,1,2-Tetrachloroethane
Concen: 51.84 ug/l
RT: 11.479 min Scan# 3029
Delta R.T. 0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion:131 Resp: 166002
Ion Ratio Lower Upper
131 100
133 96.5 47.5 142.5
119 67.0 33.3 99.8









#67
Ethyl Benzene
Concen: 51.68 ug/l
RT: 11.482 min Scan# 3030
Delta R.T. 0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 766377
Ion Ratio Lower Upper
91 100
106 31.9 25.2 37.8



#68
m/p-Xylenes
Concen: 104.55 ug/l
RT: 11.592 min Scan# 3064
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion:106 Resp: 584008
Ion Ratio Lower Upper
106 100
91 203.0 161.0 241.4









#69
o-Xylene
Concen: 52.15 ug/l
RT: 11.920 min Scan# 3166
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:106 Resp: 282815
Ion Ratio Lower Upper
106 100
91 213.0 105.5 316.5



#70
Styrene
Concen: 53.34 ug/l
RT: 11.936 min Scan# 3171
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion:104 Resp: 482615
Ion Ratio Lower Upper
104 100
78 49.2 39.0 58.6
103 54.7 44.3 66.5









#71
Bromoform
Concen: 52.06 ug/l
RT: 12.100 min Scan# 3222
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:173 Resp: 143518
Ion Ratio Lower Upper
173 100
175 48.3 24.3 72.8
254 0.0 0.0 0.0#



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.315 min Scan# 3600
Delta R.T. -0.003 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion:152 Resp: 199404
Ion Ratio Lower Upper
152 100
115 55.4 29.7 89.1
150 174.8 0.0 350.6





#73

Isopropylbenzene

Concen: 49.74 ug/l

RT: 12.222 min Scan# 3260

Delta R.T. -0.000 min

Lab File: VN065274.D





Acq: 28 Dec 2020 9:55

Tgt Ion:105 Resp: 755355

Ion Ratio Lower Upper

105 100

120 26.8 13.3 39.9



#74

N-ethyl acetate

Concen: 48.57 ug/l

RT: 12.042 min Scan# 3204

Delta R.T. -0.000 min

Lab File: VN065274.D

Acq: 28 Dec 2020 9:55

Tgt Ion: 43 Resp: 293702




Ion Ratio Lower Upper


43 100

70 40.7 34.5 51.7

55 22.9 20.4 30.6

61 24.2 20.0 30.0









#75
1,1,2,2-Tetrachloroethane
Concen: 47.64 ug/l
RT: 12.476 min Scan# 3339
Delta R.T. -0.003 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 195091
Ion Ratio Lower Upper
83 100
131 10.9 5.1 15.3
85 64.7 33.0 99.0



#76
1,2,3-Trichloropropane
Concen: 70.00 ug/l
RT: 12.527 min Scan# 3355
Delta R.T. 0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 75 Resp: 284748
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0









#77
Bromobenzene
Concen: 50.74 ug/l
RT: 12.498 min Scan# 3346
Delta R.T. -0.001 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:156 Resp: 201381
Ion Ratio Lower Upper
156 100
77 184.6 101.3 303.7
158 98.3 48.2 144.6



#78
n-propylbenzene
Concen: 49.71 ug/l
RT: 12.563 min Scan# 3366
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 91 Resp: 852323
Ion Ratio Lower Upper
91 100
120 23.9 11.6 34.7









#79
2-Chlorotoluene
Concen: 50.22 ug/l
RT: 12.646 min Scan# 3392
Delta R.T. 0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 506994
Ion Ratio Lower Upper
91 100
126 35.3 17.2 51.6



#80
1,3,5-Trimethylbenzene
Concen: 49.82 ug/l
RT: 12.704 min Scan# 3410
Delta R.T. -0.004 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 105 Resp: 645212
Ion Ratio Lower Upper
105 100
120 48.6 24.6 73.6









#81
trans-1,4-Dichloro-2-butene
Concen: 48.22 ug/l
RT: 12.273 min Scan# 3276
Delta R.T. 0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 75 Resp: 77790
Ion Ratio Lower Upper
75 100
53 91.9 67.0 100.4
89 46.1 39.6 59.4



#82
4-Chlorotoluene
Concen: 50.13 ug/l
RT: 12.746 min Scan# 3423
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 91 Resp: 521201
Ion Ratio Lower Upper
91 100
126 34.7 16.9 50.7









#83
tert-Butylbenzene
Concen: 49.67 ug/l
RT: 12.968 min Scan# 3492
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:119 Resp: 548083
Ion Ratio Lower Upper
119 100
91 64.4 32.3 96.8
134 25.5 12.6 37.6



#84
1,2,4-Trimethylbenzene
Concen: 51.14 ug/l
RT: 13.013 min Scan# 3506
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion:105 Resp: 644804
Ion Ratio Lower Upper
105 100
120 45.6 23.1 69.2









#85
 sec-Butylbenzene
 Concen: 49.34 ug/l
 RT: 13.145 min Scan# 3547
 Delta R.T. -0.000 min
 Lab File: VN065274.D
 Acq: 28 Dec 2020 9:55




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion:105 Resp: 722277
 Ion Ratio Lower Upper
 105 100
 134 20.5 10.1 30.1



#86
 p-Isopropyltoluene
 Concen: 49.91 ug/l
 RT: 13.260 min Scan# 3583
 Delta R.T. -0.004 min
 Lab File: VN065274.D
 Acq: 28 Dec 2020 9:55

Tgt Ion:119 Resp: 669739
 Ion Ratio Lower Upper
 119 100
 134 26.7 13.1 39.1
 91 23.0 11.5 34.5









#87
1,3-Dichlorobenzene
Concen: 52.15 ug/l
RT: 13.254 min Scan# 3581
Delta R.T. -0.003 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:146 Resp: 346172
Ion Ratio Lower Upper
146 100
111 37.7 20.6 61.9
148 64.1 31.9 95.5



#88
1,4-Dichlorobenzene
Concen: 51.55 ug/l
RT: 13.334 min Scan# 3606
Delta R.T. -0.004 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion:146 Resp: 343511
Ion Ratio Lower Upper
146 100
111 36.8 19.8 59.3
148 64.2 31.9 95.5









#89
n-Butylbenzene
Concen: 50.19 ug/l
RT: 13.585 min Scan# 3684
Delta R.T. -0.003 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 575700
Ion Ratio Lower Upper
91 100
92 52.3 26.0 78.0
134 27.0 12.9 38.6



#90
Hexachloroethane
Concen: 51.66 ug/l
RT: 13.846 min Scan# 3765
Delta R.T. -0.003 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 117 Resp: 119195
Ion Ratio Lower Upper
117 100
201 93.8 40.1 120.3









#91
1,2-Dichlorobenzene
Concen: 52.52 ug/l
RT: 13.627 min Scan# 3697
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:146 Resp: 334770
Ion Ratio Lower Upper
146 100
111 39.6 21.1 63.1
148 64.7 32.0 96.2



#92
1,2-Dibromo-3-Chloropropane
Concen: 47.44 ug/l
RT: 14.241 min Scan# 3888
Delta R.T. 0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion: 75 Resp: 40022
Ion Ratio Lower Upper
75 100
155 91.1 42.8 128.3
157 118.4 56.4 169.1









#93
1,2,4-Trichlorobenzene
Concen: 47.76 ug/l
RT: 14.881 min Scan# 4087
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:180 Resp: 188024
Ion Ratio Lower Upper
180 100
182 95.9 47.6 142.9
145 28.7 15.3 45.8



#94
Hexachlorobutadiene
Concen: 47.04 ug/l
RT: 14.977 min Scan# 4117
Delta R.T. -0.004 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion:225 Resp: 118355
Ion Ratio Lower Upper
225 100
223 61.9 31.0 93.0
227 63.8 32.0 96.0









#95
Naphthalene
Concen: 45.20 ug/l
RT: 15.093 min Scan# 4153
Delta R.T. -0.003 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:128 Resp: 463439
Ion Ratio Lower Upper
128 100
127 12.8 10.3 15.5
129 10.8 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 46.60 ug/l
RT: 15.264 min Scan# 4206
Delta R.T. -0.000 min
Lab File: VN065274.D
Acq: 28 Dec 2020 9:55

Tgt Ion:180 Resp: 169737
Ion Ratio Lower Upper
180 100
182 94.7 48.1 144.4
145 29.9 16.4 49.1

