

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122820\
 Data File : VN065277.D
 Acq On : 28 Dec 2020 11:29
 Operator : JC/MD
 Sample : VN1228WBS01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Dec 29 02:19:03 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	293939	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	431315	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	391566	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	184386	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.988	65	159847	48.17	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.34%
35) Dibromofluoromethane	7.550	113	134094	51.53	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.06%
50) Toluene-d8	10.058	98	497070	50.97	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.94%
62) 4-Bromofluorobenzene	12.373	95	180461	47.12	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.24%

Target Compounds					Qvalue
2) Dichlorodifluoromethane	1.840	85	48566	17.86	ug/l 97
3) Chloromethane	2.046	50	56797	18.35	ug/l 96
4) Vinyl Chloride	2.174	62	59008	18.32	ug/l 97
5) Bromomethane	2.538	94	42328	18.49	ug/l 100
6) Chloroethane	2.682	64	36751	17.66	ug/l 99
7) Trichlorofluoromethane	2.994	101	83833	17.40	ug/l 97
8) Diethyl Ether	3.380	74	31920	16.83	ug/l 100
9) 1,1,2-Trichlorotrifluo...	3.718	101	47162	17.45	ug/l 99
10) Methyl Iodide	3.907	142	70145	17.18	ug/l 95
11) Tert butyl alcohol	4.766	59	45236	83.31	ug/l 98
12) 1,1-Dichloroethene	3.695	96	49171	17.03	ug/l 99
13) Acrolein	3.570	56	31990	88.69	ug/l 98
14) Allyl chloride	4.277	41	83764	17.60	ug/l 96
15) Acrylonitrile	4.939	53	115295	87.87	ug/l 99
16) Acetone	3.782	43	111188	72.20	ug/l 98
17) Carbon Disulfide	4.004	76	131865	17.85	ug/l 99
18) Methyl Acetate	4.287	43	51978	17.44	ug/l 97
19) Methyl tert-butyl Ether	5.000	73	161813	17.65	ug/l 98
20) Methylene Chloride	4.499	84	53756	16.60	ug/l 97
21) trans-1,2-Dichloroethene	4.988	96	50478	17.81	ug/l 97
22) Diisopropyl ether	5.910	45	168119	17.75	ug/l 95
23) Vinyl Acetate	5.849	43	697787	87.40	ug/l 98
24) 1,1-Dichloroethane	5.798	63	92860	17.87	ug/l 99
25) 2-Butanone	6.798	43	157421	84.50	ug/l 100
26) 2,2-Dichloropropane	6.775	77	84755	17.07	ug/l 100
27) cis-1,2-Dichloroethene	6.782	96	60048	18.20	ug/l 99
28) Bromochloromethane	7.148	49	38687	16.80	ug/l 98
29) Tetrahydrofuran	7.174	42	102513	85.10	ug/l 98
30) Chloroform	7.332	83	95285	17.81	ug/l 98
31) Cyclohexane	7.611	56	85706	17.49	ug/l 95
32) 1,1,1-Trichloroethane	7.528	97	85810	17.70	ug/l 98
36) 1,1-Dichloropropene	7.750	75	71170	18.15	ug/l 98
37) Ethyl Acetate	6.888	43	66888	17.21	ug/l 98
38) Carbon Tetrachloride	7.730	117	72435	17.59	ug/l 98
39) Methylcyclohexane	9.042	83	85040	18.05	ug/l 98
40) Benzene	8.000	78	213114	18.27	ug/l 98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122820\
 Data File : VN065277.D
 Acq On : 28 Dec 2020 11:29
 Operator : JC/MD
 Sample : VN1228WBS01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Dec 29 02:19:03 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.129	41	31592	15.62	ug/l	90
42) 1,2-Dichloroethane	8.087	62	77393	17.95	ug/l	98
43) Isopropyl Acetate	8.129	43	115431	17.23	ug/l	99
44) Trichloroethene	8.798	130	62460	18.55	ug/l	99
45) 1,2-Dichloropropane	9.084	63	55968	18.45	ug/l	98
46) Dibromomethane	9.174	93	36608	17.77	ug/l	92
47) Bromodichloromethane	9.370	83	77163	17.77	ug/l	100
48) Methyl methacrylate	9.168	41	54867	16.87	ug/l	94
49) 1,4-Dioxane	9.171	88	18199	345.27	ug/l	96
51) 4-Methyl-2-Pentanone	9.952	43	328900	87.31	ug/l	97
52) Toluene	10.122	92	136112	18.51	ug/l	100
53) t-1,3-Dichloropropene	10.351	75	85473	17.46	ug/l	99
54) cis-1,3-Dichloropropene	9.807	75	92463	17.78	ug/l	96
55) 1,1,2-Trichloroethane	10.531	97	53332	18.33	ug/l	96
56) Ethyl methacrylate	10.399	69	78845	17.50	ug/l	93
57) 1,3-Dichloropropane	10.676	76	88633	18.46	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.663	63	171391	87.44	ug/l	100
59) 2-Hexanone	10.724	43	237814	85.88	ug/l	96
60) Dibromochloromethane	10.872	129	62975	18.16	ug/l	98
61) 1,2-Dibromoethane	10.978	107	57078	18.27	ug/l	98
64) Tetrachloroethene	10.602	164	71881	18.90	ug/l	95
65) Chlorobenzene	11.405	112	146810	18.42	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.479	131	57499	18.54	ug/l	99
67) Ethyl Benzene	11.483	91	261253	18.19	ug/l	100
68) m/p-Xylenes	11.592	106	196021	36.23	ug/l	98
69) o-Xylene	11.920	106	97321	18.53	ug/l	99
70) Styrene	11.936	104	158729	18.11	ug/l	99
71) Bromoform	12.100	173	46236	17.32	ug/l #	99
73) Isopropylbenzene	12.222	105	256260	18.25	ug/l	100
74) N-amyl acetate	12.042	43	95936	17.16	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.476	83	68810	18.17	ug/l	97
76) 1,2,3-Trichloropropane	12.528	75	95820	25.47	ug/l #	100
77) Bromobenzene	12.499	156	69178	18.85	ug/l	91
78) n-propylbenzene	12.563	91	283247	17.87	ug/l	98
79) 2-Chlorotoluene	12.647	91	169014	18.10	ug/l	98
80) 1,3,5-Trimethylbenzene	12.708	105	214775	17.93	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.274	75	24108	16.16	ug/l	88
82) 4-Chlorotoluene	12.746	91	175157	18.22	ug/l	99
83) tert-Butylbenzene	12.968	119	185093	18.14	ug/l	100
84) 1,2,4-Trimethylbenzene	13.013	105	212051	18.19	ug/l	100
85) sec-Butylbenzene	13.145	105	238092	17.59	ug/l	98
86) p-Isopropyltoluene	13.261	119	222349	17.92	ug/l	99
87) 1,3-Dichlorobenzene	13.254	146	113982	18.57	ug/l	97
88) 1,4-Dichlorobenzene	13.338	146	112720	18.29	ug/l	99
89) n-Butylbenzene	13.589	91	187631	17.69	ug/l	99
90) Hexachloroethane	13.846	117	37285	17.48	ug/l	84
91) 1,2-Dichlorobenzene	13.627	146	110894	18.81	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.245	75	13505	17.31	ug/l	94
93) 1,2,4-Trichlorobenzene	14.881	180	56326	17.39	ug/l	98
94) Hexachlorobutadiene	14.978	225	40200	17.28	ug/l	99
95) Naphthalene	15.097	128	143313	17.36	ug/l	99
96) 1,2,3-Trichlorobenzene	15.264	180	53853	18.10	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122820\
Data File : VN065277.D
Acq On : 28 Dec 2020 11:29
Operator : JC/MD
Sample : VN1228WBS01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :


Quant Time: Dec 29 02:19:03 2020
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 29 02:19:03 2020
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.624 min Scan# 1830
Delta R.T. -0.003 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:168 Resp: 293939
Ion Ratio Lower Upper
168 100
99 50.2 43.0 64.4



#2
Dichlorodifluoromethane
Concen: 17.86 ug/l
RT: 1.840 min Scan# 31
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion: 85 Resp: 48566
Ion Ratio Lower Upper
85 100
87 31.4 16.4 49.4









#3
Chloromethane
Concen: 18.35 ug/l
RT: 2.046 min Scan# 95
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 50 Resp: 56797
Ion Ratio Lower Upper
50 100
52 31.2 26.6 39.8



#4
Vinyl Chloride
Concen: 18.32 ug/l
RT: 2.174 min Scan# 135
Delta R.T. 0.003 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion: 62 Resp: 59008
Ion Ratio Lower Upper
62 100
64 31.6 26.5 39.7









#5
Bromomethane
Concen: 18.49 ug/l
RT: 2.538 min Scan# 248
Delta R.T. 0.010 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 94 Resp: 42328
Ion Ratio Lower Upper
94 100
96 94.2 75.4 113.0



#6
Chloroethane
Concen: 17.66 ug/l
RT: 2.682 min Scan# 293
Delta R.T. 0.006 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion: 64 Resp: 36751
Ion Ratio Lower Upper
64 100
66 32.6 25.6 38.4









#7
Trichlorofluoromethane
Concen: 17.40 ug/l
RT: 2.994 min Scan# 390
Delta R.T. 0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:101 Resp: 83833
Ion Ratio Lower Upper
101 100
103 65.1 50.5 75.7



#8
Diethyl Ether
Concen: 16.83 ug/l
RT: 3.380 min Scan# 510
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion: 74 Resp: 31920
Ion Ratio Lower Upper
74 100
45 97.5 48.9 146.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 17.45 ug/l
RT: 3.718 min Scan# 615
Delta R.T. 0.001 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:101 Resp: 47162
Ion Ratio Lower Upper
101 100
85 43.1 34.4 51.6
151 81.5 64.0 96.0



#10
Methyl Iodide
Concen: 17.18 ug/l
RT: 3.907 min Scan# 674
Delta R.T. 0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29


Tgt Ion:142 Resp: 70145
Ion Ratio Lower Upper
142 100
127 39.5 35.0 52.6
141 14.0 11.8 17.6








#11
Tert butyl alcohol
Concen: 83.31 ug/l
RT: 4.766 min Scan# 941
Delta R.T. -0.006 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Instrument :
MSVOA_N
ClientSampled :







Tgt Ion: 59 Resp: 45236
Ion Ratio Lower Upper
59 100
57 10.5 7.9 11.9



#12
1,1-Dichloroethene
Concen: 17.03 ug/l
RT: 3.695 min Scan# 608
Delta R.T. 0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion: 96 Resp: 49171
Ion Ratio Lower Upper
96 100
61 155.9 123.4 185.2
98 63.0 49.8 74.8









#13
Acrolein
Concen: 88.69 ug/l
RT: 3.570 min Scan# 569
Delta R.T. -0.003 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 56 Resp: 31990
Ion Ratio Lower Upper
56 100
55 69.2 56.5 84.7



#14
Allyl chloride
Concen: 17.60 ug/l
RT: 4.277 min Scan# 789
Delta R.T. 0.003 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion: 41 Resp: 83764
Ion Ratio Lower Upper
41 100
39 67.0 50.9 76.3
76 33.5 29.0 43.4









#15
Acrylonitrile
Concen: 87.87 ug/l
RT: 4.939 min Scan# 995
Delta R.T. -0.004 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 53 Resp: 115295
Ion Ratio Lower Upper
53 100
52 81.7 65.5 98.3
51 36.8 28.6 43.0



#16
Acetone
Concen: 72.20 ug/l
RT: 3.782 min Scan# 635
Delta R.T. -0.003 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion: 43 Resp: 111188
Ion Ratio Lower Upper
43 100
58 33.1 25.7 38.5









#17
Carbon Disulfide
Concen: 17.85 ug/l
RT: 4.004 min Scan# 704
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 76 Resp: 131865
Ion Ratio Lower Upper
76 100
78 9.5 7.4 11.0



#18
Methyl Acetate
Concen: 17.44 ug/l
RT: 4.287 min Scan# 792
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion: 43 Resp: 51978
Ion Ratio Lower Upper
43 100
74 23.8 20.2 30.2









#19
Methyl tert-butyl Ether
Concen: 17.65 ug/l
RT: 5.000 min Scan# 1014
Delta R.T. -0.003 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 73 Resp: 161813
Ion Ratio Lower Upper
73 100
57 22.6 17.4 26.0



#20
Methylene Chloride
Concen: 16.60 ug/l
RT: 4.499 min Scan# 858
Delta R.T. -0.003 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion: 84 Resp: 53756
Ion Ratio Lower Upper
84 100
49 120.0 93.4 140.0
51 36.3 28.2 42.4
86 61.0 51.7 77.5









#21
trans-1,2-Dichloroethene
Concen: 17.81 ug/l
RT: 4.988 min Scan# 1010
Delta R.T. 0.004 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 96 Resp: 50478
Ion Ratio Lower Upper
96 100
61 140.5 108.4 162.6
98 63.5 51.3 76.9



#22
Diisopropyl ether
Concen: 17.75 ug/l
RT: 5.910 min Scan# 1297
Delta R.T. -0.004 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion: 45 Resp: 168119
Ion Ratio Lower Upper
45 100
43 62.0 46.2 69.2
87 28.0 24.1 36.1
59 11.0 9.0 13.4









#23
 Vinyl Acetate
 Concen: 87.40 ug/l
 RT: 5.849 min Scan# 1278
 Delta R.T. -0.003 min
 Lab File: VN065277.D
 Acq: 28 Dec 2020 11:29




Instrument :
 MSVOA_N
 ClientSampled :


Tgt Ion: 43 Resp: 697787
 Ion Ratio Lower Upper
 43 100
 86 10.4 8.9 13.3



#24
 1,1-Dichloroethane
 Concen: 17.87 ug/l
 RT: 5.798 min Scan# 1262
 Delta R.T. -0.000 min
 Lab File: VN065277.D
 Acq: 28 Dec 2020 11:29

Tgt Ion: 63 Resp: 92860
 Ion Ratio Lower Upper
 63 100
 98 6.4 3.5 10.5
 100 4.3 2.3 6.8









#25
2-Butanone
Concen: 84.50 ug/l
RT: 6.798 min Scan# 1573
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 157421
Ion Ratio Lower Upper
43 100
72 26.1 21.0 31.6



#26
2,2-Dichloropropane
Concen: 17.07 ug/l
RT: 6.775 min Scan# 1566
Delta R.T. -0.003 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29


Tgt Ion: 77 Resp: 84755
Ion Ratio Lower Upper
77 100
97 23.2 11.7 35.0








#27
 cis-1,2-Dichloroethene
 Concen: 18.20 ug/l
 RT: 6.782 min Scan# 1568
 Delta R.T. -0.000 min
 Lab File: VN065277.D
 Acq: 28 Dec 2020 11:29

Instrument :
 MSVOA_N
 ClientSampleId :







Tgt Ion:	96	Resp:	60048
Ion	Ratio	Lower	Upper
96	100		
61	143.6	0.0	285.8
98	64.2	0.0	130.6



#28
 Bromochloromethane
 Concen: 16.80 ug/l
 RT: 7.148 min Scan# 1682
 Delta R.T. -0.007 min
 Lab File: VN065277.D
 Acq: 28 Dec 2020 11:29

Tgt Ion:	49	Resp:	38687
Ion	Ratio	Lower	Upper
49	100		
129	0.9	0.0	3.8
130	84.2	66.2	99.2





#29

Tetrahydrofuran

Concen: 85.10 ug/l

RT: 7.174 min Scan# 1690





Delta R.T. -0.003 min

Lab File: VN065277.D

Acq: 28 Dec 2020 11:29

Tgt Ion: 42 Resp: 102513

Ion	Ratio	Lower	Upper
42	100		
72	44.2	36.6	54.8
71	41.1	33.6	50.4



#30

Chloroform

Concen: 17.81 ug/l

RT: 7.332 min Scan# 1739




Delta R.T. -0.000 min


Lab File: VN065277.D

Acq: 28 Dec 2020 11:29

Tgt Ion: 83 Resp: 95285


Ion	Ratio	Lower	Upper
83	100		
85	64.3	52.6	78.8








#31
Cyclohexane
Concen: 17.49 ug/l
RT: 7.611 min Scan# 1826
Delta R.T. 0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion: 56 Resp: 85706
Ion Ratio Lower Upper
56 100
69 37.0 27.5 41.3
84 86.9 73.9 110.9




#32
1,1,1-Trichloroethane
Concen: 17.70 ug/l
RT: 7.528 min Scan# 1800
Delta R.T. -0.003 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29



Tgt Ion: 97 Resp: 85810
Ion Ratio Lower Upper
97 100
99 62.8 51.7 77.5
61 42.7 33.9 50.9









#33
1,2-Dichloroethane-d4
Concen: 48.17 ug/l
RT: 7.988 min Scan# 1943
Delta R.T. -0.003 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 65 Resp: 159847
Ion Ratio Lower Upper
65 100
67 52.9 0.0 108.2



#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.550 min Scan# 2118
Delta R.T. 0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion: 114 Resp: 431315
Ion Ratio Lower Upper
114 100
63 19.5 0.0 38.6
88 15.4 0.0 32.6









#35
Dibromofluoromethane
Concen: 51.53 ug/l
RT: 7.550 min Scan# 1807
Delta R.T. 0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:113 Resp: 134094
Ion Ratio Lower Upper
113 100
111 102.0 81.6 122.4
192 23.4 16.2 24.2



#36
1,1-Dichloropropene
Concen: 18.15 ug/l
RT: 7.750 min Scan# 1869
Delta R.T. -0.003 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion: 75 Resp: 71170
Ion Ratio Lower Upper
75 100
110 37.1 17.8 53.3
77 31.5 25.0 37.4





#37

Ethyl Acetate

Concen: 17.21 ug/l

RT: 6.888 min Scan# 1601





Delta R.T. -0.003 min

Lab File: VN065277.D

Acq: 28 Dec 2020 11:29

Tgt Ion:	43	Resp:	66888
Ion	Ratio	Lower	Upper
43	100		
61	15.0	12.8	19.2
70	10.6	8.8	13.2

Instrument :
MSVOA_N
ClientSampled :



#38

Carbon Tetrachloride

Concen: 17.59 ug/l




RT: 7.730 min Scan# 1863


Delta R.T. -0.003 min

Lab File: VN065277.D

Acq: 28 Dec 2020 11:29

Tgt Ion:	117	Resp:	72435
Ion	Ratio	Lower	Upper
117	100		
119	97.1	76.7	115.1
121	30.0	25.4	38.2









#39
Methylcyclohexane
Concen: 18.05 ug/l
RT: 9.042 min Scan# 2271
Delta R.T. -0.003 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 85040
Ion Ratio Lower Upper
83 100
55 72.5 60.2 90.2
98 46.6 37.4 56.2



#40
Benzene
Concen: 18.27 ug/l
RT: 8.000 min Scan# 1947
Delta R.T. -0.004 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29


Tgt Ion: 78 Resp: 213114
Ion Ratio Lower Upper
78 100
77 23.3 17.9 26.9








#41
Methacrylonitrile
Concen: 15.62 ug/l
RT: 7.129 min Scan# 1676
Delta R.T. 0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:	41	Resp:	31592
Ion Ratio	Lower	Upper	
41	100		
39	59.6	51.0	76.6
67	85.4	79.8	119.8
52	34.4	31.5	47.3



#42
1,2-Dichloroethane
Concen: 17.95 ug/l
RT: 8.087 min Scan# 1974
Delta R.T. 0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion:	62	Resp:	77393
Ion Ratio	Lower	Upper	
62	100		
98	9.0	0.0	19.6









#43
Isopropyl Acetate
Concen: 17.23 ug/l
RT: 8.129 min Scan# 1987
Delta R.T. -0.003 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 115431
Ion Ratio Lower Upper
43 100
61 21.0 17.2 25.8
87 13.7 11.4 17.0



#44
Trichloroethene
Concen: 18.55 ug/l
RT: 8.798 min Scan# 2195
Delta R.T. -0.003 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion:130 Resp: 62460
Ion Ratio Lower Upper
130 100
95 99.5 0.0 197.8









#45
1,2-Dichloropropane
Concen: 18.45 ug/l
RT: 9.084 min Scan# 2284
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 63 Resp: 55968
Ion Ratio Lower Upper
63 100
65 32.1 25.0 37.4



#46
Dibromomethane
Concen: 17.77 ug/l
RT: 9.174 min Scan# 2312
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion: 93 Resp: 36608
Ion Ratio Lower Upper
93 100
95 84.0 68.8 103.2
174 116.9 83.1 124.7









#47
Bromodichloromethane
Concen: 17.77 ug/l
RT: 9.370 min Scan# 2373
Delta R.T. 0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:	83	Resp:	77163
Ion Ratio	Lower	Upper	
83	100		
85	65.0	51.8	77.8
127	8.6	6.8	10.2



#48
Methyl methacrylate
Concen: 16.87 ug/l
RT: 9.168 min Scan# 2310
Delta R.T. 0.001 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion:	41	Resp:	54867
Ion Ratio	Lower	Upper	
41	100		
69	85.5	71.8	107.8
39	60.5	44.1	66.1









#49
1,4-Dioxane
Concen: 345.27 ug/l
RT: 9.171 min Scan# 2311
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 88 Resp: 18199
Ion Ratio Lower Upper
88 100
43 31.5 24.2 36.4
58 69.8 52.5 78.7



#50
Toluene-d8
Concen: 50.97 ug/l
RT: 10.058 min Scan# 2587
Delta R.T. 0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion: 98 Resp: 497070
Ion Ratio Lower Upper
98 100
100 64.4 51.7 77.5









#51
4-Methyl-2-Pentanone
Concen: 87.31 ug/l
RT: 9.952 min Scan# 2554
Delta R.T. -0.003 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 328900
Ion Ratio Lower Upper
43 100
58 38.8 32.4 48.6



#52
Toluene
Concen: 18.51 ug/l
RT: 10.122 min Scan# 2607
Delta R.T. -0.004 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion: 92 Resp: 136112
Ion Ratio Lower Upper
92 100
91 174.3 139.1 208.7









#53
t-1,3-Dichloropropene
Concen: 17.46 ug/l
RT: 10.351 min Scan# 2678
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 75 Resp: 85473
Ion Ratio Lower Upper
75 100
77 31.8 25.8 38.6



#54
cis-1,3-Dichloropropene
Concen: 17.78 ug/l
RT: 9.807 min Scan# 2509
Delta R.T. 0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion: 75 Resp: 92463
Ion Ratio Lower Upper
75 100
77 31.3 25.5 38.3
39 48.4 35.6 53.4









#55
1,1,2-Trichloroethane
Concen: 18.33 ug/l
RT: 10.531 min Scan# 2734
Delta R.T. -0.003 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:	97	Resp:	53332
Ion	Ratio	Lower	Upper
97	100		
83	85.6	72.1	108.1
85	54.7	46.8	70.2
99	64.4	49.6	74.4



#56
Ethyl methacrylate
Concen: 17.50 ug/l
RT: 10.399 min Scan# 2693
Delta R.T. 0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion:	69	Resp:	78845
Ion	Ratio	Lower	Upper
69	100		
41	65.5	49.4	74.0
39	38.7	25.9	38.9









#57
1,3-Dichloropropane
Concen: 18.46 ug/l
RT: 10.676 min Scan# 2779
Delta R.T. -0.003 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 76 Resp: 88633
Ion Ratio Lower Upper
76 100
78 32.5 26.1 39.1



#58
2-Chloroethyl Vinyl ether
Concen: 87.44 ug/l
RT: 9.663 min Scan# 2464
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion: 63 Resp: 171391
Ion Ratio Lower Upper
63 100
106 28.5 22.7 34.1









#59
2-Hexanone
Concen: 85.88 ug/l
RT: 10.724 min Scan# 2794
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 237814
Ion Ratio Lower Upper
43 100
58 53.9 28.3 84.9



#60
Dibromochloromethane
Concen: 18.16 ug/l
RT: 10.872 min Scan# 2840
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion:129 Resp: 62975
Ion Ratio Lower Upper
129 100
127 76.1 39.1 117.2









#61
1,2-Dibromoethane
Concen: 18.27 ug/l
RT: 10.978 min Scan# 2873
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:107 Resp: 57078
Ion Ratio Lower Upper
107 100
109 92.8 76.0 114.0



#62
4-Bromofluorobenzene
Concen: 47.12 ug/l
RT: 12.373 min Scan# 3307
Delta R.T. -0.003 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion: 95 Resp: 180461
Ion Ratio Lower Upper
95 100
174 91.2 0.0 160.4
176 88.4 0.0 158.4









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.380 min Scan# 2998
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:117 Resp: 391566
Ion Ratio Lower Upper
117 100
82 52.7 44.2 66.2
119 31.9 25.8 38.6



#64
Tetrachloroethene
Concen: 18.90 ug/l
RT: 10.602 min Scan# 2756
Delta R.T. 0.001 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion:164 Resp: 71881
Ion Ratio Lower Upper
164 100
166 130.9 101.4 152.0
129 88.8 76.1 114.1
131 86.9 74.2 111.2









#65
Chlorobenzene
Concen: 18.42 ug/l
RT: 11.405 min Scan# 3006
Delta R.T. 0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:112 Resp: 146810
Ion Ratio Lower Upper
112 100
114 32.9 25.1 37.7



#66
1,1,1,2-Tetrachloroethane
Concen: 18.54 ug/l
RT: 11.479 min Scan# 3029
Delta R.T. 0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion:131 Resp: 57499
Ion Ratio Lower Upper
131 100
133 95.4 47.5 142.5
119 65.9 33.3 99.8









#67
Ethyl Benzene
Concen: 18.19 ug/l
RT: 11.483 min Scan# 3030
Delta R.T. 0.001 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 261253
Ion Ratio Lower Upper
91 100
106 31.6 25.2 37.8



#68
m/p-Xylenes
Concen: 36.23 ug/l
RT: 11.592 min Scan# 3064
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion:106 Resp: 196021
Ion Ratio Lower Upper
106 100
91 204.9 161.0 241.4









#69
o-Xylene
Concen: 18.53 ug/l
RT: 11.920 min Scan# 3166
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:106 Resp: 97321
Ion Ratio Lower Upper
106 100
91 209.7 105.5 316.5



#70
Styrene
Concen: 18.11 ug/l
RT: 11.936 min Scan# 3171
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion:104 Resp: 158729
Ion Ratio Lower Upper
104 100
78 49.8 39.0 58.6
103 55.4 44.3 66.5





#71

Bromoform

Concen: 17.32 ug/l

RT: 12.100 min Scan# 3222





Delta R.T. -0.000 min

Lab File: VN065277.D

Acq: 28 Dec 2020 11:29

Tgt Ion:173	Resp:	46236
Ion Ratio	Lower	Upper
173	100	
175	49.2	24.3 72.8
254	0.0	0.0 0.0

Instrument :
MSVOA_N
ClientSampleId :



#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l




RT: 13.315 min Scan# 3600


Delta R.T. -0.003 min

Lab File: VN065277.D

Acq: 28 Dec 2020 11:29

Tgt Ion:152	Resp:	184386
Ion Ratio	Lower	Upper
152	100	
115	55.7	29.7 89.1
150	163.6	0.0 350.6





#73

Isopropylbenzene

Concen: 18.25 ug/l

RT: 12.222 min Scan# 3260

Delta R.T. 0.000 min

Lab File: VN065277.D





Acq: 28 Dec 2020 11:29

Tgt Ion:105 Resp: 256260

Ion Ratio Lower Upper

105 100

120 26.6 13.3 39.9



#74

N-ethyl acetate

Concen: 17.16 ug/l

RT: 12.042 min Scan# 3204

Delta R.T. 0.000 min

Lab File: VN065277.D

Acq: 28 Dec 2020 11:29

Tgt Ion: 43 Resp: 95936




Ion Ratio Lower Upper


43 100

70 40.6 34.5 51.7

55 23.6 20.4 30.6

61 23.8 20.0 30.0









#75
 1,1,2,2-Tetrachloroethane
 Concen: 18.17 ug/l
 RT: 12.476 min Scan# 3339
 Delta R.T. -0.003 min
 Lab File: VN065277.D
 Acq: 28 Dec 2020 11:29




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 83 Resp: 68810
 Ion Ratio Lower Upper
 83 100
 131 10.8 5.1 15.3
 85 63.5 33.0 99.0



#76
 1,2,3-Trichloropropane
 Concen: 25.47 ug/l
 RT: 12.528 min Scan# 3355
 Delta R.T. 0.001 min
 Lab File: VN065277.D
 Acq: 28 Dec 2020 11:29

Tgt Ion: 75 Resp: 95820
 Ion Ratio Lower Upper
 75 100
 77 0.0 0.0 0.0









#77
Bromobenzene
Concen: 18.85 ug/l
RT: 12.499 min Scan# 3346
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:156 Resp: 69178
Ion Ratio Lower Upper
156 100
77 182.9 101.3 303.7
158 97.2 48.2 144.6



#78
n-propylbenzene
Concen: 17.87 ug/l
RT: 12.563 min Scan# 3366
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion: 91 Resp: 283247
Ion Ratio Lower Upper
91 100
120 23.9 11.6 34.7









#79
2-Chlorotoluene
Concen: 18.10 ug/l
RT: 12.647 min Scan# 3392
Delta R.T. 0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 91 Resp: 169014
Ion Ratio Lower Upper
91 100
126 35.8 17.2 51.6



#80
1,3,5-Trimethylbenzene
Concen: 17.93 ug/l
RT: 12.708 min Scan# 3411
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion:105 Resp: 214775
Ion Ratio Lower Upper
105 100
120 48.6 24.6 73.6









#81
trans-1,4-Dichloro-2-butene
Concen: 16.16 ug/l
RT: 12.274 min Scan# 3276
Delta R.T. 0.001 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion	Ratio	Lower	Upper
75	100		
53	100.0	67.0	100.4
89	48.7	39.6	59.4



#82
4-Chlorotoluene
Concen: 18.22 ug/l
RT: 12.746 min Scan# 3423
Delta R.T. 0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion	Ratio	Lower	Upper
91	100		
126	34.3	16.9	50.7









#83
tert-Butylbenzene
Concen: 18.14 ug/l
RT: 12.968 min Scan# 3492
Delta R.T. 0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:119 Resp: 185093
Ion Ratio Lower Upper
119 100
91 64.4 32.3 96.8
134 25.3 12.6 37.6



#84
1,2,4-Trimethylbenzene
Concen: 18.19 ug/l
RT: 13.013 min Scan# 3506
Delta R.T. 0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion:105 Resp: 212051
Ion Ratio Lower Upper
105 100
120 45.9 23.1 69.2









#85
 sec-Butylbenzene
 Concen: 17.59 ug/l
 RT: 13.145 min Scan# 3547
 Delta R.T. -0.000 min
 Lab File: VN065277.D
 Acq: 28 Dec 2020 11:29




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:105 Resp: 238092
 Ion Ratio Lower Upper
 105 100
 134 20.9 10.1 30.1



#86
 p-Isopropyltoluene
 Concen: 17.92 ug/l
 RT: 13.261 min Scan# 3583
 Delta R.T. -0.003 min
 Lab File: VN065277.D
 Acq: 28 Dec 2020 11:29


Tgt Ion:119 Resp: 222349
 Ion Ratio Lower Upper
 119 100
 134 26.6 13.1 39.1
 91 23.0 11.5 34.5








#87
1,3-Dichlorobenzene
Concen: 18.57 ug/l
RT: 13.254 min Scan# 3581
Delta R.T. -0.003 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:146 Resp: 113982
Ion Ratio Lower Upper
146 100
111 36.9 20.6 61.9
148 63.8 31.9 95.5



#88
1,4-Dichlorobenzene
Concen: 18.29 ug/l
RT: 13.338 min Scan# 3607
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion:146 Resp: 112720
Ion Ratio Lower Upper
146 100
111 37.6 19.8 59.3
148 64.1 31.9 95.5









#89
n-Butylbenzene
Concen: 17.69 ug/l
RT: 13.589 min Scan# 3685
Delta R.T. 0.001 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 187631
Ion Ratio Lower Upper
91 100
92 52.5 26.0 78.0
134 26.6 12.9 38.6



#90
Hexachloroethane
Concen: 17.48 ug/l
RT: 13.846 min Scan# 3765
Delta R.T. -0.003 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion: 117 Resp: 37285
Ion Ratio Lower Upper
117 100
201 94.6 40.1 120.3









#91
1,2-Dichlorobenzene
Concen: 18.81 ug/l
RT: 13.627 min Scan# 3697
Delta R.T. 0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:146 Resp: 110894
Ion Ratio Lower Upper
146 100
111 38.3 21.1 63.1
148 63.5 32.0 96.2



#92
1,2-Dibromo-3-Chloropropane
Concen: 17.31 ug/l
RT: 14.245 min Scan# 3889
Delta R.T. 0.004 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion: 75 Resp: 13505
Ion Ratio Lower Upper
75 100
155 91.1 42.8 128.3
157 118.8 56.4 169.1









#93
1,2,4-Trichlorobenzene
Concen: 17.39 ug/l
RT: 14.881 min Scan# 4087
Delta R.T. 0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:180 Resp: 56326
Ion Ratio Lower Upper
180 100
182 96.2 47.6 142.9
145 28.7 15.3 45.8



#94
Hexachlorobutadiene
Concen: 17.28 ug/l
RT: 14.978 min Scan# 4117
Delta R.T. -0.003 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion:225 Resp: 40200
Ion Ratio Lower Upper
225 100
223 61.9 31.0 93.0
227 63.0 32.0 96.0









#95
Naphthalene
Concen: 17.36 ug/l
RT: 15.097 min Scan# 4154
Delta R.T. 0.001 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion:128 Resp: 143313
Ion Ratio Lower Upper
128 100
127 13.3 10.3 15.5
129 10.8 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 18.10 ug/l
RT: 15.264 min Scan# 4206
Delta R.T. -0.000 min
Lab File: VN065277.D
Acq: 28 Dec 2020 11:29

Tgt Ion:180 Resp: 53853
Ion Ratio Lower Upper
180 100
182 93.0 48.1 144.4
145 29.4 16.4 49.1

