

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122820\
 Data File : VN065293.D
 Acq On : 28 Dec 2020 18:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 29 02:24:24 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
 Quant Title : Sw846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	50.000	48.338	3.3	92	0.00
3 P	Chloromethane	50.000	49.863	0.3	97	0.00
4 C	Vinyl Chloride	50.000	51.862	-3.7#	100	0.00
5 T	Bromomethane	50.000	35.402	29.2#	70	0.00
6 T	Chloroethane	50.000	50.882	-1.8	98	0.00
7 T	Trichlorofluoromethane	50.000	46.753	6.5	92	0.00
8 T	Diethyl Ether	50.000	47.803	4.4	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	47.059	5.9	94	0.00
10 T	Methyl Iodide	50.000	34.592	30.8#	66	0.00
11 T	Tert butyl alcohol	250.000	228.493	8.6	89	0.00
12 CM	1,1-Dichloroethene	50.000	46.676	6.6#	96	0.00
13 T	Acrolein	250.000	209.214	16.3	83	0.00
14 T	Allyl chloride	50.000	46.763	6.5	92	0.00
15 T	Acrylonitrile	250.000	258.113	-3.2	100	0.00
16 T	Acetone	250.000	196.698	21.3	78	0.00
17 T	Carbon Disulfide	50.000	47.083	5.8	91	0.00
18 T	Methyl Acetate	50.000	51.646	-3.3	101	0.00
19 T	Methyl tert-butyl Ether	50.000	48.512	3.0	97	0.00
20 T	Methylene Chloride	50.000	45.896	8.2	99	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.290	1.4	97	0.00
22 T	Diisopropyl ether	50.000	49.095	1.8	99	0.00
23 T	Vinyl Acetate	250.000	248.599	0.6	96	0.00
24 P	1,1-Dichloroethane	50.000	50.387	-0.8	99	0.00
25 T	2-Butanone	250.000	241.552	3.4	94	0.00
26 T	2,2-Dichloropropane	50.000	44.531	10.9	90	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.935	-1.9	101	0.00
28 T	Bromochloromethane	50.000	46.525	7.0	100	0.00
29 T	Tetrahydrofuran	250.000	250.471	-0.2	100	0.00
30 C	Chloroform	50.000	49.822	0.4#	99	0.00
31 T	Cyclohexane	50.000	46.219	7.6	96	0.00
32 T	1,1,1-Trichloroethane	50.000	48.188	3.6	95	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.953	0.1	100	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	100	0.00
35 S	Dibromofluoromethane	50.000	52.329	-4.7	104	0.00
36 T	1,1-Dichloropropene	50.000	49.584	0.8	98	0.00
37 T	Ethyl Acetate	50.000	50.133	-0.3	98	0.00
38 T	Carbon Tetrachloride	50.000	48.393	3.2	96	0.00
39 T	Methylcyclohexane	50.000	47.893	4.2	93	0.00
40 TM	Benzene	50.000	50.717	-1.4	101	0.00
41 T	Methacrylonitrile	50.000	48.182	3.6	100	0.00
42 TM	1,2-Dichloroethane	50.000	48.104	3.8	97	0.00
43 T	Isopropyl Acetate	50.000	48.833	2.3	95	0.00
44 TM	Trichloroethene	50.000	49.907	0.2	99	0.00
45 C	1,2-Dichloropropane	50.000	50.974	-1.9#	100	0.00
46 T	Dibromomethane	50.000	50.497	-1.0	101	0.00
47 T	Bromodichloromethane	50.000	49.178	1.6	97	0.00
48 T	Methyl methacrylate	50.000	49.345	1.3	98	0.00
49 T	1,4-Dioxane	1000.000	981.194	1.9	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122820\
 Data File : VN065293.D
 Acq On : 28 Dec 2020 18:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 29 02:24:24 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
 Quant Title : Sw846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

50 S	Toluene-d8	50.000	52.186	-4.4	104	0.00
51 T	4-Methyl-2-Pentanone	250.000	252.137	-0.9	98	0.00
52 CM	Toluene	50.000	50.845	-1.7#	100	0.00
53 T	t-1,3-Dichloropropene	50.000	48.388	3.2	95	0.00
54 T	cis-1,3-Dichloropropene	50.000	48.858	2.3	98	0.00
55 T	1,1,2-Trichloroethane	50.000	50.864	-1.7	100	0.00
56 T	Ethyl methacrylate	50.000	49.863	0.3	98	0.00
57 T	1,3-Dichloropropane	50.000	51.334	-2.7	100	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	245.499	1.8	98	0.00
59 T	2-Hexanone	250.000	248.684	0.5	95	0.00
60 T	Dibromochloromethane	50.000	49.963	0.1	98	0.00
61 T	1,2-Dibromoethane	50.000	50.981	-2.0	100	0.00
62 S	4-Bromofluorobenzene	50.000	48.626	2.7	103	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	102	0.00
64 T	Tetrachloroethene	50.000	46.395	7.2	97	0.00
65 PM	Chlorobenzene	50.000	49.694	0.6	100	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.874	0.3	100	0.00
67 C	Ethyl Benzene	50.000	49.321	1.4#	99	0.00
68 T	m/p-Xylenes	100.000	98.705	1.3	99	0.00
69 T	o-Xylene	50.000	49.904	0.2	100	0.00
70 T	Styrene	50.000	50.238	-0.5	99	0.00
71 P	Bromoform	50.000	49.560	0.9	98	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	101	0.00
73 T	Isopropylbenzene	50.000	47.400	5.2	99	0.00
74 T	N-aryl acetate	50.000	47.910	4.2	95	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.759	-1.5	105	0.00
76 T	1,2,3-Trichloropropane	50.000	49.068	1.9	106	0.00
77 T	Bromobenzene	50.000	49.437	1.1	102	0.00
78 T	n-propylbenzene	50.000	46.793	6.4	96	0.00
79 T	2-Chlorotoluene	50.000	47.793	4.4	97	0.00
80 T	1,3,5-Trimethylbenzene	50.000	47.371	5.3	98	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	44.833	10.3	90	0.00
82 T	4-Chlorotoluene	50.000	47.848	4.3	98	0.00
83 T	tert-Butylbenzene	50.000	47.169	5.7	97	0.00
84 T	1,2,4-Trimethylbenzene	50.000	48.025	4.0	98	0.00
85 T	sec-Butylbenzene	50.000	46.548	6.9	96	0.00
86 T	p-Isopropyltoluene	50.000	46.582	6.8	96	0.00
87 T	1,3-Dichlorobenzene	50.000	49.325	1.3	99	0.00
88 T	1,4-Dichlorobenzene	50.000	48.588	2.8	98	0.00
89 T	n-Butylbenzene	50.000	45.617	8.8	92	0.00
90 T	Hexachloroethane	50.000	47.311	5.4	93	0.00
91 T	1,2-Dichlorobenzene	50.000	50.489	-1.0	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.084	1.8	95	0.00
93 T	1,2,4-Trichlorobenzene	50.000	46.017	8.0	97	0.00
94 T	Hexachlorobutadiene	50.000	41.983	16.0	90	0.00
95 T	Naphthalene	50.000	45.414	9.2	98	0.00
96 T	1,2,3-Trichlorobenzene	50.000	45.049	9.9	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122820\
Data File : VN065293.D
Acq On : 28 Dec 2020 18:10
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Dec 29 02:24:24 2020
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
Quant Title : Sw846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6