

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122821\
 Data File : VN070329.D
 Acq On : 28 Dec 2021 17:03
 Operator : JC/MD
 Sample : M5188-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SASP2-119-1-COMP

Quant Time: Dec 28 17:27:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	842523	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1466509	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1327672	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	425524	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	765045	63.154	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	126.300%
35) Dibromofluoromethane	8.024	113	535515	60.353	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	120.700%
50) Toluene-d8	10.443	98	2025634	55.883	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	111.760%
62) 4-Bromofluorobenzene	12.731	95	604011	46.136	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.280%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	399831	50.845	ug/l	99
20) Methylene Chloride	5.098	84	132184	12.149	ug/l #	85
25) 2-Butanone	7.345	43	31456	3.177	ug/l	94
43) Isopropyl Acetate	8.563	43	145061	4.836	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122821\
Data File : VN070329.D
Acq On : 28 Dec 2021 17:03
Operator : JC/MD
Sample : M5188-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SASP2-119-1-COMP

Quant Time: Dec 28 17:27:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

