

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122821\
 Data File : VN070335.D
 Acq On : 28 Dec 2021 19:36
 Operator : JC/MD
 Sample : M5198-11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-01-VOC

Quant Time: Dec 28 20:31:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1028976	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1767881	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1603805	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	549296	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	790847	53.455	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 106.900%	
35) Dibromofluoromethane	8.024	113	554556	51.844	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.680%	
50) Toluene-d8	10.443	98	2154271	49.301	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 98.600%	
62) 4-Bromofluorobenzene	12.731	95	673103	42.649	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 85.300%	
Target Compounds						
					Qvalue	
16) Acetone	4.291	43	188117	19.587	ug/l	97
20) Methylene Chloride	5.095	84	145015	10.913	ug/l	87
25) 2-Butanone	7.351	43	26186	2.165	ug/l	96
43) Isopropyl Acetate	8.560	43	142477	3.940	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122821\
Data File : VN070335.D
Acq On : 28 Dec 2021 19:36
Operator : JC/MD
Sample : M5198-11
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-01-VOC

Quant Time: Dec 28 20:31:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

