

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122822\
 Data File : VN076002.D
 Acq On : 28 Dec 2022 16:09
 Operator : JC\MD
 Sample : N6213-04 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-4

Quant Time: Dec 29 04:53:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	418863	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	652427	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	558573	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	242961	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	210718	46.242	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	92.480%		
35) Dibromofluoromethane	8.177	113	191092	47.295	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.600%		
50) Toluene-d8	10.571	98	742338	49.140	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.280%		
62) 4-Bromofluorobenzene	12.853	95	239822	47.931	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.860%		
Target Compounds					Qvalue	
40) Benzene	8.612	78	227694	13.750	ug/l	100
52) Toluene	10.635	92	123078	11.412	ug/l	99
68) m/p-Xylenes	12.071	106	39057	5.100	ug/l	97
69) o-Xylene	12.400	106	35865	4.732	ug/l	99
84) 1,2,4-Trimethylbenzene	13.488	105	30333	2.081	ug/l	94
95) Naphthalene	15.641	128	251814	20.392	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122822\
Data File : VN076002.D
Acq On : 28 Dec 2022 16:09
Operator : JC\MD
Sample : N6213-04 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-4

Quant Time: Dec 29 04:53:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

