

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122822\
 Data File : VN076011.D
 Acq On : 28 Dec 2022 19:44
 Operator : JC\MD
 Sample : N6220-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-CR

Quant Time: Dec 29 04:57:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	395461	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	620724	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	545056	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	231392	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	196963	45.781	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.560%
35) Dibromofluoromethane	8.171	113	181211	47.141	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.280%
50) Toluene-d8	10.571	98	715318	49.770	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.540%
62) 4-Bromofluorobenzene	12.853	95	228806	48.065	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.140%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	84879	57.232	ug/l	100
20) Methylene Chloride	5.294	84	22429	5.078	ug/l	92
25) 2-Butanone	7.494	43	30777	14.821	ug/l	97
37) Ethyl Acetate	7.577	43	37107	9.047	ug/l	99
43) Isopropyl Acetate	8.694	43	246068	36.368	ug/l #	80
95) Naphthalene	15.641	128	18113	3.553	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122822\
Data File : VN076011.D
Acq On : 28 Dec 2022 19:44
Operator : JC\MD
Sample : N6220-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-CR

Quant Time: Dec 29 04:57:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

