

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122822\
 Data File : VN076023.D
 Acq On : 29 Dec 2022 01:19
 Operator : JC\MD
 Sample : PB149923ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149923ZHE#04

Quant Time: Dec 29 07:49:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	378086	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	603617	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	533247	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	227164	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	194780	47.355	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.700%
35) Dibromofluoromethane	8.177	113	175295	46.894	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.780%
50) Toluene-d8	10.571	98	705286	50.463	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.920%
62) 4-Bromofluorobenzene	12.853	95	221559	47.862	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.720%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	69176	48.787	ug/l	93
20) Methylene Chloride	5.295	84	18405	4.267	ug/l	89
25) 2-Butanone	7.501	43	23778	11.977	ug/l	97
37) Ethyl Acetate	7.571	43	34324	8.605	ug/l	98
43) Isopropyl Acetate	8.695	43	361477	54.939	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122822\
Data File : VN076023.D
Acq On : 29 Dec 2022 01:19
Operator : JC\MD
Sample : PB149923ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149923ZHE#04

Quant Time: Dec 29 07:49:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

