

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122822\
 Data File : VN076025.D
 Acq On : 29 Dec 2022 02:07
 Operator : JC\MD
 Sample : PB149923ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149923ZHE#06

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 12/29/2022
 Supervised By : Mahesh Dadoda 12/31/2022

Quant Time: Dec 29 07:49:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	390172	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	610993	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	538573	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	226946	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	194135	45.736	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	91.480%		
35) Dibromofluoromethane	8.177	113	178245	47.108	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.220%		
50) Toluene-d8	10.571	98	704974	49.831	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.660%		
62) 4-Bromofluorobenzene	12.853	95	228291	48.721	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.440%		
Target Compounds						Qvalue
16) Acetone	4.442	43	64121	43.821	ug/l	95
20) Methylene Chloride	5.294	84	14506	3.105	ug/l	88
25) 2-Butanone	7.500	43	20484	9.998	ug/l	95
37) Ethyl Acetate	7.565	43	36572	9.058	ug/l	97
43) Isopropyl Acetate	8.694	43	333252m	50.038	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122822\
Data File : VN076025.D
Acq On : 29 Dec 2022 02:07
Operator : JC\MD
Sample : PB149923ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149923ZHE#06

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 12/29/2022
Supervised By : Mahesh Dadoda 12/31/2022

Quant Time: Dec 29 07:49:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

