

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122822\
 Data File : VN076034.D
 Acq On : 29 Dec 2022 05:43
 Operator : JC\MD
 Sample : PB149923ZHE#15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149923ZHE#15

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 12/29/2022
 Supervised By : Mahesh Dadoda 12/31/2022

Quant Time: Dec 29 07:51:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	365447	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	579141	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	530815	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	224183	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	181711	45.705	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	91.420%		
35) Dibromofluoromethane	8.177	113	171648	47.859	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.720%		
50) Toluene-d8	10.570	98	678789	50.619	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	101.240%		
62) 4-Bromofluorobenzene	12.853	95	218588	49.216	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	98.440%		
Target Compounds						Qvalue
16) Acetone	4.447	43	64928	47.375	ug/l	99
20) Methylene Chloride	5.283	84	17156	4.091	ug/l #	89
25) 2-Butanone	7.494	43	22719	11.839	ug/l	98
37) Ethyl Acetate	7.571	43	35765	9.346	ug/l	97
43) Isopropyl Acetate	8.694	43	302206m	47.872	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122822\
Data File : VN076034.D
Acq On : 29 Dec 2022 05:43
Operator : JC\MD
Sample : PB149923ZHE#15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149923ZHE#15

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/29/2022
Supervised By :Mahesh Dadoda 12/31/2022

Quant Time: Dec 29 07:51:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

