

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122920\
 Data File : VN065295.D
 Acq On : 29 Dec 2020 9:46
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 30 00:52:50 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
 Quant Title : Sw846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	103	0.00
2 T	Dichlorodifluoromethane	50.000	48.733	2.5	94	0.00
3 P	Chloromethane	50.000	50.071	-0.1	99	0.00
4 C	Vinyl Chloride	50.000	51.541	-3.1#	100	0.00
5 T	Bromomethane	50.000	46.521	7.0	93	0.00
6 T	Chloroethane	50.000	50.696	-1.4	99	0.00
7 T	Trichlorofluoromethane	50.000	46.584	6.8	92	0.00
8 T	Diethyl Ether	50.000	49.453	1.1	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	47.940	4.1	97	0.00
10 T	Methyl Iodide	50.000	41.652	16.7	81	0.00
11 T	Tert butyl alcohol	250.000	240.360	3.9	95	0.00
12 CM	1,1-Dichloroethene	50.000	45.985	8.0#	95	0.00
13 T	Acrolein	250.000	183.220	26.7#	73	0.00
14 T	Allyl chloride	50.000	46.730	6.5	93	0.00
15 T	Acrylonitrile	250.000	263.879	-5.6	103	0.00
16 T	Acetone	250.000	258.902	-3.6	103	0.00
17 T	Carbon Disulfide	50.000	47.541	4.9	93	0.00
18 T	Methyl Acetate	50.000	52.305	-4.6	104	0.00
19 T	Methyl tert-butyl Ether	50.000	49.207	1.6	99	0.00
20 T	Methylene Chloride	50.000	46.088	7.8	100	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.543	0.9	99	0.00
22 T	Diisopropyl ether	50.000	49.145	1.7	100	0.00
23 T	Vinyl Acetate	250.000	246.829	1.3	97	0.00
24 P	1,1-Dichloroethane	50.000	50.159	-0.3	100	0.00
25 T	2-Butanone	250.000	257.018	-2.8	101	0.00
26 T	2,2-Dichloropropane	50.000	46.117	7.8	94	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.503	-1.0	101	0.00
28 T	Bromochloromethane	50.000	47.036	5.9	102	0.00
29 T	Tetrahydrofuran	250.000	249.362	0.3	101	0.00
30 C	Chloroform	50.000	49.561	0.9#	99	0.00
31 T	Cyclohexane	50.000	45.837	8.3	96	0.00
32 T	1,1,1-Trichloroethane	50.000	47.318	5.4	95	0.00
33 S	1,2-Dichloroethane-d4	50.000	47.464	5.1	96	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	100	0.00
35 S	Dibromofluoromethane	50.000	50.468	-0.9	100	0.00
36 T	1,1-Dichloropropene	50.000	49.692	0.6	98	0.00
37 T	Ethyl Acetate	50.000	49.697	0.6	97	0.00
38 T	Carbon Tetrachloride	50.000	47.983	4.0	95	0.00
39 T	Methylcyclohexane	50.000	48.847	2.3	94	0.00
40 TM	Benzene	50.000	50.787	-1.6	101	0.00
41 T	Methacrylonitrile	50.000	45.151	9.7	94	0.00
42 TM	1,2-Dichloroethane	50.000	48.506	3.0	97	0.00
43 T	Isopropyl Acetate	50.000	48.952	2.1	95	0.00
44 TM	Trichloroethene	50.000	50.664	-1.3	100	0.00
45 C	1,2-Dichloropropane	50.000	52.034	-4.1#	102	0.00
46 T	Dibromomethane	50.000	50.798	-1.6	101	0.00
47 T	Bromodichloromethane	50.000	49.835	0.3	98	0.00
48 T	Methyl methacrylate	50.000	48.448	3.1	96	0.00
49 T	1,4-Dioxane	1000.000	1069.058	-6.9	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122920\
 Data File : VN065295.D
 Acq On : 29 Dec 2020 9:46
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 30 00:52:50 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
 Quant Title : Sw846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

50 S	Toluene-d8	50.000	50.529	-1.1	101	0.00
51 T	4-Methyl-2-Pentanone	250.000	252.767	-1.1	98	0.00
52 CM	Toluene	50.000	51.337	-2.7#	101	0.00
53 T	t-1,3-Dichloropropene	50.000	49.778	0.4	98	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.418	-0.8	100	0.00
55 T	1,1,2-Trichloroethane	50.000	52.594	-5.2	103	0.00
56 T	Ethyl methacrylate	50.000	50.852	-1.7	100	0.00
57 T	1,3-Dichloropropane	50.000	52.518	-5.0	102	0.00
58 T	2-Chloroethyl vinyl ether	250.000	239.857	4.1	96	0.00
59 T	2-Hexanone	250.000	256.448	-2.6	98	0.00
60 T	Dibromochloromethane	50.000	51.063	-2.1	100	0.00
61 T	1,2-Dibromoethane	50.000	51.903	-3.8	102	0.00
62 S	4-Bromofluorobenzene	50.000	46.983	6.0	99	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	102	0.00
64 T	Tetrachloroethene	50.000	48.699	2.6	101	0.00
65 PM	Chlorobenzene	50.000	50.685	-1.4	102	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.978	-2.0	102	0.00
67 C	Ethyl Benzene	50.000	50.088	-0.2#	101	0.00
68 T	m/p-Xylenes	100.000	101.298	-1.3	102	0.00
69 T	o-Xylene	50.000	50.565	-1.1	101	0.00
70 T	Styrene	50.000	51.855	-3.7	102	0.00
71 P	Bromoform	50.000	50.370	-0.7	99	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	101	0.00
73 T	Isopropylbenzene	50.000	48.416	3.2	101	0.00
74 T	N-aryl acetate	50.000	47.880	4.2	94	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.517	-1.0	104	0.00
76 T	1,2,3-Trichloropropane	50.000	49.477	1.0	106	0.00
77 T	Bromobenzene	50.000	50.093	-0.2	103	0.00
78 T	n-propylbenzene	50.000	48.320	3.4	99	0.00
79 T	2-Chlorotoluene	50.000	48.700	2.6	99	0.00
80 T	1,3,5-Trimethylbenzene	50.000	48.623	2.8	101	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	48.516	3.0	97	0.00
82 T	4-Chlorotoluene	50.000	49.089	1.8	101	0.00
83 T	tert-Butylbenzene	50.000	48.080	3.8	99	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.403	1.2	101	0.00
85 T	sec-Butylbenzene	50.000	47.443	5.1	98	0.00
86 T	p-Isopropyltoluene	50.000	48.101	3.8	99	0.00
87 T	1,3-Dichlorobenzene	50.000	51.362	-2.7	103	0.00
88 T	1,4-Dichlorobenzene	50.000	50.595	-1.2	102	0.00
89 T	n-Butylbenzene	50.000	47.294	5.4	95	0.00
90 T	Hexachloroethane	50.000	48.415	3.2	95	0.00
91 T	1,2-Dichlorobenzene	50.000	51.797	-3.6	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.860	0.3	96	0.00
93 T	1,2,4-Trichlorobenzene	50.000	47.200	5.6	99	0.00
94 T	Hexachlorobutadiene	50.000	43.357	13.3	92	0.00
95 T	Naphthalene	50.000	46.525	7.0	100	0.00
96 T	1,2,3-Trichlorobenzene	50.000	46.483	7.0	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122920\
Data File : VN065295.D
Acq On : 29 Dec 2020 9:46
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Dec 30 00:52:50 2020
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
Quant Title : Sw846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6