

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122920\
 Data File : VN065309.D
 Acq On : 29 Dec 2020 16:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 30 00:57:42 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
 Quant Title : Sw846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	103	0.00
2 T	Dichlorodifluoromethane	0.463	0.434	6.3	91	0.00
3 P	Chloromethane	0.527	0.527	0.0	99	0.00
4 C	Vinyl Chloride	0.548	0.556	-1.5#	99	0.00
5 T	Bromomethane	0.389	0.359	7.7	93	0.00
6 T	Chloroethane	0.354	0.353	0.3	98	0.00
7 T	Trichlorofluoromethane	0.820	0.754	8.0	91	0.00
8 T	Diethyl Ether	0.323	0.299	7.4	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.460	0.438	4.8	97	0.00
10 T	Methyl Iodide	0.694	0.623	10.2	87	0.00
11 T	Tert butyl alcohol	0.092	0.078	15.2	84	0.00
12 CM	1,1-Dichloroethene	0.491	0.445	9.4#	94	0.00
13 T	Acrolein	0.065	0.052	20.0	84	0.00
14 T	Allyl chloride	0.810	0.734	9.4	91	0.00
15 T	Acrylonitrile	0.223	0.215	3.6	95	0.00
16 T	Acetone	0.260	0.181	30.4#	76	0.00
17 T	Carbon Disulfide	1.257	1.165	7.3	91	0.00
18 T	Methyl Acetate	0.507	0.490	3.4	96	0.00
19 T	Methyl tert-butyl Ether	1.560	1.444	7.4	94	0.00
20 T	Methylene Chloride	0.551	0.491	10.9	97	0.00
21 T	trans-1,2-Dichloroethene	0.482	0.469	2.7	97	0.00
22 T	Diisopropyl ether	1.611	1.526	5.3	97	0.00
23 T	Vinyl Acetate	1.358	1.270	6.5	92	0.00
24 P	1,1-Dichloroethane	0.884	0.861	2.6	97	0.00
25 T	2-Butanone	0.317	0.285	10.1	89	0.00
26 T	2,2-Dichloropropane	0.844	0.747	11.5	91	0.00
27 T	cis-1,2-Dichloroethene	0.561	0.552	1.6	99	0.00
28 T	Bromochloromethane	0.379	0.372	1.8	101	0.00
29 T	Tetrahydrofuran	0.205	0.191	6.8	94	0.00
30 C	Chloroform	0.910	0.875	3.8#	97	0.00
31 T	Cyclohexane	0.833	0.761	8.6	96	0.00
32 T	1,1,1-Trichloroethane	0.824	0.766	7.0	93	0.00
33 S	1,2-Dichloroethane-d4	0.564	0.530	6.0	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	102	0.00
35 S	Dibromofluoromethane	0.302	0.303	-0.3	101	0.00
36 T	1,1-Dichloropropene	0.455	0.439	3.5	97	0.00
37 T	Ethyl Acetate	0.451	0.415	8.0	91	0.00
38 T	Carbon Tetrachloride	0.477	0.439	8.0	92	0.00
39 T	Methylcyclohexane	0.546	0.529	3.1	95	0.00
40 TM	Benzene	1.352	1.340	0.9	100	0.00
41 T	Methacrylonitrile	0.234	0.211	9.8	95	0.00
42 TM	1,2-Dichloroethane	0.500	0.456	8.8	93	0.00
43 T	Isopropyl Acetate	0.777	0.703	9.5	89	0.00
44 TM	Trichloroethene	0.390	0.384	1.5	99	0.00
45 C	1,2-Dichloropropane	0.352	0.345	2.0#	98	0.00
46 T	Dibromomethane	0.239	0.231	3.3	98	0.00
47 T	Bromodichloromethane	0.503	0.479	4.8	95	0.00
48 T	Methyl methacrylate	0.377	0.344	8.8	91	0.00
49 T	1,4-Dioxane	0.006	0.006	0.0	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122920\
 Data File : VN065309.D
 Acq On : 29 Dec 2020 16:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 30 00:57:42 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
 Quant Title : Sw846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

50 S	Toluene-d8	1.131	1.137	-0.5	101	0.00
51 T	4-Methyl-2-Pentanone	0.437	0.409	6.4	92	0.00
52 CM	Toluene	0.852	0.854	-0.2#	100	0.00
53 T	t-1,3-Dichloropropene	0.567	0.537	5.3	94	0.00
54 T	cis-1,3-Dichloropropene	0.603	0.571	5.3	96	0.00
55 T	1,1,2-Trichloroethane	0.337	0.335	0.6	99	0.00
56 T	Ethyl methacrylate	0.522	0.498	4.6	95	0.00
57 T	1,3-Dichloropropane	0.557	0.559	-0.4	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.227	0.216	4.8	96	0.00
59 T	2-Hexanone	0.321	0.297	7.5	90	0.00
60 T	Dibromochloromethane	0.402	0.385	4.2	95	0.00
61 T	1,2-Dibromoethane	0.362	0.355	1.9	97	0.00
62 S	4-Bromofluorobenzene	0.424	0.416	1.9	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.486	0.458	5.8	99	0.00
65 PM	Chlorobenzene	1.018	1.011	0.7	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.396	0.387	2.3	98	0.00
67 C	Ethyl Benzene	1.834	1.804	1.6#	100	0.00
68 T	m/p-Xylenes	0.691	0.686	0.7	100	0.00
69 T	o-Xylene	0.671	0.665	0.9	100	0.00
70 T	Styrene	1.119	1.128	-0.8	100	0.00
71 P	Bromoform	0.341	0.323	5.3	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	3.808	3.581	6.0	100	0.00
74 T	N-aryl acetate	1.516	1.375	9.3	91	0.00
75 P	1,1,2,2-Tetrachloroethane	1.027	0.994	3.2	102	0.00
76 T	1,2,3-Trichloropropane	1.020	1.032	-1.2	111	0.00
77 T	Bromobenzene	0.995	0.966	2.9	102	0.00
78 T	n-propylbenzene	4.299	4.058	5.6	99	0.00
79 T	2-Chlorotoluene	2.532	2.389	5.6	98	0.00
80 T	1,3,5-Trimethylbenzene	3.247	3.039	6.4	99	0.00
81 T	trans-1,4-Dichloro-2-butene	0.405	0.362	10.6	91	0.00
82 T	4-Chlorotoluene	2.607	2.485	4.7	100	0.00
83 T	tert-Butylbenzene	2.767	2.587	6.5	98	0.00
84 T	1,2,4-Trimethylbenzene	3.162	3.024	4.4	99	0.00
85 T	sec-Butylbenzene	3.671	3.409	7.1	98	0.00
86 T	p-Isopropyltoluene	3.365	3.160	6.1	98	0.00
87 T	1,3-Dichlorobenzene	1.664	1.654	0.6	102	0.00
88 T	1,4-Dichlorobenzene	1.671	1.622	2.9	100	0.00
89 T	n-Butylbenzene	2.876	2.662	7.4	94	0.00
90 T	Hexachloroethane	0.579	0.542	6.4	94	0.00
91 T	1,2-Dichlorobenzene	1.598	1.611	-0.8	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.212	0.198	6.6	92	0.00
93 T	1,2,4-Trichlorobenzene	0.806	0.920	-14.1	100	0.00
94 T	Hexachlorobutadiene	0.631	0.541	14.3	93	0.00
95 T	Naphthalene	2.105	2.351	-11.7	100	0.00
96 T	1,2,3-Trichlorobenzene	0.764	0.842	-10.2	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122920\
Data File : VN065309.D
Acq On : 29 Dec 2020 16:34
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Dec 30 00:57:42 2020
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
Quant Title : Sw846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area%	Dev(min)
----------	-------	------	------------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6