

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122920\
 Data File : VN065309.D
 Acq On : 29 Dec 2020 16:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 30 00:57:42 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
 Quant Title : Sw846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	103	0.00
2 T	Dichlorodifluoromethane	50.000	46.916	6.2	91	0.00
3 P	Chloromethane	50.000	50.030	-0.1	99	0.00
4 C	Vinyl Chloride	50.000	50.741	-1.5#	99	0.00
5 T	Bromomethane	50.000	46.163	7.7	93	0.00
6 T	Chloroethane	50.000	49.917	0.2	98	0.00
7 T	Trichlorofluoromethane	50.000	46.018	8.0	91	0.00
8 T	Diethyl Ether	50.000	46.405	7.2	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	47.657	4.7	97	0.00
10 T	Methyl Iodide	50.000	44.836	10.3	87	0.00
11 T	Tert butyl alcohol	250.000	211.876	15.2	84	0.00
12 CM	1,1-Dichloroethene	50.000	45.298	9.4#	94	0.00
13 T	Acrolein	250.000	208.921	16.4	84	0.00
14 T	Allyl chloride	50.000	45.316	9.4	91	0.00
15 T	Acrylonitrile	250.000	240.981	3.6	95	0.00
16 T	Acetone	250.000	188.541	24.6	76	0.00
17 T	Carbon Disulfide	50.000	46.360	7.3	91	0.00
18 T	Methyl Acetate	50.000	48.296	3.4	96	0.00
19 T	Methyl tert-butyl Ether	50.000	46.284	7.4	94	0.00
20 T	Methylene Chloride	50.000	44.536	10.9	97	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.655	2.7	97	0.00
22 T	Diisopropyl ether	50.000	47.362	5.3	97	0.00
23 T	Vinyl Acetate	250.000	233.814	6.5	92	0.00
24 P	1,1-Dichloroethane	50.000	48.695	2.6	97	0.00
25 T	2-Butanone	250.000	225.053	10.0	89	0.00
26 T	2,2-Dichloropropane	50.000	44.252	11.5	91	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.221	1.6	99	0.00
28 T	Bromochloromethane	50.000	46.316	7.4	101	0.00
29 T	Tetrahydrofuran	250.000	232.732	6.9	94	0.00
30 C	Chloroform	50.000	48.100	3.8#	97	0.00
31 T	Cyclohexane	50.000	45.639	8.7	96	0.00
32 T	1,1,1-Trichloroethane	50.000	46.481	7.0	93	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.980	6.0	95	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	102	0.00
35 S	Dibromofluoromethane	50.000	50.304	-0.6	101	0.00
36 T	1,1-Dichloropropene	50.000	48.274	3.5	97	0.00
37 T	Ethyl Acetate	50.000	46.081	7.8	91	0.00
38 T	Carbon Tetrachloride	50.000	46.000	8.0	92	0.00
39 T	Methylcyclohexane	50.000	48.400	3.2	95	0.00
40 TM	Benzene	50.000	49.550	0.9	100	0.00
41 T	Methacrylonitrile	50.000	44.964	10.1	95	0.00
42 TM	1,2-Dichloroethane	50.000	45.651	8.7	93	0.00
43 T	Isopropyl Acetate	50.000	45.283	9.4	89	0.00
44 TM	Trichloroethene	50.000	49.226	1.5	99	0.00
45 C	1,2-Dichloropropane	50.000	49.126	1.7#	98	0.00
46 T	Dibromomethane	50.000	48.438	3.1	98	0.00
47 T	Bromodichloromethane	50.000	47.612	4.8	95	0.00
48 T	Methyl methacrylate	50.000	45.587	8.8	91	0.00
49 T	1,4-Dioxane	1000.000	943.347	5.7	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122920\
 Data File : VN065309.D
 Acq On : 29 Dec 2020 16:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 30 00:57:42 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
 Quant Title : Sw846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

50 S	Toluene-d8	50.000	50.285	-0.6	101	0.00
51 T	4-Methyl-2-Pentanone	250.000	234.033	6.4	92	0.00
52 CM	Toluene	50.000	50.090	-0.2#	100	0.00
53 T	t-1,3-Dichloropropene	50.000	47.308	5.4	94	0.00
54 T	cis-1,3-Dichloropropene	50.000	47.333	5.3	96	0.00
55 T	1,1,2-Trichloroethane	50.000	49.654	0.7	99	0.00
56 T	Ethyl methacrylate	50.000	47.649	4.7	95	0.00
57 T	1,3-Dichloropropane	50.000	50.202	-0.4	99	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	237.243	5.1	96	0.00
59 T	2-Hexanone	250.000	231.108	7.6	90	0.00
60 T	Dibromochloromethane	50.000	47.870	4.3	95	0.00
61 T	1,2-Dibromoethane	50.000	49.033	1.9	97	0.00
62 S	4-Bromofluorobenzene	50.000	46.869	6.3	100	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	102	0.00
64 T	Tetrachloroethene	50.000	47.168	5.7	99	0.00
65 PM	Chlorobenzene	50.000	49.674	0.7	100	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	48.833	2.3	98	0.00
67 C	Ethyl Benzene	50.000	49.184	1.6#	100	0.00
68 T	m/p-Xylenes	100.000	99.345	0.7	100	0.00
69 T	o-Xylene	50.000	49.601	0.8	100	0.00
70 T	Styrene	50.000	50.395	-0.8	100	0.00
71 P	Bromoform	50.000	47.320	5.4	94	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	102	0.00
73 T	Isopropylbenzene	50.000	47.018	6.0	100	0.00
74 T	N-aryl acetate	50.000	45.330	9.3	91	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.401	3.2	102	0.00
76 T	1,2,3-Trichloropropane	50.000	50.585	-1.2	111	0.00
77 T	Bromobenzene	50.000	48.528	2.9	102	0.00
78 T	n-propylbenzene	50.000	47.200	5.6	99	0.00
79 T	2-Chlorotoluene	50.000	47.190	5.6	98	0.00
80 T	1,3,5-Trimethylbenzene	50.000	46.795	6.4	99	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	44.752	10.5	91	0.00
82 T	4-Chlorotoluene	50.000	47.664	4.7	100	0.00
83 T	tert-Butylbenzene	50.000	46.749	6.5	98	0.00
84 T	1,2,4-Trimethylbenzene	50.000	47.822	4.4	99	0.00
85 T	sec-Butylbenzene	50.000	46.438	7.1	98	0.00
86 T	p-Isopropyltoluene	50.000	46.950	6.1	98	0.00
87 T	1,3-Dichlorobenzene	50.000	49.700	0.6	102	0.00
88 T	1,4-Dichlorobenzene	50.000	48.544	2.9	100	0.00
89 T	n-Butylbenzene	50.000	46.276	7.4	94	0.00
90 T	Hexachloroethane	50.000	46.849	6.3	94	0.00
91 T	1,2-Dichlorobenzene	50.000	50.390	-0.8	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	46.708	6.6	92	0.00
93 T	1,2,4-Trichlorobenzene	50.000	46.646	6.7	100	0.00
94 T	Hexachlorobutadiene	50.000	42.832	14.3	93	0.00
95 T	Naphthalene	50.000	45.685	8.6	100	0.00
96 T	1,2,3-Trichlorobenzene	50.000	46.152	7.7	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122920\
Data File : VN065309.D
Acq On : 29 Dec 2020 16:34
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Dec 30 00:57:42 2020
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
Quant Title : Sw846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6