

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
 Data File : VN070369.D
 Acq On : 30 Dec 2021 1:38
 Operator : JC/MD
 Sample : M5198-11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Dec 30 03:16:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	432935	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	939180	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	713521	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	162649	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	655672	105.333	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 210.660%#	
35) Dibromofluoromethane	8.024	113	433899	76.357	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 152.720%#	
50) Toluene-d8	10.441	98	984744	42.421	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 84.840%	
62) 4-Bromofluorobenzene	12.731	95	251773	30.029	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 60.060%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.081	85	1297	0.264	ug/l	99
3) Chloromethane	2.303	50	7643	1.144	ug/l	94
4) Vinyl Chloride	2.451	62	2907	0.478	ug/l #	79
5) Bromomethane	2.877	94	4711	1.501	ug/l	86
7) Trichlorofluoromethane	3.384	101	1691	0.231	ug/l	91
10) Methyl Iodide	4.441	142	1690	0.279	ug/l #	67
11) Tert butyl alcohol	5.321	59	5588	5.391	ug/l #	71
13) Acrolein	4.033	56	802	0.954	ug/l #	1
15) Acrylonitrile	5.570	53	1948	0.602	ug/l #	75
16) Acetone	4.293	43	167320	41.407	ug/l	97
17) Carbon Disulfide	4.529	76	11710	0.973	ug/l #	94
18) Methyl Acetate	4.870	43	24946	2.134	ug/l #	87
19) Methyl tert-butyl Ether	5.624	73	7908	0.483	ug/l #	80
20) Methylene Chloride	5.098	84	128909	23.057	ug/l	86
24) 1,1-Dichloroethane	6.383	63	2093	0.207	ug/l #	83
25) 2-Butanone	7.343	43	25823	5.075	ug/l #	84
27) cis-1,2-Dichloroethene	7.327	96	2322	0.417	ug/l	58
28) Bromochloromethane	7.657	49	2767	0.560	ug/l #	72
29) Tetrahydrofuran	7.694	42	2095	0.690	ug/l #	47
30) Chloroform	7.823	83	3969	0.407	ug/l #	83
31) Cyclohexane	8.083	56	11699	1.191	ug/l #	28
32) 1,1,1-Trichloroethane	8.013	97	1754	0.214	ug/l #	44
37) Ethyl Acetate	7.343	43	25823	2.119	ug/l #	68
38) Carbon Tetrachloride	8.086	117	43630	4.861	ug/l #	16
40) Benzene	8.461	78	9047	0.314	ug/l	95
42) 1,2-Dichloroethane	8.531	62	6050	0.559	ug/l	89
43) Isopropyl Acetate	8.563	43	128151	6.671	ug/l #	85
44) Trichloroethene	9.223	130	2118	0.309	ug/l	76
45) 1,2-Dichloropropane	9.486	63	2509	0.323	ug/l #	58
47) Bromodichloromethane	9.765	83	3776	0.394	ug/l #	77
48) Methyl methacrylate	9.547	41	32589	3.556	ug/l #	9
49) 1,4-Dioxane	9.566	88	186	1.387	ug/l #	1
51) 4-Methyl-2-Pentanone	10.333	43	43743	3.624	ug/l	100
52) Toluene	10.508	92	5310	0.304	ug/l	92
53) t-1,3-Dichloropropene	10.717	75	3531	0.340	ug/l #	88
54) cis-1,3-Dichloropropene	10.245	75	157233	14.014	ug/l #	68

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
 Data File : VN070369.D
 Acq On : 30 Dec 2021 1:38
 Operator : JC/MD
 Sample : M5198-11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 35 Sample Multiplier: 1


Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Dec 30 03:16:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

55) 1,1,2-Trichloroethane	10.905	97	2805	0.406	ug/l	94
57) 1,3-Dichloropropane	11.092	76	75567	6.190	ug/l #	43
58) 2-Chloroethyl Vinyl ether	10.036	63	159	18.564	ug/l #	45
59) 2-Hexanone	11.092	43	1392249	158.156	ug/l #	49
60) Dibromochloromethane	11.240	129	2796	0.415	ug/l	95
61) 1,2-Dibromoethane	11.344	107	2643	0.377	ug/l	86
65) Chlorobenzene	11.771	112	7015	0.463	ug/l	95
67) Ethyl Benzene	11.843	91	8561	0.309	ug/l	89
68) m/p-Xylenes	11.950	106	5097	0.505	ug/l	93
69) o-Xylene	12.283	106	3136	0.312	ug/l	95
70) Styrene	12.296	104	6203	0.390	ug/l	96
71) Bromoform	12.463	173	2521	0.629	ug/l #	90
73) Isopropylbenzene	12.578	105	6119	0.405	ug/l	98
74) N-amyl acetate	12.393	43	13802	2.173	ug/l #	77
75) 1,1,2,2-Tetrachloroethane	12.830	83	6272	1.582	ug/l	96
76) 1,2,3-Trichloropropane	12.881	75	7117	4.131	ug/l #	1
77) Bromobenzene	12.862	156	2527	0.729	ug/l	72
78) n-propylbenzene	12.924	91	10123	0.605	ug/l	96
79) 2-Chlorotoluene	13.010	91	5008	0.466	ug/l	88
80) 1,3,5-Trimethylbenzene	13.061	105	6899	0.570	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.734	75	132500	99.530	ug/l #	5
82) 4-Chlorotoluene	13.106	91	8250	0.820	ug/l	86
83) tert-Butylbenzene	13.326	119	2631	0.247	ug/l #	88
84) 1,2,4-Trimethylbenzene	13.367	105	9434	0.809	ug/l	87
85) sec-Butylbenzene	13.503	105	6926	0.479	ug/l	97
86) p-Isopropyltoluene	13.613	119	7160	0.630	ug/l	93
87) 1,3-Dichlorobenzene	13.619	146	6709	1.156	ug/l	94
88) 1,4-Dichlorobenzene	13.619	146	6709	1.172	ug/l	94
89) n-Butylbenzene	13.943	91	12167	1.283	ug/l	95
91) 1,2-Dichlorobenzene	13.991	146	7888	1.386	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	14.603	75	2565	2.931	ug/l	72
93) 1,2,4-Trichlorobenzene	15.265	180	9804	3.656	ug/l	92
94) Hexachlorobutadiene	15.373	225	1476	0.947	ug/l	78
95) Naphthalene	15.509	128	67579	10.588	ug/l	98
96) 1,2,3-Trichlorobenzene	15.697	180	14007	6.197	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.088 min Scan# 21

Delta R.T. 0.002 min

Lab File: VN070369.D

Acq: 30 Dec 2021 1:38

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:168 Resp: 432935

Ion Ratio Lower Upper

168 100

99 57.8 41.3 61.9



#2

Dichlorodifluoromethane

Concen: 0.264 ug/l

RT: 2.081 min Scan# 34

Delta R.T. 0.008 min

Lab File: VN070369.D




Acq: 30 Dec 2021 1:38


Tgt Ion: 85 Resp: 1297

Ion Ratio Lower Upper

85 100


87 33.1 16.3 48.9








#3
Chloromethane
Concen: 1.144 ug/l
RT: 2.303 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion: 50 Resp: 7643
Ion Ratio Lower Upper
50 100
52 28.3 25.4 38.0




#4
Vinyl Chloride
Concen: 0.478 ug/l
RT: 2.451 min Scan# 172
Delta R.T. 0.003 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38



Tgt Ion: 62 Resp: 2907
Ion Ratio Lower Upper
62 100
64 20.7 25.8 38.6#









#5
Bromomethane
Concen: 1.501 ug/l
RT: 2.877 min Scan# 31
Delta R.T. 0.024 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 94 Resp: 4711
Ion Ratio Lower Upper
94 100
96 79.2 73.8 110.8



#7
Trichlorofluoromethane
Concen: 0.231 ug/l
RT: 3.384 min Scan# 520
Delta R.T. 0.008 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38


Tgt Ion: 101 Resp: 1691
Ion Ratio Lower Upper
101 100
103 55.7 50.1 75.1








#10
Methyl Iodide
Concen: 0.279 ug/l
RT: 4.441 min Scan# 914
Delta R.T. 0.019 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:142 Resp: 1690
Ion Ratio Lower Upper
142 100
127 23.3 35.4 53.0#
141 0.0 11.6 17.4#



#11
Tert butyl alcohol
Concen: 5.391 ug/l
RT: 5.321 min Scan# 1242
Delta R.T. -0.056 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38


Tgt Ion: 59 Resp: 5588
Ion Ratio Lower Upper
59 100
57 0.0 8.7 13.1#








#13
Acrolein
Concen: 0.954 ug/l
RT: 4.033 min Scan# 765
Delta R.T. -0.008 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 56 Resp: 802
Ion Ratio Lower Upper
56 100
55 210.8 56.6 85.0#



#15
Acrylonitrile
Concen: 0.602 ug/l
RT: 5.570 min Scan# 1335
Delta R.T. 0.016 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Tgt Ion: 53 Resp: 1948
Ion Ratio Lower Upper
53 100
52 61.8 66.2 99.2#
51 19.5 29.6 44.4#









#16
Acetone
Concen: 41.407 ug/l
RT: 4.293 min Scan# 859
Delta R.T. 0.008 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 167320
Ion Ratio Lower Upper
43 100
58 27.6 23.1 34.7



#17
Carbon Disulfide
Concen: 0.973 ug/l
RT: 4.529 min Scan# 947
Delta R.T. 0.000 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38


Tgt Ion: 76 Resp: 11710
Ion Ratio Lower Upper
76 100
78 11.4 7.5 11.3#








#18
Methyl Acetate
Concen: 2.134 ug/l
RT: 4.870 min Scan# 1068
Delta R.T. 0.016 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 43 Resp: 24946
Ion Ratio Lower Upper
43 100
74 18.1 19.4 29.2#



#19
Methyl tert-butyl Ether
Concen: 0.483 ug/l
RT: 5.624 min Scan# 1355
Delta R.T. 0.008 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38


Tgt Ion: 73 Resp: 7908
Ion Ratio Lower Upper
73 100
57 30.5 16.8 25.2#








#20
Methylene Chloride
Concen: 23.057 ug/l
RT: 5.098 min Scan# 1158
Delta R.T. 0.003 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:	84	Resp:	128909
Ion	Ratio	Lower	Upper
84	100		
49	140.3	93.5	140.3
51	41.9	29.4	44.2
86	62.2	53.0	79.4



#24
1,1-Dichloroethane
Concen: 0.207 ug/l
RT: 6.383 min Scan# 1638
Delta R.T. -0.005 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Tgt Ion:	63	Resp:	2093
Ion	Ratio	Lower	Upper
63	100		
98	0.0	3.3	9.9#
100	0.0	2.1	6.5#





#25

2-Butanone

Concen: 5.075 ug/l

RT: 7.343 min Scan# 1996

Delta R.T. 0.008 min

Lab File: VN070369.D

Acq: 30 Dec 2021 1:38

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 43 Resp: 25823

Ion Ratio Lower Upper

43 100

72 18.6 21.7 32.5#



#27

cis-1,2-Dichloroethene

Concen: 0.417 ug/l

RT: 7.327 min Scan# 1990

Delta R.T. 0.003 min

Lab File: VN070369.D

Acq: 30 Dec 2021 1:38




Tgt Ion: 96 Resp: 2322


Ion Ratio Lower Upper

96 100

61 189.6 0.0 275.0

98 33.2 0.0 129.6









#28
Bromochloromethane
Concen: 0.560 ug/l
RT: 7.657 min Scan# 2113
Delta R.T. -0.000 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 49 Resp: 2767
Ion Ratio Lower Upper
49 100
129 0.0 0.0 4.0
130 59.1 67.8 101.8#



#29
Tetrahydrofuran
Concen: 0.690 ug/l
RT: 7.694 min Scan# 2127
Delta R.T. 0.011 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Tgt Ion: 42 Resp: 2095
Ion Ratio Lower Upper
42 100
72 94.4 36.3 54.5#
71 62.9 34.7 52.1#









#30
Chloroform
Concen: 0.407 ug/l
RT: 7.823 min Scan# 2175
Delta R.T. 0.006 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 3969
Ion Ratio Lower Upper
83 100
85 51.3 51.9 77.9#



#31
Cyclohexane
Concen: 1.191 ug/l
RT: 8.083 min Scan# 2272
Delta R.T. -0.011 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Tgt Ion: 56 Resp: 11699
Ion Ratio Lower Upper
56 100
69 154.6 27.1 40.7#
84 111.6 73.4 110.0#









#32
1,1,1-Trichloroethane
Concen: 0.214 ug/l
RT: 8.013 min Scan# 21
Delta R.T. 0.000 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 97 Resp: 1754
Ion Ratio Lower Upper
97 100
99 0.0 50.1 75.1#
61 57.1 35.2 52.8#



#33
1,2-Dichloroethane-d4
Concen: 105.333 ug/l
RT: 8.440 min Scan# 2405
Delta R.T. 0.003 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Tgt Ion: 65 Resp: 655672
Ion Ratio Lower Upper
65 100
67 52.4 0.0 103.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.968 min Scan# 21
Delta R.T. 0.000 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:114 Resp: 939180
Ion Ratio Lower Upper
114 100
63 23.5 0.0 37.8
88 16.0 0.0 27.4



#35
Dibromofluoromethane
Concen: 76.357 ug/l
RT: 8.024 min Scan# 2250
Delta R.T. 0.003 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Tgt Ion:113 Resp: 433899
Ion Ratio Lower Upper
113 100
111 102.6 82.2 123.2
192 19.3 16.9 25.3





#37

Ethyl Acetate

Concen: 2.119 ug/l

RT: 7.343 min Scan# 1996

Delta R.T. -0.069 min

Lab File: VN070369.D





Acq: 30 Dec 2021 1:38

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion:	43	Resp:	25823
Ion Ratio	Lower	Upper	
43	100		
61	0.0	11.5	17.3#
70	0.0	7.9	11.9#



#38

Carbon Tetrachloride

Concen: 4.861 ug/l




RT: 8.086 min Scan# 2273


Delta R.T. -0.120 min

Lab File: VN070369.D

Acq: 30 Dec 2021 1:38

Tgt Ion:	117	Resp:	43630
Ion Ratio	Lower	Upper	
117	100		
119	4.6	76.1	114.1#
121	0.0	24.2	36.2#









#40
Benzene
Concen: 0.314 ug/l
RT: 8.461 min Scan# 2413
Delta R.T. 0.003 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 78 Resp: 9047
Ion Ratio Lower Upper
78 100
77 25.9 18.6 27.8



#42
1,2-Dichloroethane
Concen: 0.559 ug/l
RT: 8.531 min Scan# 2439
Delta R.T. 0.003 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38


Tgt Ion: 62 Resp: 6050
Ion Ratio Lower Upper
62 100
98 4.7 0.0 17.6








#43
Isopropyl Acetate
Concen: 6.671 ug/l
RT: 8.563 min Scan# 2450
Delta R.T. 0.003 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 43 Resp: 128151
Ion Ratio Lower Upper
43 100
61 16.2 20.4 30.6#
87 10.9 11.2 16.8#



#44
Trichloroethene
Concen: 0.309 ug/l
RT: 9.223 min Scan# 2697
Delta R.T. 0.008 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Tgt Ion:130 Resp: 2118
Ion Ratio Lower Upper
130 100
95 72.1 0.0 191.2





#45

1,2-Dichloropropane

Concen: 0.323 ug/l

RT: 9.486 min Scan# 21

Delta R.T. -0.002 min


Lab File: VN070369.D

Acq: 30 Dec 2021 1:38

Instrument :

MSVOA_N

ClientSampleId :






Tgt Ion: 63 Resp: 2509

Ion Ratio Lower Upper

63 100

65 53.7 24.7 37.1#



#47

Bromodichloromethane

Concen: 0.394 ug/l

RT: 9.765 min Scan# 2899

Delta R.T. 0.003 min

Lab File: VN070369.D

Acq: 30 Dec 2021 1:38




Tgt Ion: 83 Resp: 3776


Ion Ratio Lower Upper

83 100

85 42.8 50.9 76.3#

127 8.6 7.1 10.7









#48
Methyl methacrylate
Concen: 3.556 ug/l
RT: 9.547 min Scan# 21
Delta R.T. -0.011 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 41 Resp: 32589
Ion Ratio Lower Upper
41 100
69 1.9 72.4 108.6#
39 0.0 58.4 87.6#



#49
1,4-Dioxane
Concen: 1.387 ug/l
RT: 9.566 min Scan# 2825
Delta R.T. -0.003 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Tgt Ion: 88 Resp: 186
Ion Ratio Lower Upper
88 100
43 0.0 23.4 35.2#
58 2875.8 50.6 75.8#





#50

Toluene-d8

Concen: 42.421 ug/l

RT: 10.441 min Scan# 3151

Delta R.T. 0.001 min

Lab File: VN070369.D

Acq: 30 Dec 2021 1:38

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 98 Resp: 984744

Ion Ratio Lower Upper

98 100

100 64.0 50.4 75.6



#51

4-Methyl-2-Pentanone

Concen: 3.624 ug/l

RT: 10.333 min Scan# 3111

Delta R.T. 0.003 min

Lab File: VN070369.D




Acq: 30 Dec 2021 1:38


Tgt Ion: 43 Resp: 43743

Ion Ratio Lower Upper

43 100

58 37.8 30.4 45.6









#52
Toluene
Concen: 0.304 ug/l
RT: 10.508 min Scan# 3175
Delta R.T. 0.003 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 92 Resp: 5310
Ion Ratio Lower Upper
92 100
91 166.8 142.0 213.0



#53
t-1,3-Dichloropropene
Concen: 0.340 ug/l
RT: 10.717 min Scan# 3254
Delta R.T. -0.000 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38


Tgt Ion: 75 Resp: 3531
Ion Ratio Lower Upper
75 100
77 38.7 25.6 38.4#








#54
 cis-1,3-Dichloropropene
 Concen: 14.014 ug/l
 RT: 10.245 min Scan# 3057
 Delta R.T. 0.057 min
 Lab File: VN070369.D
 Acq: 30 Dec 2021 1:38

Instrument :
 MSVOA_N
 ClientSampleId :







Tgt Ion: 75 Resp: 157233
 Ion Ratio Lower Upper
 75 100
 77 0.6 26.4 39.6#
 39 41.4 42.5 63.7#



#55
 1,1,2-Trichloroethane
 Concen: 0.406 ug/l
 RT: 10.905 min Scan# 3324
 Delta R.T. 0.008 min
 Lab File: VN070369.D
 Acq: 30 Dec 2021 1:38

Tgt Ion: 97 Resp: 2805
 Ion Ratio Lower Upper
 97 100
 83 83.9 67.8 101.6
 85 64.9 45.4 68.2
 99 55.8 49.4 74.0





#57

1,3-Dichloropropane

Concen: 6.190 ug/l

RT: 11.092 min Scan# 31

Delta R.T. 0.048 min

Lab File: VN070369.D

Acq: 30 Dec 2021 1:38

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 76 Resp: 75567

Ion Ratio Lower Upper

76 100

78 0.5 25.7 38.5#



#58

2-Chloroethyl Vinyl ether

Concen: 18.564 ug/l

RT: 10.036 min Scan# 3000

Delta R.T. -0.005 min

Lab File: VN070369.D




Acq: 30 Dec 2021 1:38


Tgt Ion: 63 Resp: 159

Ion Ratio Lower Upper

63 100

106 0.0 23.9 35.9#









#59
2-Hexanone
Concen: 158.156 ug/l
RT: 11.092 min Scan# 31
Delta R.T. 0.008 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 1392249
Ion Ratio Lower Upper
43 100
58 17.5 26.9 80.7#



#60
Dibromochloromethane
Concen: 0.415 ug/l
RT: 11.240 min Scan# 3449
Delta R.T. 0.003 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Tgt Ion:129 Resp: 2796
Ion Ratio Lower Upper
129 100
127 73.3 38.6 116.0









#61
1,2-Dibromoethane
Concen: 0.377 ug/l
RT: 11.344 min Scan# 3488
Delta R.T. 0.000 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:107 Resp: 2643
Ion Ratio Lower Upper
107 100
109 108.1 75.7 113.5



#62
4-Bromofluorobenzene
Concen: 30.029 ug/l
RT: 12.731 min Scan# 4005
Delta R.T. -0.000 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Tgt Ion: 95 Resp: 251773
Ion Ratio Lower Upper
95 100
174 77.2 0.0 187.2
176 73.7 0.0 181.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.744 min Scan# 30

Delta R.T. -0.000 min

Lab File: VN070369.D

Acq: 30 Dec 2021 1:38

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:117 Resp: 713521

Ion Ratio Lower Upper

117 100

82 59.8 42.4 63.6

119 31.8 25.4 38.2



#65

Chlorobenzene

Concen: 0.463 ug/l

RT: 11.771 min Scan# 3647

Delta R.T. 0.003 min

Lab File: VN070369.D




Acq: 30 Dec 2021 1:38


Tgt Ion:112 Resp: 7015

Ion Ratio Lower Upper

112 100

114 29.2 25.4 38.0









#67
Ethyl Benzene
Concen: 0.309 ug/l
RT: 11.843 min Scan# 3674
Delta R.T. 0.000 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 8561
Ion Ratio Lower Upper
91 100
106 26.2 25.7 38.5



#68
m/p-Xylenes
Concen: 0.505 ug/l
RT: 11.950 min Scan# 3714
Delta R.T. -0.003 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Tgt Ion: 106 Resp: 5097
Ion Ratio Lower Upper
106 100
91 211.9 160.6 240.8









#69
o-Xylene
Concen: 0.312 ug/l
RT: 12.283 min Scan# 3837
Delta R.T. 0.005 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:106 Resp: 3136
Ion Ratio Lower Upper
106 100
91 202.9 105.8 317.3



#70
Styrene
Concen: 0.390 ug/l
RT: 12.296 min Scan# 3843
Delta R.T. 0.002 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Tgt Ion:104 Resp: 6203
Ion Ratio Lower Upper
104 100
78 51.9 38.7 58.1
103 58.7 45.0 67.6









#71
Bromoform
Concen: 0.629 ug/l
RT: 12.463 min Scan# 3905
Delta R.T. 0.006 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:173 Resp: 2521
Ion Ratio Lower Upper
173 100
175 41.7 24.2 72.6
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.675 min Scan# 4357
Delta R.T. 0.003 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Tgt Ion:152 Resp: 162649
Ion Ratio Lower Upper
152 100
115 59.3 27.5 82.5
150 159.5 0.0 344.4









#73
Isopropylbenzene
Concen: 0.405 ug/l
RT: 12.578 min Scan# 3948
Delta R.T. 0.000 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:105 Resp: 6119
Ion Ratio Lower Upper
105 100
120 25.8 13.4 40.1



#74
N-ethyl acetate
Concen: 2.173 ug/l
RT: 12.393 min Scan# 3879
Delta R.T. 0.008 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Tgt Ion: 43 Resp: 13802
Ion Ratio Lower Upper
43 100
70 15.9 33.4 50.0#
55 19.1 19.3 28.9#
61 20.2 18.5 27.7









#75
1,1,2,2-Tetrachloroethane
Concen: 1.582 ug/l
RT: 12.830 min Scan# 4040
Delta R.T. 0.005 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 6272
Ion Ratio Lower Upper
83 100
131 9.3 5.5 16.5
85 61.7 32.4 97.2



#76
1,2,3-Trichloropropane
Concen: 4.131 ug/l
RT: 12.881 min Scan# 4061
Delta R.T. 0.003 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Tgt Ion: 75 Resp: 7117
Ion Ratio Lower Upper
75 100
77 0.0 82.6 247.8#





#77

Bromobenzene

Concen: 0.729 ug/l

RT: 12.862 min Scan# 4053

Delta R.T. 0.002 min

Lab File: VN070369.D

Acq: 30 Dec 2021 1:38

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:156 Resp: 2527

Ion Ratio Lower Upper

156 100

77 231.3 90.6 271.8

158 82.9 48.9 146.7



#78

n-propylbenzene

Concen: 0.605 ug/l

RT: 12.924 min Scan# 4077

Delta R.T. 0.005 min

Lab File: VN070369.D




Acq: 30 Dec 2021 1:38


Tgt Ion: 91 Resp: 10123

Ion Ratio Lower Upper

91 100

120 21.7 11.9 35.9









#79
2-Chlorotoluene
Concen: 0.466 ug/l
RT: 13.010 min Scan# 4108
Delta R.T. 0.003 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 5008
Ion Ratio Lower Upper
91 100
126 29.2 18.3 54.8



#80
1,3,5-Trimethylbenzene
Concen: 0.570 ug/l
RT: 13.061 min Scan# 4128
Delta R.T. 0.003 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Tgt Ion: 105 Resp: 6899
Ion Ratio Lower Upper
105 100
120 47.1 24.6 73.6









#81
trans-1,4-Dichloro-2-butene
Concen: 99.530 ug/l
RT: 12.734 min Scan# 4006
Delta R.T. 0.110 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 75 Resp: 132500
Ion Ratio Lower Upper
75 100
53 0.1 90.5 135.7#
89 0.0 38.2 57.4#



#82
4-Chlorotoluene
Concen: 0.820 ug/l
RT: 13.106 min Scan# 4145
Delta R.T. 0.002 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Tgt Ion: 91 Resp: 8250
Ion Ratio Lower Upper
91 100
126 28.4 18.2 54.6









#83
tert-Butylbenzene
Concen: 0.247 ug/l
RT: 13.326 min Scan# 4227
Delta R.T. 0.005 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:119 Resp: 2631
Ion Ratio Lower Upper
119 100
91 68.0 31.5 94.5
134 11.9 12.4 37.4#



#84
1,2,4-Trimethylbenzene
Concen: 0.809 ug/l
RT: 13.367 min Scan# 4242
Delta R.T. 0.001 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Tgt Ion:105 Resp: 9434
Ion Ratio Lower Upper
105 100
120 36.9 22.8 68.4





#85

sec-Butylbenzene

Concen: 0.479 ug/l

RT: 13.503 min Scan# 4292

Delta R.T. 0.002 min

Lab File: VN070369.D

Acq: 30 Dec 2021 1:38

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:105 Resp: 6926

Ion Ratio Lower Upper

105 100

134 19.1 10.4 31.1



#86

p-Isopropyltoluene

Concen: 0.630 ug/l

RT: 13.613 min Scan# 4334

Delta R.T. 0.000 min

Lab File: VN070369.D

Acq: 30 Dec 2021 1:38




Tgt Ion:119 Resp: 7160


Ion Ratio Lower Upper

119 100

134 23.5 13.6 40.7

91 26.0 11.4 34.1





#87

1,3-Dichlorobenzene

Concen: 1.156 ug/l

RT: 13.619 min Scan# 4334

Delta R.T. 0.006 min

Lab File: VN070369.D

Acq: 30 Dec 2021 1:38

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:146 Resp: 6709

Ion Ratio Lower Upper

146 100

111 45.4 18.6 55.6

148 63.8 31.6 95.0



#88

1,4-Dichlorobenzene

Concen: 1.172 ug/l

RT: 13.619 min Scan# 4336

Delta R.T. -0.075 min

Lab File: VN070369.D

Acq: 30 Dec 2021 1:38




Tgt Ion:146 Resp: 6709


Ion Ratio Lower Upper

146 100

111 45.4 18.3 54.8

148 63.8 31.8 95.4









#89
n-Butylbenzene
Concen: 1.283 ug/l
RT: 13.943 min Scan# 4456
Delta R.T. 0.003 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 12167
Ion Ratio Lower Upper
91 100
92 52.0 25.4 76.0
134 21.8 13.6 40.8



#91
1,2-Dichlorobenzene
Concen: 1.386 ug/l
RT: 13.991 min Scan# 4475
Delta R.T. 0.005 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38

Tgt Ion:146 Resp: 7888
Ion Ratio Lower Upper
146 100
111 44.7 19.1 57.3
148 65.4 32.1 96.3









#92
 1,2-Dibromo-3-Chloropropane
 Concen: 2.931 ug/l
 RT: 14.603 min Scan# 41
 Delta R.T. 0.003 min
 Lab File: VN070369.D
 Acq: 30 Dec 2021 1:38




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 75 Resp: 2565
 Ion Ratio Lower Upper
 75 100
 155 69.3 46.9 140.7
 157 88.1 60.6 181.7



#93
 1,2,4-Trichlorobenzene
 Concen: 3.656 ug/l
 RT: 15.265 min Scan# 4950
 Delta R.T. 0.002 min
 Lab File: VN070369.D
 Acq: 30 Dec 2021 1:38

Tgt Ion: 180 Resp: 9804
 Ion Ratio Lower Upper
 180 100
 182 102.5 48.3 144.9
 145 36.1 14.5 43.5









#94
Hexachlorobutadiene
Concen: 0.947 ug/l
RT: 15.373 min Scan# 4988
Delta R.T. 0.006 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:225 Resp: 1476
Ion Ratio Lower Upper
225 100
223 33.5 31.9 95.7
227 59.7 32.1 96.3



#95
Naphthalene
Concen: 10.588 ug/l
RT: 15.509 min Scan# 5041
Delta R.T. 0.005 min
Lab File: VN070369.D
Acq: 30 Dec 2021 1:38


Tgt Ion:128 Resp: 67579
Ion Ratio Lower Upper
128 100
127 12.7 10.5 15.7
129 11.8 8.7 13.1





#96
 1,2,3-Trichlorobenzene
 Concen: 6.197 ug/l
 RT: 15.697 min Scan# 51
 Delta R.T. -0.003 min
 Lab File: VN070369.D
 Acq: 30 Dec 2021 1:38

Instrument :
 MSVOA_N
 ClientSampleId :



Tgt Ion	Ratio	Lower	Upper
180	100		
182	82.5	47.4	142.2
145	16.7	15.0	45.1

