

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
 Data File : VN070342.D
 Acq On : 29 Dec 2021 12:10
 Operator : JC/MD
 Sample : VN1229WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1229WBL01

Quant Time: Dec 30 01:48:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1066080	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1762831	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1577875	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	532310	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	810657	52.887	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	105.780%
35) Dibromofluoromethane	8.021	113	573946	53.811	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.620%
50) Toluene-d8	10.443	98	2190389	50.271	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.540%
62) 4-Bromofluorobenzene	12.733	95	767691	48.782	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
Data File : VN070342.D
Acq On : 29 Dec 2021 12:10
Operator : JC/MD
Sample : VN1229WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1229WBL01

Quant Time: Dec 30 01:48:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

