

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
 Data File : VN070351.D
 Acq On : 29 Dec 2021 16:16
 Operator : JC/MD
 Sample : M5115-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-6

Quant Time: Dec 30 01:51:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	898081	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1531371	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1372779	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	456485	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	682116	52.825	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 105.660%	
35) Dibromofluoromethane	8.024	113	483446	52.177	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.360%	
50) Toluene-d8	10.443	98	1829171	48.326	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 96.660%	
62) 4-Bromofluorobenzene	12.733	95	628260	45.956	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 91.920%	
Target Compounds						
						Qvalue
16) Acetone	4.304	43	13691	1.633	ug/l	97
19) Methyl tert-butyl Ether	5.618	73	76432	2.250	ug/l	99
85) sec-Butylbenzene	13.503	105	22878	0.563	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
Data File : VN070351.D
Acq On : 29 Dec 2021 16:16
Operator : JC/MD
Sample : M5115-08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-6

Quant Time: Dec 30 01:51:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

