

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
 Data File : VN070354.D
 Acq On : 29 Dec 2021 17:32
 Operator : JC/MD
 Sample : M5115-11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-2021-12-15

Quant Time: Dec 30 01:51:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	886317	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1509206	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	1343969	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	431837	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	672587	52.779	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 105.560%			
35) Dibromofluoromethane	8.024	113	479715	52.534	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 105.060%			
50) Toluene-d8	10.443	98	1788571	47.947	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 95.900%			
62) 4-Bromofluorobenzene	12.734	95	609187	45.215	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 90.440%			
Target Compounds						
16) Acetone	4.296	43	22761	2.751	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
Data File : VN070354.D
Acq On : 29 Dec 2021 17:32
Operator : JC/MD
Sample : M5115-11
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-2021-12-15

Quant Time: Dec 30 01:51:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

