

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
 Data File : VN070356.D
 Acq On : 29 Dec 2021 18:23
 Operator : JC/MD
 Sample : M5128-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2

Quant Time: Dec 30 01:51:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	858145	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1461344	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	1300420	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	422134	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	668029	54.142	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.280%	
35) Dibromofluoromethane	8.024	113	468048	52.935	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.880%	
50) Toluene-d8	10.443	98	1765445	48.877	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 97.760%	
62) 4-Bromofluorobenzene	12.734	95	604909	46.368	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 92.740%	
Target Compounds						Qvalue
16) Acetone	4.296	43	30589	3.819	ug/l	97
65) Chlorobenzene	11.771	112	20666	0.749	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
Data File : VN070356.D
Acq On : 29 Dec 2021 18:23
Operator : JC/MD
Sample : M5128-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-2

Quant Time: Dec 30 01:51:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

