

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
 Data File : VN070357.D
 Acq On : 29 Dec 2021 18:49
 Operator : JC/MD
 Sample : M5128-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-3

Quant Time: Dec 30 01:52:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	877424	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1503175	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	1334706	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	420844	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	675278	53.527	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.060%	
35) Dibromofluoromethane	8.024	113	480006	52.777	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.560%	
50) Toluene-d8	10.443	98	1801426	48.486	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 96.980%	
62) 4-Bromofluorobenzene	12.734	95	599723	44.691	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 89.380%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
Data File : VN070357.D
Acq On : 29 Dec 2021 18:49
Operator : JC/MD
Sample : M5128-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-3

Quant Time: Dec 30 01:52:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

