

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
 Data File : VN070365.D
 Acq On : 29 Dec 2021 23:05
 Operator : JC/MD
 Sample : VN1229MBL02
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1229MBL02

Quant Time: Dec 30 01:53:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1021258	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1989179	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2413327	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	889397	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	783713	53.373	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.740%
35) Dibromofluoromethane	8.024	113	598303	49.711	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.420%
50) Toluene-d8	10.443	98	2706134	55.040	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	110.080%
62) 4-Bromofluorobenzene	12.731	95	1114608	62.767	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	125.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
Data File : VN070365.D
Acq On : 29 Dec 2021 23:05
Operator : JC/MD
Sample : VN1229MBL02
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1229MBL02

Quant Time: Dec 30 01:53:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

