

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
 Data File : VN070369.D
 Acq On : 30 Dec 2021 1:38
 Operator : JC/MD
 Sample : M5198-11RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-01-VOCRE

Quant Time: Dec 30 03:16:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	432935	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	939180	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	713521	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	162649	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	655672	105.333	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 210.660%#	
35) Dibromofluoromethane	8.024	113	433899	76.357	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 152.720%#	
50) Toluene-d8	10.441	98	984744	42.421	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 84.840%	
62) 4-Bromofluorobenzene	12.731	95	251773	30.029	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 60.060%	
Target Compounds						
						Qvalue
16) Acetone	4.293	43	167320	41.407	ug/l	97
20) Methylene Chloride	5.098	84	128909	23.057	ug/l	86
43) Isopropyl Acetate	8.563	43	128151	6.671	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
Data File : VN070369.D
Acq On : 30 Dec 2021 1:38
Operator : JC/MD
Sample : M5198-11RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-01-VOCRE

Quant Time: Dec 30 03:16:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

