

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
 Data File : VN070370.D
 Acq On : 30 Dec 2021 2:03
 Operator : JC/MD
 Sample : M5241-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2S

Quant Time: Dec 30 03:17:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	903092	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1570751	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	1441684	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	491794	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	707047	54.452	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.900%	
35) Dibromofluoromethane	8.024	113	489650	51.521	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.040%	
50) Toluene-d8	10.443	98	1941058	49.996	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.000%	
62) 4-Bromofluorobenzene	12.731	95	693928	49.487	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 98.980%	
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	237285	28.151	ug/l	100
20) Methylene Chloride	5.098	84	144508	12.391	ug/l	89
39) Methylcyclohexane	9.461	83	28500	1.497	ug/l #	84
40) Benzene	8.461	78	642887	13.338	ug/l	99
43) Isopropyl Acetate	8.563	43	136325	4.243	ug/l #	87
67) Ethyl Benzene	11.846	91	134088	2.392	ug/l	93
68) m/p-Xylenes	11.958	106	37518	1.839	ug/l	91
73) Isopropylbenzene	12.581	105	207508	4.544	ug/l	98
78) n-propylbenzene	12.921	91	206194	4.075	ug/l	96
85) sec-Butylbenzene	13.503	105	67139	1.535	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
Data File : VN070370.D
Acq On : 30 Dec 2021 2:03
Operator : JC/MD
Sample : M5241-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2S

Quant Time: Dec 30 03:17:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

