

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
 Data File : VN070375.D
 Acq On : 30 Dec 2021 4:11
 Operator : JC/MD
 Sample : M5007-03RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TOLL-MW-04RE

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/30/2021
 Supervised By : Amit Patel 12/30/2021

Quant Time: Dec 30 07:17:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	886012	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1558142	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	1358281	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	453578	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	675343	53.013	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 106.020%			
35) Dibromofluoromethane	8.024	113	478935	50.802	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.600%			
50) Toluene-d8	10.443	98	1743649	45.275	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 90.540%			
62) 4-Bromofluorobenzene	12.731	95	629228	45.236	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 90.480%			
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	53945	6.523	ug/l	98
18) Methyl Acetate	4.867	43	16481m	0.689	ug/l	
19) Methyl tert-butyl Ether	5.621	73	135175	4.034	ug/l	92
27) cis-1,2-Dichloroethene	7.327	96	90871	7.971	ug/l	97
44) Trichloroethene	9.217	130	2952	0.260	ug/l	85
95) Naphthalene	15.507	128	9543	2.242	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
Data File : VN070375.D
Acq On : 30 Dec 2021 4:11
Operator : JC/MD
Sample : M5007-03RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TOLL-MW-04RE

Manual Integrations
APPROVED

Reviewed By : John Carlone 12/30/2021
Supervised By : Amit Patel 12/30/2021

Quant Time: Dec 30 07:17:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

