

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
 Data File : VN070377.D
 Acq On : 30 Dec 2021 5:02
 Operator : JC/MD
 Sample : M5007-09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TOLL-MW-08RE

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/30/2021
 Supervised By : Amit Patel 12/30/2021

Quant Time: Dec 30 07:18:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	877782	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	1550758	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	1347820	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	430736	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	689512	54.633	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 109.260%			
35) Dibromofluoromethane	8.024	113	493994	52.649	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 105.300%			
50) Toluene-d8	10.443	98	1855457	48.408	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 96.820%			
62) 4-Bromofluorobenzene	12.731	95	619305	44.735	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 89.460%			
Target Compounds					Qvalue	
16) Acetone	4.296	43	28869	3.524	ug/l	94
19) Methyl tert-butyl Ether	5.613	73	20985m	0.632	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
Data File : VN070377.D
Acq On : 30 Dec 2021 5:02
Operator : JC/MD
Sample : M5007-09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TOLL-MW-08RE

Quant Time: Dec 30 07:18:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/30/2021
Supervised By :Amit Patel 12/30/2021

