

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
 Data File : VN070379.D
 Acq On : 30 Dec 2021 5:52
 Operator : JC/MD
 Sample : M5007-12
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-2021.12.08RE

Quant Time: Dec 30 07:18:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	854125	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1478763	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	1303927	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	423359	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	664858	54.138	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.280%	
35) Dibromofluoromethane	8.021	113	472168	52.772	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.540%	
50) Toluene-d8	10.443	98	1791289	49.009	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 98.020%	
62) 4-Bromofluorobenzene	12.734	95	613775	46.494	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 92.980%	
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	8571	1.075	ug/l	97
18) Methyl Acetate	4.867	43	12715	0.551	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
Data File : VN070379.D
Acq On : 30 Dec 2021 5:52
Operator : JC/MD
Sample : M5007-12
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-2021.12.08RE

Quant Time: Dec 30 07:18:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

