

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
 Data File : VN070380.D
 Acq On : 30 Dec 2021 6:18
 Operator : JC/MD
 Sample : M5007-13
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-2021.12.08RE

Quant Time: Dec 30 07:19:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	889084	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1539390	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1346092	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	425639	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	669419	52.367	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.740%	
35) Dibromofluoromethane	8.024	113	474114	50.903	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.800%	
50) Toluene-d8	10.443	98	1801116	47.337	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 94.680%	
62) 4-Bromofluorobenzene	12.734	95	598781	43.572	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 87.140%	
Target Compounds						
						Qvalue
16) Acetone	4.296	43	12707	1.531	ug/l	90
18) Methyl Acetate	4.865	43	13939	0.581	ug/l #	59

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
Data File : VN070380.D
Acq On : 30 Dec 2021 6:18
Operator : JC/MD
Sample : M5007-13
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-2021.12.08RE

Quant Time: Dec 30 07:19:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

