

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
 Data File : VN070383.D
 Acq On : 30 Dec 2021 7:34
 Operator : JC/MD
 Sample : M5007-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TOLL-MW-06RE

Quant Time: Dec 30 08:03:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	883762	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1540822	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1339582	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	420815	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	690357	54.330	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 108.660%			
35) Dibromofluoromethane	8.024	113	486894	52.226	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 104.460%			
50) Toluene-d8	10.443	98	1907793	50.094	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 100.180%			
62) 4-Bromofluorobenzene	12.731	95	612563	44.533	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 89.060%			
Target Compounds						
16) Acetone	4.296	43	29065	3.524	ug/l	99
18) Methyl Acetate	4.873	43	11687	0.490	ug/l #	82
19) Methyl tert-butyl Ether	5.624	73	183982	5.504	ug/l #	90
22) Diisopropyl ether	6.498	45	30782	0.774	ug/l #	89
27) cis-1,2-Dichloroethene	7.327	96	15068	1.325	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
Data File : VN070383.D
Acq On : 30 Dec 2021 7:34
Operator : JC/MD
Sample : M5007-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TOLL-MW-06RE

Quant Time: Dec 30 08:03:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

