

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
 Data File : VN070385.D
 Acq On : 30 Dec 2021 8:25
 Operator : JC/MD
 Sample : M4873-07DL 10X
 Misc : 5.18g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 A2-8-6.5DL

Quant Time: Dec 30 09:50:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	991457	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1903157	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2400741	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1058520	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	743754	52.174	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.340%	
35) Dibromofluoromethane	8.021	113	569659	49.471	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 98.940%	
50) Toluene-d8	10.443	98	2610148	55.488	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 110.980%	
62) 4-Bromofluorobenzene	12.731	95	1257956	74.041	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 148.080%#	
Target Compounds						
					Qvalue	
39) Methylcyclohexane	9.459	83	78941	3.423	ug/l #	73
52) Toluene	10.505	92	130840	3.701	ug/l	97
67) Ethyl Benzene	11.843	91	549740	5.890	ug/l	95
68) m/p-Xylenes	11.950	106	800698	23.569	ug/l	93
69) o-Xylene	12.280	106	325910	9.645	ug/l	92
73) Isopropylbenzene	12.578	105	127088	1.293	ug/l	99
78) n-propylbenzene	12.919	91	688320	6.320	ug/l	95
80) 1,3,5-Trimethylbenzene	13.058	105	993968	12.618	ug/l	96
84) 1,2,4-Trimethylbenzene	13.366	105	3310934	43.650	ug/l	98
85) sec-Butylbenzene	13.501	105	91235m	0.969	ug/l	
86) p-Isopropyltoluene	13.616	119	53611	0.725	ug/l	97
89) n-Butylbenzene	13.940	91	153969	2.495	ug/l #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
Data File : VN070385.D
Acq On : 30 Dec 2021 8:25
Operator : JC/MD
Sample : M4873-07DL 10X
Misc : 5.18g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
A2-8-6.5DL

Quant Time: Dec 30 09:50:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

