

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
 Data File : VN070340.D
 Acq On : 29 Dec 2021 10:36
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 30 01:48:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.567	0.600	-5.8	99	0.00
3 P	Chloromethane	0.771	0.778	-0.9	102	0.00
4 C	Vinyl Chloride	0.702	0.709	-1.0#	100	0.00
5 T	Bromomethane	0.362	0.369	-1.9	93	-0.02
6 T	Chloroethane	0.423	0.417	1.4	98	0.00
7 T	Trichlorofluoromethane	0.845	0.842	0.4	98	0.00
8 T	Diethyl Ether	0.373	0.378	-1.3	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.517	0.550	-6.4	103	0.00
10 T	Methyl Iodide	0.700	0.747	-6.7	99	0.00
11 T	Tert butyl alcohol	0.120	0.128	-6.7	105	0.00
12 CM	1,1-Dichloroethene	0.509	0.524	-2.9#	100	0.00
13 T	Acrolein	0.097	0.099	-2.1	103	0.00
14 T	Allyl chloride	1.155	1.225	-6.1	103	0.00
15 T	Acrylonitrile	0.374	0.403	-7.8	104	0.00
16 T	Acetone	0.467	0.511	-9.4	102	0.00
17 T	Carbon Disulfide	1.390	1.561	-12.3	109	0.00
18 T	Methyl Acetate	1.350	1.399	-3.6	103	0.00
19 T	Methyl tert-butyl Ether	1.891	1.966	-4.0	99	0.00
20 T	Methylene Chloride	0.646	0.641	0.8	101	0.00
21 T	trans-1,2-Dichloroethene	0.539	0.559	-3.7	100	0.00
22 T	Diisopropyl ether	2.249	2.389	-6.2	103	0.00
23 T	Vinyl Acetate	1.745	1.913	-9.6	103	0.00
24 P	1,1-Dichloroethane	1.166	1.202	-3.1	101	0.00
25 T	2-Butanone	0.588	0.646	-9.9	104	0.00
26 T	2,2-Dichloropropane	0.920	0.957	-4.0	101	0.00
27 T	cis-1,2-Dichloroethene	0.643	0.675	-5.0	101	0.00
28 T	Bromochloromethane	0.570	0.493	13.5	85	0.00
29 T	Tetrahydrofuran	0.351	0.384	-9.4	104	0.00
30 C	Chloroform	1.125	1.158	-2.9#	101	0.00
31 T	Cyclohexane	1.134	1.163	-2.6	101	0.00
32 T	1,1,1-Trichloroethane	0.948	0.987	-4.1	100	0.00
33 S	1,2-Dichloroethane-d4	0.719	0.657	8.6	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.303	0.299	1.3	92	0.00
36 T	1,1-Dichloropropene	0.500	0.549	-9.8	103	0.00
37 T	Ethyl Acetate	0.649	0.706	-8.8	101	0.00
38 T	Carbon Tetrachloride	0.478	0.532	-11.3	104	0.00
39 T	Methylcyclohexane	0.606	0.682	-12.5	103	0.00
40 TM	Benzene	1.534	1.647	-7.4	101	0.00
41 T	Methacrylonitrile	0.324	0.377	-16.4	105	0.00
42 TM	1,2-Dichloroethane	0.577	0.608	-5.4	101	0.00
43 T	Isopropyl Acetate	1.023	1.126	-10.1	103	0.00
44 TM	Trichloroethene	0.365	0.390	-6.8	102	0.00
45 C	1,2-Dichloropropane	0.413	0.454	-9.9#	102	0.00
46 T	Dibromomethane	0.257	0.280	-8.9	102	0.00
47 T	Bromodichloromethane	0.510	0.590	-15.7	106	0.00
48 T	Methyl methacrylate	0.488	0.572	-17.2	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
 Data File : VN070340.D
 Acq On : 29 Dec 2021 10:36
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 30 01:48:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.009	-28.6#	113	0.00
50 S	Toluene-d8	1.236	1.196	3.2	92	0.00
51 T	4-Methyl-2-Pentanone	0.643	0.738	-14.8	104	0.00
52 CM	Toluene	0.929	1.026	-10.4#	102	0.00
53 T	t-1,3-Dichloropropene	0.552	0.643	-16.5	104	0.00
54 T	cis-1,3-Dichloropropene	0.597	0.700	-17.3	104	0.00
55 T	1,1,2-Trichloroethane	0.367	0.402	-9.5	102	0.00
56 T	Ethyl methacrylate	0.607	0.693	-14.2	102	0.00
57 T	1,3-Dichloropropane	0.650	0.711	-9.4	102	0.00
58 T	2-Chloroethyl Vinyl ether	0.127	0.176	-38.6#	119	0.00
59 T	2-Hexanone	0.469	0.563	-20.0	106	0.00
60 T	Dibromochloromethane	0.359	0.433	-20.6	108	0.00
61 T	1,2-Dibromoethane	0.373	0.412	-10.5	102	0.00
62 S	4-Bromofluorobenzene	0.446	0.451	-1.1	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.362	0.366	-1.1	100	0.00
65 PM	Chlorobenzene	1.061	1.129	-6.4	102	0.00
66 T	1,1,1,2-Tetrachloroethane	0.379	0.416	-9.8	105	0.00
67 C	Ethyl Benzene	1.944	2.127	-9.4#	102	0.00
68 T	m/p-Xylenes	0.708	0.791	-11.7	103	0.00
69 T	o-Xylene	0.704	0.776	-10.2	101	0.00
70 T	Styrene	1.114	1.302	-16.9	104	0.00
71 P	Bromoform	0.281	0.358	-27.4#	115	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	4.643	4.772	-2.8	102	0.00
74 T	N-amyl acetate	1.952	2.124	-8.8	107	0.00
75 P	1,1,2,2-Tetrachloroethane	1.559	1.541	1.2	102	0.00
76 T	1,2,3-Trichloropropane	1.501	1.328	11.5	97	0.00
77 T	Bromobenzene	1.066	1.097	-2.9	102	0.00
78 T	n-propylbenzene	5.145	5.640	-9.6	106	0.00
79 T	2-Chlorotoluene	3.301	3.370	-2.1	103	0.00
80 T	1,3,5-Trimethylbenzene	3.721	3.926	-5.5	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.409	0.505	-23.5	113	0.00
82 T	4-Chlorotoluene	3.092	3.319	-7.3	106	0.00
83 T	tert-Butylbenzene	3.276	3.372	-2.9	102	0.00
84 T	1,2,4-Trimethylbenzene	3.583	3.857	-7.6	103	0.00
85 T	sec-Butylbenzene	4.448	4.843	-8.9	105	0.00
86 T	p-Isopropyltoluene	3.494	3.879	-11.0	105	0.00
87 T	1,3-Dichlorobenzene	1.785	1.883	-5.5	105	0.00
88 T	1,4-Dichlorobenzene	1.760	1.837	-4.4	106	0.00
89 T	n-Butylbenzene	2.915	3.337	-14.5	110	0.00
90 T	Hexachloroethane	0.607	0.751	-23.7	117	0.00
91 T	1,2-Dichlorobenzene	1.749	1.786	-2.1	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.269	0.296	-10.0	109	0.00
93 T	1,2,4-Trichlorobenzene	0.824	0.925	-12.3	112	0.00
94 T	Hexachlorobutadiene	0.479	0.530	-10.6	111	0.00
95 T	Naphthalene	2.240	2.436	-8.7	108	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122921\
Data File : VN070340.D
Acq On : 29 Dec 2021 10:36
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Dec 30 01:48:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.804	0.875	-8.8	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6