

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122923\
 Data File : VN080605.D
 Acq On : 29 Dec 2023 10:23
 Operator : JC\MD
 Sample : VN1229WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1229WBL01

Quant Time: Dec 29 23:37:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122723W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 28 03:07:01 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	388605	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	830773	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	723496	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	319322	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	223769	44.891	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.780%
35) Dibromofluoromethane	8.165	113	201338	37.472	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	74.940%#
50) Toluene-d8	10.571	98	660404	35.202	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	70.400%#
62) 4-Bromofluorobenzene	12.853	95	300285	47.561	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122923\
Data File : VN080605.D
Acq On : 29 Dec 2023 10:23
Operator : JC\MD
Sample : VN1229WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1229WBL01

Quant Time: Dec 29 23:37:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122723W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 28 03:07:01 2023
Response via : Initial Calibration

