

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123021\
 Data File : VN070390.D
 Acq On : 30 Dec 2021 10:53
 Operator : JC/MD
 Sample : VN1230WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1230WBL01

Quant Time: Dec 30 17:32:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	974691	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1681355	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	1512398	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	490764	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	762664	54.421	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	108.840%
35) Dibromofluoromethane	8.021	113	548771	53.944	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.880%
50) Toluene-d8	10.443	98	2066379	49.723	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.440%
62) 4-Bromofluorobenzene	12.733	95	726308	48.389	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123021\
Data File : VN070390.D
Acq On : 30 Dec 2021 10:53
Operator : JC/MD
Sample : VN1230WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1230WBL01

Quant Time: Dec 30 17:32:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

