

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123021\
 Data File : VN070402.D
 Acq On : 30 Dec 2021 16:24
 Operator : JC/MD
 Sample : M5177-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Dec 31 05:20:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	958022	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1644058	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	1440107	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	463468	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	671493	48.749	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.500%
35) Dibromofluoromethane	8.024	113	473182	47.569	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.140%
50) Toluene-d8	10.443	98	1804359	44.403	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	88.800%
62) 4-Bromofluorobenzene	12.733	95	623650	42.492	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	84.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123021\
Data File : VN070402.D
Acq On : 30 Dec 2021 16:24
Operator : JC/MD
Sample : M5177-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Dec 31 05:20:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

