

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123021\
 Data File : VN070404.D
 Acq On : 30 Dec 2021 17:15
 Operator : JC/MD
 Sample : M5212-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-50DB

Quant Time: Dec 31 05:21:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	873798	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	1507493	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	1333480	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	430787	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	670924	53.402	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.800%
35) Dibromofluoromethane	8.024	113	477340	52.334	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.660%
50) Toluene-d8	10.443	98	1799374	48.292	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.580%
62) 4-Bromofluorobenzene	12.733	95	614313	45.648	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.300%
Target Compounds						
						Qvalue
16) Acetone	4.296	43	29678	3.639	ug/l	98
19) Methyl tert-butyl Ether	5.626	73	331081	10.018	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123021\
Data File : VN070404.D
Acq On : 30 Dec 2021 17:15
Operator : JC/MD
Sample : M5212-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-50DB

Quant Time: Dec 31 05:21:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

