

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123021\
 Data File : VN070410.D
 Acq On : 30 Dec 2021 19:49
 Operator : JC/MD
 Sample : M5234-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 P-1

Quant Time: Dec 31 05:23:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	874377	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	1494003	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	1321508	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	429235	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	658345	52.366	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.740%	
35) Dibromofluoromethane	8.021	113	468799	51.861	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.720%	
50) Toluene-d8	10.443	98	1768476	47.891	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 95.780%	
62) 4-Bromofluorobenzene	12.733	95	614458	46.071	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 92.140%	
Target Compounds						
						Qvalue
16) Acetone	4.299	43	29825	3.655	ug/l	100
30) Chloroform	7.820	83	34228	1.740	ug/l	98
64) Tetrachloroethene	10.974	164	2870	0.300	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123021\
Data File : VN070410.D
Acq On : 30 Dec 2021 19:49
Operator : JC/MD
Sample : M5234-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
P-1

Quant Time: Dec 31 05:23:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

